

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092118\
 Data File : P0049215.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Sep 2018 20:03
 Operator : SM/SJ
 Sample : J4777-31
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PL-04-092018-D

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 00:08:43 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 20 02:02:28 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.392	3.523	3268803	3344309	17.605	16.183
2) SA Decachlor...	10.109	8.458	2456042	1602516	8.935	7.828

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092118\
Data File : PO049215.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Sep 2018 20:03
Operator : SM/SJ
Sample : J4777-31
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PL-04-092018-D

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 00:08:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 20 02:02:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.392 min
Delta R.T.: -0.004 min
Response: 3268803
Conc: 17.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
PL-04-092018-D




#1 Tetrachloro-m-xylene

R.T.: 3.523 min
Delta R.T.: 0.000 min
Response: 3344309
Conc: 16.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.109 min
Delta R.T.: -0.005 min
Response: 2456042
Conc: 8.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.458 min
Delta R.T.: -0.003 min
Response: 1602516
Conc: 7.83 ng/ml