

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092518\
 Data File : P0049309.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Sep 2018 00:04
 Operator : SM/SJ
 Sample : J4779-09
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 01:25:21 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 20 02:02:28 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.395	3.528	3491310	3073703	18.803	14.874
2)	SA Decachlor...	10.112	8.460	3980120	2498633	14.479	12.205
Target Compounds							
6)	L1 AR-1016-4	0.000	4.695	0	192873	N.D.	24.017 #
7)	L1 AR-1016-5	0.000	5.055	0	339631	N.D.	49.065 #
10)	L2 AR-1221-3	0.000	3.939	0	540013	N.D.	96.668 #
11)	L3 AR-1232-1	0.000	4.415	0	304874	N.D.	130.227 #
13)	L3 AR-1232-3	0.000	4.883	0	368963	N.D.	169.674 #
14)	L3 AR-1232-4	0.000	5.194	0	257109	N.D.	108.373 #
15)	L3 AR-1232-5	0.000	5.787	0	1734266	N.D.	6963.652 #
18)	L4 AR-1242-3	0.000	4.883	0	368963	N.D.	205.720 #
19)	L4 AR-1242-4	0.000	5.194	0	257109	N.D.	161.997 #
20)	L4 AR-1242-5	0.000	5.647	0	1825651	N.D.	437.600 #
21)	L5 AR-1248-1	0.000	4.883	0	368963	N.D.	52.892 #
22)	L5 AR-1248-2	0.000	5.389	0	496284	N.D.	38.188 #
23)	L5 AR-1248-3	0.000	5.437	0	378462	N.D.	40.908 #
24)	L5 AR-1248-4	0.000	5.601	0	1015990	N.D.	125.181 #
25)	L5 AR-1248-5	0.000	5.949	0	376186	N.D.	61.357 #
26)	L6 AR-1254-1	0.000	5.527	0	337602	N.D.	27.516 #
27)	L6 AR-1254-2	0.000	5.841	0	422728	N.D.	39.502 #
31)	L7 AR-1260-1	0.000	6.053	0	211182	N.D.	14.687 #
36)	L8 AR-1262-1	0.000	5.949	0	376186	N.D.	52.704 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092518\
Data File : PO049309.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Sep 2018 00:04
Operator : SM/SJ
Sample : J4779-09
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 01:25:21 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 20 02:02:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: 0.000 min
Response: 3491310
Conc: 18.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.528 min
Delta R.T.: 0.004 min
Response: 3073703
Conc: 14.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.112 min
Delta R.T.: -0.002 min
Response: 3980120
Conc: 14.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.460 min
Delta R.T.: 0.000 min
Response: 2498633
Conc: 12.21 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 5.652 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#6 AR-1016-4

R.T.: 4.695 min
Delta R.T.: 0.029 min
Response: 192873
Conc: 24.02 ng/ml




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T.: 6.046 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.055 min
Delta R.T.: 0.018 min
Response: 339631
Conc: 49.07 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.769 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 3.939 min
Delta R.T.: 0.045 min
Response: 540013
Conc: 96.67 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 5.496 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.415 min
Delta R.T.: 0.040 min
Response: 304874
Conc: 130.23 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 5.924 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.883 min
Delta R.T.: 0.015 min
Response: 368963
Conc: 169.67 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T.: 6.254 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.194 min
Delta R.T.: 0.009 min
Response: 257109
Conc: 108.37 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 7.132 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.787 min
Delta R.T.: 0.030 min
Response: 1734266
Conc: 6963.65 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T.: 6.087 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 4.883 min
Delta R.T.: -0.016 min
Response: 368963
Conc: 205.72 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 6.131 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.194 min
Delta R.T.: -0.061 min
Response: 257109
Conc: 162.00 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 6.188 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.647 min
Delta R.T.: 0.004 min
Response: 1825651
Conc: 437.60 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 5.840 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.883 min
Delta R.T.: 0.056 min
Response: 368963
Conc: 52.89 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T.: 6.421 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 5.389 min
Delta R.T.: 0.005 min
Response: 496284
Conc: 38.19 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 6.450 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.437 min
Delta R.T.: 0.012 min
Response: 378462
Conc: 40.91 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T.: 6.488 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.601 min
Delta R.T.: 0.003 min
Response: 1015990
Conc: 125.18 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 6.686 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.949 min
Delta R.T.: 0.019 min
Response: 376186
Conc: 61.36 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.641 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.527 min
Delta R.T.: -0.004 min
Response: 337602
Conc: 27.52 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.923 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.841 min
Delta R.T.: 0.000 min
Response: 422728
Conc: 39.50 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T.: 7.172 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.053 min
Delta R.T.: -0.007 min
Response: 211182
Conc: 14.69 ng/ml



#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.064 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#36 AR-1262-1

R.T.: 5.949 min
Delta R.T.: 0.003 min
Response: 376186
Conc: 52.70 ng/ml