

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092619\
 Data File : P0061390.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Sep 2019 17:11
 Operator : SM/AJ
 Sample : PB123414BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB123414BL

Manual Integrations
 APPROVED

Ankita
 9/27/2019 1:27:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 05:18:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.363	3.519	713196	587600	19.385	21.044m
2) SA Decachlor...	10.028	8.395	983781	681781	21.412	19.451m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092619\
Data File : P0061390.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Sep 2019 17:11
Operator : SM/AJ
Sample : PB123414BL
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB123414BL


Manual Integrations
APPROVED

Ankita
9/27/2019 1:27:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 05:18:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.363 min
Delta R.T.: -0.007 min
Response: 713196
Conc: 19.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123414BL


Manual Integrations
APPROVED

Ankita
9/27/2019 1:27:58 PM




#1 Tetrachloro-m-xylene

R.T.: 3.519 min
Delta R.T.: -0.012 min
Response: 587600
Conc: 21.04 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.028 min
Delta R.T.: 0.000 min
Response: 983781
Conc: 21.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.395 min
Delta R.T.: -0.018 min
Response: 681781
Conc: 19.45 ng/ml m