

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100118\
 Data File : P0049501.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2018 8:01
 Operator : SM/SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 05:15:52 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 01 14:47:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
2)	SA Decachlor...	0.000	8.442	0	156288	N.D.	0.512 #
Target Compounds							
28)	L6 AR-1254-3	0.000	5.917	0	324207	N.D.	9.895 #
29)	L6 AR-1254-4	0.000	6.143	0	136808	N.D.	6.007 #
30)	L6 AR-1254-5	0.000	6.557	0	531649	N.D.	17.713 #
31)	L7 AR-1260-1	0.000	6.045	0	284186	N.D.	13.844 #
32)	L7 AR-1260-2	0.000	6.232	0	423277	N.D.	15.899 #
33)	L7 AR-1260-3	0.000	6.384	0	400253	N.D.	16.887 #
34)	L7 AR-1260-4	0.000	6.852	0	162319	N.D.	8.998 #
35)	L7 AR-1260-5	0.000	7.092	0	576871	N.D.	12.346 #
36)	L8 AR-1262-1	0.000	6.594	0	213398	N.D.	7.389 #
37)	L8 AR-1262-2	0.000	7.092	0	576871	N.D.	11.566 #
38)	L8 AR-1262-3	0.000	7.437	0	339308	N.D.	17.147 #
39)	L8 AR-1262-4	0.000	7.437	0	339308	N.D.	9.630 #
41)	L9 AR-1268-1	0.000	7.437	0	339308	N.D.	5.345 #
42)	L9 AR-1268-2	0.000	7.437	0	339308	N.D.	5.886 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100118\
Data File : PO049501.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Oct 2018 8:01
Operator : SM/SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 05:15:52 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 01 14:47:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.383 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.514 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.094 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 8.442 min
Delta R.T.: -0.002 min
Response: 156288
Conc: 0.51 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 7.000 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#28 AR-1254-3

R.T.: 5.917 min
Delta R.T.: -0.006 min
Response: 324207
Conc: 9.90 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T.: 7.287 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.143 min
Delta R.T.: -0.008 min
Response: 136808
Conc: 6.01 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.705 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#30 AR-1254-5

R.T.: 6.557 min
Delta R.T.: -0.008 min
Response: 531649
Conc: 17.71 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T.: 7.159 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.045 min
Delta R.T.: -0.002 min
Response: 284186
Conc: 13.84 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.415 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#32 AR-1260-2

R.T.: 6.232 min
Delta R.T.: -0.002 min
Response: 423277
Conc: 15.90 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 7.774 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 6.384 min
Delta R.T.: -0.002 min
Response: 400253
Conc: 16.89 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 8.000 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#34 AR-1260-4

R.T.: 6.852 min
Delta R.T.: -0.002 min
Response: 162319
Conc: 9.00 ng/ml




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T.: 8.317 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 7.092 min
Delta R.T.: -0.002 min
Response: 576871
Conc: 12.35 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 7.779 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#36 AR-1262-1

R.T.: 6.594 min
Delta R.T.: -0.008 min
Response: 213398
Conc: 7.39 ng/ml




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T.: 8.323 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 7.092 min
Delta R.T.: -0.007 min
Response: 576871
Conc: 11.57 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 8.628 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#38 AR-1262-3

R.T.: 7.437 min
Delta R.T.: 0.056 min
Response: 339308
Conc: 17.15 ng/ml




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 8.719 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 7.437 min
Delta R.T.: -0.008 min
Response: 339308
Conc: 9.63 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.626 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#41 AR-1268-1

R.T.: 7.437 min
Delta R.T.: 0.055 min
Response: 339308
Conc: 5.35 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 8.720 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 7.437 min
Delta R.T.: -0.009 min
Response: 339308
Conc: 5.89 ng/ml