

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100118\
 Data File : P0049525.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2018 17:20
 Operator : SM/SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 02:01:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 01 14:47:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.383	3.515	8199749	7633178	23.220	22.443
2)	SA Decachlor...	10.094	8.446	9675999	6251158	20.420	20.460

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100118\
Data File : PO049525.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Oct 2018 17:20
Operator : SM/SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 02:01:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 01 14:47:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.383 min
Delta R.T.: 0.000 min
Response: 8199749
Conc: 23.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.515 min
Delta R.T.: 0.000 min
Response: 7633178
Conc: 22.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.000 min
Response: 9675999
Conc: 20.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.446 min
Delta R.T.: 0.002 min
Response: 6251158
Conc: 20.46 ng/ml