

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO100119\
 Data File : PO061519.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2019 15:20
 Operator : HP/AJ
 Sample : PO100119ICV500
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 01:31:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	41.014	40.286 E3	1.8	94	0.00
2 SA Decachlorobiphenyl	46.571	47.373 E3	-1.7	103	0.00
3 L1 AR-1016-1	1.834	1.784 E3	2.7	99	0.00
4 L1 AR-1016-2	2.585	2.560 E3	1.0	99	0.00
5 L1 AR-1016-3	1.606	1.591 E3	0.9	100	0.00
6 L1 AR-1016-4	1.337	1.335 E3	0.1	100	0.00
7 L1 AR-1016-5	1.371	1.383 E3	-0.9	101	0.00
31 L7 AR-1260-1	2.611	2.650 E3	-1.5	102	0.00
32 L7 AR-1260-2	3.188	3.231 E3	-1.3	104	0.00
33 L7 AR-1260-3	2.402	2.490 E3	-3.7	105	0.00
34 L7 AR-1260-4	2.672	2.756 E3	-3.1	104	0.00
35 L7 AR-1260-5	4.948	5.037 E3	-1.8	103	0.00

Signal #2

1 SA Tetrachloro-m-xylene	32.339	31.026 E3	4.1	95	0.00
2 SA Decachlorobiphenyl	32.376	32.672 E3	-0.9	104	0.00
3 L1 AR-1016-1	1.345	1.274 E3	5.3	98	0.00
4 L1 AR-1016-2	1.853	1.789 E3	3.5	99	0.00
5 L1 AR-1016-3	1001.720	1014.123	-1.2	103	0.00
6 L1 AR-1016-4	866.564	849.183	2.0	100	0.00
7 L1 AR-1016-5	1.141	1.109 E3	2.8	101	0.00
31 L7 AR-1260-1	2.040	1.997 E3	2.1	104	0.00
32 L7 AR-1260-2	2.402	2.345 E3	2.4	103	0.00
33 L7 AR-1260-3	2.253	2.231 E3	1.0	104	0.00
34 L7 AR-1260-4	1.846	1.844 E3	0.1	105	0.00
35 L7 AR-1260-5	3.894	3.916 E3	-0.6	104	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	462.284	0.000	100.0#	0#	-4.57#
9 L2 AR-1221-2	359.022	0.000	100.0#	0#	-4.66#
10 L2 AR-1221-3	1.139	0.000 E3	100.0#	0#	-4.74#
11 L3 AR-1232-1	1.398	0.000 E3	100.0#	0#	-4.74#
12 L3 AR-1232-2	784.559	0.000	100.0#	0#	-5.26#
13 L3 AR-1232-3	1.709	0.000 E3	100.0#	0#	-5.55#
14 L3 AR-1232-4	884.786	0.000	100.0#	0#	-5.71#
15 L3 AR-1232-5	693.332	0.000	100.0#	0#	-5.80#
16 L4 AR-1242-1	1.420	0.000 E3	100.0#	0#	-5.53#
17 L4 AR-1242-2	2.010	0.000 E3	100.0#	0#	-5.55#
18 L4 AR-1242-3	1.262	0.000 E3	100.0#	0#	-5.61#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO100119\
 Data File : PO061519.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2019 15:20
 Operator : HP/AJ
 Sample : PO100119ICV500
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 01:31:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound		AvgRF	CCRF	%Dev	Area%	Dev(Min)

19 L4	AR-1242-4	1.051	0.000 E3	100.0#	0#	-5.71#
20 L4	AR-1242-5	1.337	0.000 E3	100.0#	0#	-6.44#
21 L5	AR-1248-1	1.175	0.000 E3	100.0#	0#	-5.53#
22 L5	AR-1248-2	1.687	0.000 E3	100.0#	0#	-5.80#
23 L5	AR-1248-3	1.869	0.000 E3	100.0#	0#	-6.00#
24 L5	AR-1248-4	2.363	0.000 E3	100.0#	0#	-6.41#
25 L5	AR-1248-5	2.267	0.000 E3	100.0#	0#	-6.44#
26 L6	AR-1254-1	2.281	0.000 E3	100.0#	0#	-6.38#
27 L6	AR-1254-2	3.613	0.000 E3	100.0#	0#	-6.60#
28 L6	AR-1254-3	3.907	0.000 E3	100.0#	0#	-6.96#
29 L6	AR-1254-4	3.029	0.000 E3	100.0#	0#	-7.25#
30 L6	AR-1254-5	3.140	0.000 E3	100.0#	0#	-7.67#
36 L8	AR-1262-1	3.831	0.000 E3	100.0#	0#	-7.74#
37 L8	AR-1262-2	6.208	0.000 E3	100.0#	0#	-8.28#
38 L8	AR-1262-3	4.156	0.000 E3	100.0#	0#	-8.58#
39 L8	AR-1262-4	3.157	0.000 E3	100.0#	0#	-8.67#
40 L8	AR-1262-5	2.034	0.000 E3	100.0#	0#	-9.31#
41 L9	AR-1268-1	6.724	0.000 E3	100.0#	0#	-8.58#
42 L9	AR-1268-2	6.192	0.000 E3	100.0#	0#	-8.67#
43 L9	AR-1268-3	5.112	0.000 E3	100.0#	0#	-8.89#
44 L9	AR-1268-4	2.187	0.000 E3	100.0#	0#	-9.31#
45 L9	AR-1268-5	15.023	0.000 E3	100.0#	0#	-9.71#

Signal #2

8 L2	AR-1221-1	395.975	0.000	100.0#	0#	-3.73#
9 L2	AR-1221-2	308.393	0.000	100.0#	0#	-3.82#
10 L2	AR-1221-3	912.139	0.000	100.0#	0#	-3.89#
11 L3	AR-1232-1	1.123	0.000 E3	100.0#	0#	-3.89#
12 L3	AR-1232-2	1.221	0.000 E3	100.0#	0#	-4.59#
13 L3	AR-1232-3	660.254	0.000	100.0#	0#	-4.77#
14 L3	AR-1232-4	630.353	0.000	100.0#	0#	-4.85#
15 L3	AR-1232-5	686.150	0.000	100.0#	0#	-5.01#
16 L4	AR-1242-1	1.025	0.000 E3	100.0#	0#	-4.58#
17 L4	AR-1242-2	1.410	0.000 E3	100.0#	0#	-4.60#
18 L4	AR-1242-3	780.884	0.000	100.0#	0#	-4.77#
19 L4	AR-1242-4	820.599	0.000	100.0#	0#	-4.85#
20 L4	AR-1242-5	1.071	0.000 E3	100.0#	0#	-5.36#
21 L5	AR-1248-1	894.816	0.000	100.0#	0#	-4.58#
22 L5	AR-1248-2	1.178	0.000 E3	100.0#	0#	-4.81#
23 L5	AR-1248-3	1.241	0.000 E3	100.0#	0#	-4.85#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0100119\
 Data File : P0061519.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2019 15:20
 Operator : HP/AJ
 Sample : P0100119ICV500
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 01:31:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.503	0.000 E3	100.0#	0#	-5.01#
25 L5 AR-1248-5	1.534	0.000 E3	100.0#	0#	-5.40#
26 L6 AR-1254-1	2.193	0.000 E3	100.0#	0#	-5.36#
27 L6 AR-1254-2	1.960	0.000 E3	100.0#	0#	-5.50#
28 L6 AR-1254-3	3.184	0.000 E3	100.0#	0#	-5.89#
29 L6 AR-1254-4	2.029	0.000 E3	100.0#	0#	-6.12#
30 L6 AR-1254-5	2.828	0.000 E3	100.0#	0#	-6.53#
36 L8 AR-1262-1	1.221	0.000 E3	100.0#	0#	-6.62#
37 L8 AR-1262-2	4.598	0.000 E3	100.0#	0#	-7.06#
38 L8 AR-1262-3	1.884	0.000 E3	100.0#	0#	-7.34#
39 L8 AR-1262-4	3.361	0.000 E3	100.0#	0#	-7.40#
40 L8 AR-1262-5	1.516	0.000 E3	100.0#	0#	-7.89#
41 L9 AR-1268-1	4.829	0.000 E3	100.0#	0#	-7.34#
42 L9 AR-1268-2	4.310	0.000 E3	100.0#	0#	-7.40#
43 L9 AR-1268-3	3.584	0.000 E3	100.0#	0#	-7.61#
44 L9 AR-1268-4	1.609	0.000 E3	100.0#	0#	-7.89#
45 L9 AR-1268-5	10.056	0.000 E3	100.0#	0#	-8.16#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0