Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100319\
Data File : PO061647.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Oct 2019 17:27
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 01:29:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.389	3.539	1995529	1411591	48.654	43.649	
2)	SA Decachlor...	10.061	8.409	1791743	1461578	38.473	45.144	
	Target Compounds							
3)	L1 AR-1016-1	5.553	4.593	922926	634845	503.300	472.122	
4)	L1 AR-1016-2	5.575	4.610	1320260	879886	510.775	474.752	
5)	L1 AR-1016-3	5.636	4.781	814666	499021	507.125	498.164	
6)	L1 AR-1016-4	5.735	4.823	687582	421910	514.294	486.877	
7)	L1 AR-1016-5	6.027	5.029	706300	554103	515.011	485.443	
8)	L2 AR-1221-1	4.557	3.747	5765	46112	12.470	116.451	#
9)	L2 AR-1221-2	4.685	3.831	206833	78830	576.101	255.615	#
10)	L2 AR-1221-3	4.760	3.904	390622	281688	343.095	308.822	
11)	L3 AR-1232-1	4.760	3.904	390622	281688	279.333	250.881	
12)	L3 AR-1232-2	5.238	4.593	48804	634845	62.206	520.072	#
13)	L3 AR-1232-3	5.553	4.781	922926	499021	540.139	755.802	#
14)	L3 AR-1232-4	5.693	4.863	476	519441	0.538	824.048	#
15)	L3 AR-1232-5	5.824	5.029	609223	554103	878.689	807.553	
16)	L4 AR-1242-1	5.553	4.593	922926	634845	649.873	619.661	
17)	L4 AR-1242-2	5.575	4.610	1320260	879886	656.899	623.934	
18)	L4 AR-1242-3	5.636	4.781	814666	499021	645.600	639.047	
19)	L4 AR-1242-4	5.735	4.863	687582	519441	654.362	633.002	
20)	L4 AR-1242-5	6.469	5.371	131503	431903	98.321	403.179	#
21)	L5 AR-1248-1	5.553	4.593	922926	634845	785.591	709.469	
22)	L5 AR-1248-2	5.824	4.823	609223	421910	361.092	358.078	
23)	L5 AR-1248-3	6.027	4.863	706300	519441	377.935	418.472	
24)	L5 AR-1248-4	6.429	5.029	164446	554103	69.597	368.591	#
25)	L5 AR-1248-5	6.469	5.410	131503	79371	57.996	51.736	
26)	L6 AR-1254-1	6.402	5.371	573349	431903	251.315	196.920	
27)	L6 AR-1254-2	6.620	5.515	624993	411730	172.981	210.014	
28)	L6 AR-1254-3	6.985	5.889	367184	183895	93.973	57.760	#
29)	L6 AR-1254-4	7.269	6.133	279306	130171	92.222	64.160	#
30)	L6 AR-1254-5	7.686	6.543	2033040	1383342	647.381	489.079	
31)	L7 AR-1260-1	7.147	6.037	1478355	1068238	566.302	523.698	
32)	L7 AR-1260-2	7.404	6.222	1788075	1273907	560.878	530.411	
33)	L7 AR-1260-3	7.761	6.372	1350698	1186055	562.354	526.407	
34)	L7 AR-1260-4	7.986	6.835	1480640	933785	554.122	505.785	
35)	L7 AR-1260-5	8.302	7.075	2549597	1869331	515.265	480.100	
36)	L8 AR-1262-1	7.761	6.633	1350698	544917	352.527	446.420	#
37)	L8 AR-1262-2	8.302	7.075	2549597	1869331	410.690	406.551	
38)	L8 AR-1262-3	8.621	7.353	1528043	427668	367.644	227.021	#
39)	L8 AR-1262-4	8.677	7.416	530605	1280640	168.080	381.059	#
40)	L8 AR-1262-5	9.281	7.904	518	461385	0.255	304.350	#
41)	L9 AR-1268-1	8.621	7.353	1528043	427668	227.238	88.560	#

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100319\
Data File : PO061647.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Oct 2019 17:27
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 01:29:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.677	7.416	530605	1280640	85.686	297.144 #
43) L9	AR-1268-3	8.916	7.618	62814	42134	12.288	11.755
44) L9	AR-1268-4	9.334	7.904	653298	461385	298.775	286.684
45) L9	AR-1268-5	9.685	8.175	305	129948	0.020	12.922 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100319\
Data File : PO061647.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Oct 2019 17:27
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 01:29:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.389 min
Delta R.T.: 0.000 min
Response: 1995529
Conc: 48.65 ng/ml

Instrument :
ECD_O
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 1411591
Conc: 43.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: 0.025 min
Response: 1791743
Conc: 38.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.409 min
Delta R.T.: 0.000 min
Response: 1461578
Conc: 45.14 ng/ml



#3 AR-1016-1

R.T.: 5.553 min
Delta R.T.: 0.003 min
Response: 922926
Conc: 503.30 ng/ml

Instrument :
ECD_O
ClientSampleId :



#3 AR-1016-1

R.T.: 4.593 min
Delta R.T.: 0.003 min
Response: 634845
Conc: 472.12 ng/ml



#4 AR-1016-2

R.T.: 5.575 min
Delta R.T.: 0.005 min
Response: 1320260
Conc: 510.77 ng/ml



#4 AR-1016-2

R.T.: 4.610 min
Delta R.T.: 0.000 min
Response: 879886
Conc: 474.75 ng/ml



#5 AR-1016-3

R.T.: 5.636 min
Delta R.T.: -0.004 min
Response: 814666
Conc: 507.12 ng/ml

Instrument :
ECD_O
ClientSampleId :



#5 AR-1016-3

R.T.: 4.781 min
Delta R.T.: 0.000 min
Response: 499021
Conc: 498.16 ng/ml



#6 AR-1016-4

R.T.: 5.735 min
Delta R.T.: 0.005 min
Response: 687582
Conc: 514.29 ng/ml



#6 AR-1016-4

R.T.: 4.823 min
Delta R.T.: 0.003 min
Response: 421910
Conc: 486.88 ng/ml



#7 AR-1016-5

R.T.: 6.027 min
Delta R.T.: -0.003 min
Response: 706300
Conc: 515.01 ng/ml

Instrument :
ECD_O
ClientSampleId :



#7 AR-1016-5

R.T.: 5.029 min
Delta R.T.: 0.000 min
Response: 554103
Conc: 485.44 ng/ml



#8 AR-1221-1

R.T.: 4.557 min
Delta R.T.: -0.017 min
Response: 5765
Conc: 12.47 ng/ml



#8 AR-1221-1

R.T.: 3.747 min
Delta R.T.: 0.013 min
Response: 46112
Conc: 116.45 ng/ml



#9 AR-1221-2

R.T.: 4.685 min
Delta R.T.: 0.025 min
Response: 206833
Conc: 576.10 ng/ml

Instrument :
ECD_O
ClientSampleId :



#9 AR-1221-2

R.T.: 3.831 min
Delta R.T.: 0.014 min
Response: 78830
Conc: 255.62 ng/ml



#10 AR-1221-3

R.T.: 4.760 min
Delta R.T.: 0.024 min
Response: 390622
Conc: 343.09 ng/ml



#10 AR-1221-3

R.T.: 3.904 min
Delta R.T.: 0.014 min
Response: 281688
Conc: 308.82 ng/ml



#11 AR-1232-1

R.T.: 4.760 min
Delta R.T.: 0.024 min
Response: 390622
Conc: 279.33 ng/ml

Instrument :
ECD_O
ClientSampleId :



#11 AR-1232-1

R.T.: 3.904 min
Delta R.T.: 0.014 min
Response: 281688
Conc: 250.88 ng/ml



#12 AR-1232-2

R.T.: 5.238 min
Delta R.T.: -0.023 min
Response: 48804
Conc: 62.21 ng/ml



#12 AR-1232-2

R.T.: 4.593 min
Delta R.T.: -0.002 min
Response: 634845
Conc: 520.07 ng/ml



#13 AR-1232-3

R.T.: 5.553 min
Delta R.T.: 0.003 min
Response: 922926
Conc: 540.14 ng/ml

Instrument :
ECD_O
ClientSampleId :



#13 AR-1232-3

R.T.: 4.781 min
Delta R.T.: 0.016 min
Response: 499021
Conc: 755.80 ng/ml



#14 AR-1232-4

R.T.: 5.693 min
Delta R.T.: -0.017 min
Response: 476
Conc: 0.54 ng/ml



#14 AR-1232-4

R.T.: 4.863 min
Delta R.T.: 0.015 min
Response: 519441
Conc: 824.05 ng/ml



#15 AR-1232-5

R.T.: 5.824 min
Delta R.T.: 0.025 min
Response: 609223
Conc: 878.69 ng/ml

Instrument :
ECD_O
ClientSampleId :



#15 AR-1232-5

R.T.: 5.029 min
Delta R.T.: 0.016 min
Response: 554103
Conc: 807.55 ng/ml



#16 AR-1242-1

R.T.: 5.553 min
Delta R.T.: 0.003 min
Response: 922926
Conc: 649.87 ng/ml



#16 AR-1242-1

R.T.: 4.593 min
Delta R.T.: 0.003 min
Response: 634845
Conc: 619.66 ng/ml



#17 AR-1242-2

R.T.: 5.575 min
Delta R.T.: 0.005 min
Response: 1320260
Conc: 656.90 ng/ml

Instrument :
ECD_O
ClientSampleId :



#17 AR-1242-2

R.T.: 4.610 min
Delta R.T.: 0.000 min
Response: 879886
Conc: 623.93 ng/ml



#18 AR-1242-3

R.T.: 5.636 min
Delta R.T.: -0.004 min
Response: 814666
Conc: 645.60 ng/ml



#18 AR-1242-3

R.T.: 4.781 min
Delta R.T.: 0.000 min
Response: 499021
Conc: 639.05 ng/ml



#19 AR-1242-4

R.T.: 5.735 min
Delta R.T.: 0.005 min
Response: 687582
Conc: 654.36 ng/ml

Instrument :
ECD_O
ClientSampleId :



#19 AR-1242-4

R.T.: 4.863 min
Delta R.T.: 0.003 min
Response: 519441
Conc: 633.00 ng/ml



#20 AR-1242-5

R.T.: 6.469 min
Delta R.T.: -0.001 min
Response: 131503
Conc: 98.32 ng/ml



#20 AR-1242-5

R.T.: 5.371 min
Delta R.T.: 0.001 min
Response: 431903
Conc: 403.18 ng/ml



#21 AR-1248-1

R.T.: 5.553 min
Delta R.T.: 0.003 min
Response: 922926
Conc: 785.59 ng/ml

Instrument :
ECD_O
ClientSampleId :



#21 AR-1248-1

R.T.: 4.593 min
Delta R.T.: 0.003 min
Response: 634845
Conc: 709.47 ng/ml



#22 AR-1248-2

R.T.: 5.824 min
Delta R.T.: 0.004 min
Response: 609223
Conc: 361.09 ng/ml



#22 AR-1248-2

R.T.: 4.823 min
Delta R.T.: 0.003 min
Response: 421910
Conc: 358.08 ng/ml



#23 AR-1248-3

R.T.: 6.027 min
Delta R.T.: -0.003 min
Response: 706300
Conc: 377.93 ng/ml

Instrument :
ECD_O
ClientSampleId :



#23 AR-1248-3

R.T.: 4.863 min
Delta R.T.: 0.003 min
Response: 519441
Conc: 418.47 ng/ml



#24 AR-1248-4

R.T.: 6.429 min
Delta R.T.: 0.000 min
Response: 164446
Conc: 69.60 ng/ml



#24 AR-1248-4

R.T.: 5.029 min
Delta R.T.: 0.000 min
Response: 554103
Conc: 368.59 ng/ml



#25 AR-1248-5

R.T.: 6.469 min
Delta R.T.: -0.001 min
Response: 131503
Conc: 58.00 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 5.410 min
Delta R.T.: 0.000 min
Response: 79371
Conc: 51.74 ng/ml



#26 AR-1254-1

R.T.: 6.402 min
Delta R.T.: 0.021 min
Response: 573349
Conc: 251.32 ng/ml



#26 AR-1254-1

R.T.: 5.371 min
Delta R.T.: 0.014 min
Response: 431903
Conc: 196.92 ng/ml



#27 AR-1254-2

R.T.: 6.620 min
Delta R.T.: 0.020 min
Response: 624993
Conc: 172.98 ng/ml

Instrument :
ECD_O
ClientSampleId :



#27 AR-1254-2

R.T.: 5.515 min
Delta R.T.: 0.014 min
Response: 411730
Conc: 210.01 ng/ml



#28 AR-1254-3

R.T.: 6.985 min
Delta R.T.: 0.021 min
Response: 367184
Conc: 93.97 ng/ml



#28 AR-1254-3

R.T.: 5.889 min
Delta R.T.: -0.006 min
Response: 183895
Conc: 57.76 ng/ml



#29 AR-1254-4

R.T.: 7.269 min
Delta R.T.: 0.020 min
Response: 279306
Conc: 92.22 ng/ml

Instrument :
ECD_O
ClientSampleId :



#29 AR-1254-4

R.T.: 6.133 min
Delta R.T.: 0.014 min
Response: 130171
Conc: 64.16 ng/ml



#30 AR-1254-5

R.T.: 7.686 min
Delta R.T.: 0.019 min
Response: 2033040
Conc: 647.38 ng/ml



#30 AR-1254-5

R.T.: 6.543 min
Delta R.T.: 0.013 min
Response: 1383342
Conc: 489.08 ng/ml



#31 AR-1260-1

R.T.: 7.147 min
Delta R.T.: -0.003 min
Response: 1478355
Conc: 566.30 ng/ml

Instrument :
ECD_O
ClientSampleId :



#31 AR-1260-1

R.T.: 6.037 min
Delta R.T.: -0.003 min
Response: 1068238
Conc: 523.70 ng/ml



#32 AR-1260-2

R.T.: 7.404 min
Delta R.T.: 0.004 min
Response: 1788075
Conc: 560.88 ng/ml



#32 AR-1260-2

R.T.: 6.222 min
Delta R.T.: 0.002 min
Response: 1273907
Conc: 530.41 ng/ml



#33 AR-1260-3

R.T.: 7.761 min
Delta R.T.: 0.001 min
Response: 1350698
Conc: 562.35 ng/ml

Instrument :
ECD_O
ClientSampleId :



#33 AR-1260-3

R.T.: 6.372 min
Delta R.T.: 0.002 min
Response: 1186055
Conc: 526.41 ng/ml



#34 AR-1260-4

R.T.: 7.986 min
Delta R.T.: -0.004 min
Response: 1480640
Conc: 554.12 ng/ml



#34 AR-1260-4

R.T.: 6.835 min
Delta R.T.: -0.005 min
Response: 933785
Conc: 505.79 ng/ml



#35 AR-1260-5

R.T.: 8.302 min
Delta R.T.: 0.002 min
Response: 2549597
Conc: 515.27 ng/ml

Instrument :
ECD_O
ClientSampleId :



#35 AR-1260-5

R.T.: 7.075 min
Delta R.T.: 0.005 min
Response: 1869331
Conc: 480.10 ng/ml



#36 AR-1262-1

R.T.: 7.761 min
Delta R.T.: 0.022 min
Response: 1350698
Conc: 352.53 ng/ml



#36 AR-1262-1

R.T.: 6.633 min
Delta R.T.: 0.013 min
Response: 544917
Conc: 446.42 ng/ml



#37 AR-1262-2

R.T.: 8.302 min
Delta R.T.: 0.023 min
Response: 2549597
Conc: 410.69 ng/ml

Instrument :
ECD_O
ClientSampleId :



#37 AR-1262-2

R.T.: 7.075 min
Delta R.T.: 0.014 min
Response: 1869331
Conc: 406.55 ng/ml



#38 AR-1262-3

R.T.: 8.621 min
Delta R.T.: 0.039 min
Response: 1528043
Conc: 367.64 ng/ml



#38 AR-1262-3

R.T.: 7.353 min
Delta R.T.: 0.013 min
Response: 427668
Conc: 227.02 ng/ml



#39 AR-1262-4

R.T.: 8.677 min
Delta R.T.: 0.008 min
Response: 530605
Conc: 168.08 ng/ml

Instrument :
ECD_O
ClientSampleId :



#39 AR-1262-4

R.T.: 7.416 min
Delta R.T.: 0.012 min
Response: 1280640
Conc: 381.06 ng/ml



#40 AR-1262-5

R.T.: 9.281 min
Delta R.T.: -0.025 min
Response: 518
Conc: 0.25 ng/ml



#40 AR-1262-5

R.T.: 7.904 min
Delta R.T.: 0.013 min
Response: 461385
Conc: 304.35 ng/ml



#41 AR-1268-1

R.T.: 8.621 min
Delta R.T.: 0.039 min
Response: 1528043
Conc: 227.24 ng/ml

Instrument :
ECD_O
ClientSampleId :



#41 AR-1268-1

R.T.: 7.353 min
Delta R.T.: 0.012 min
Response: 427668
Conc: 88.56 ng/ml



#42 AR-1268-2

R.T.: 8.677 min
Delta R.T.: 0.003 min
Response: 530605
Conc: 85.69 ng/ml



#42 AR-1268-2

R.T.: 7.416 min
Delta R.T.: 0.011 min
Response: 1280640
Conc: 297.14 ng/ml



#43 AR-1268-3

R.T.: 8.916 min
Delta R.T.: 0.022 min
Response: 62814
Conc: 12.29 ng/ml

Instrument :
ECD_O
ClientSampleId :



#43 AR-1268-3

R.T.: 7.618 min
Delta R.T.: 0.011 min
Response: 42134
Conc: 11.76 ng/ml



#44 AR-1268-4

R.T.: 9.334 min
Delta R.T.: 0.025 min
Response: 653298
Conc: 298.77 ng/ml



#44 AR-1268-4

R.T.: 7.904 min
Delta R.T.: 0.013 min
Response: 461385
Conc: 286.68 ng/ml



#45 AR-1268-5

R.T.: 9.685 min
Delta R.T.: -0.023 min
Response: 305
Conc: 0.02 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.175 min
Delta R.T.: 0.012 min
Response: 129948
Conc: 12.92 ng/ml