

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100319\
 Data File : P0061665.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Oct 2019 22:21
 Operator : HP/AJ
 Sample : K5160-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CSA-3-2

Manual Integrations
 APPROVED

Ankita
 10/4/2019 5:12:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 01:37:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.379	3.538	1250129	941910	30.480	29.126
2)	SA Decachlor...	10.046	8.403	843098	558842	18.103m	17.261

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100319\
Data File : PO061665.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Oct 2019 22:21
Operator : HP/AJ
Sample : K5160-03
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CSA-3-2


Manual Integrations
APPROVED

Ankita
10/4/2019 5:12:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 01:37:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.379 min
Delta R.T.: -0.009 min
Response: 1250129
Conc: 30.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-3-2


Manual Integrations
APPROVED

Ankita
10/4/2019 5:12:04 PM




#1 Tetrachloro-m-xylene

R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 941910
Conc: 29.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: 0.010 min
Response: 843098
Conc: 18.10 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.403 min
Delta R.T.: -0.006 min
Response: 558842
Conc: 17.26 ng/ml