

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100918\
 Data File : PO049805.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2018 17:54
 Operator : SM/SJ
 Sample : J5293-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 VW-CONCRETE-WALL

Manual Integrations
 APPROVED

Sohil
 10/10/2018 3:51:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 11:42:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 08 15:27:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.382	3.512	5279045	4823123	23.353m	19.997m
2)	SA Decachlor...	10.089	8.437	36606300	24619208	109.864	94.123
Target Compounds							
26)	L6 AR-1254-1	6.409	5.369	10235987	13308868	744.360	678.899m
27)	L6 AR-1254-2	6.628	5.514	20026686	12721531	907.028	760.067m
28)	L6 AR-1254-3	6.995	5.913	18714778	25876856	767.027	927.027
29)	L6 AR-1254-4	7.280	6.140	17556664	17503208	856.051	872.271
30)	L6 AR-1254-5	7.699	6.553	19428170	21645061	1008.151	808.926
31)	L7 AR-1260-1	7.156	6.042	9941508	14106098	538.810	683.151 #
32)	L7 AR-1260-2	7.415	6.229	12770832	17292349	532.551	616.725
33)	L7 AR-1260-3	7.773	6.379	6069932	12800334	321.050	525.266 #
34)	L7 AR-1260-4	7.993	6.848	8530153	4233513	449.248	219.205 #
35)	L7 AR-1260-5	8.316	7.087	8313366	12494981	196.515	243.712

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100918\
Data File : P0049805.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2018 17:54
Operator : SM/SJ
Sample : J5293-01
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:42:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100818.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 08 15:27:38 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.382 min
Delta R.T.: 0.000 min
Response: 5279045
Conc: 23.35 ng/ml m

Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM




#1 Tetrachloro-m-xylene

R.T.: 3.512 min
Delta R.T.: -0.002 min
Response: 4823123
Conc: 20.00 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.089 min
Delta R.T.: 0.003 min
Response: 36606300
Conc: 109.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.437 min
Delta R.T.: 0.000 min
Response: 24619208
Conc: 94.12 ng/ml




#26 AR-1254-1

R.T.: 6.409 min
Delta R.T.: -0.006 min
Response: 10235987
Conc: 744.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM




#26 AR-1254-1

R.T.: 5.369 min
Delta R.T.: -0.010 min
Response: 13308868
Conc: 678.90 ng/ml m




#27 AR-1254-2

R.T.: 6.628 min
Delta R.T.: -0.006 min
Response: 20026686
Conc: 907.03 ng/ml



#27 AR-1254-2

R.T.: 5.514 min
Delta R.T.: -0.010 min
Response: 12721531
Conc: 760.07 ng/ml m




#28 AR-1254-3

R.T.: 6.995 min
Delta R.T.: -0.005 min
Response: 18714778
Conc: 767.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM




#28 AR-1254-3

R.T.: 5.913 min
Delta R.T.: -0.011 min
Response: 25876856
Conc: 927.03 ng/ml




#29 AR-1254-4

R.T.: 7.280 min
Delta R.T.: -0.006 min
Response: 17556664
Conc: 856.05 ng/ml



#29 AR-1254-4

R.T.: 6.140 min
Delta R.T.: -0.010 min
Response: 17503208
Conc: 872.27 ng/ml




#30 AR-1254-5

R.T.: 7.699 min
Delta R.T.: -0.006 min
Response: 19428170
Conc: 1008.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM




#30 AR-1254-5

R.T.: 6.553 min
Delta R.T.: -0.012 min
Response: 21645061
Conc: 808.93 ng/ml




#31 AR-1260-1

R.T.: 7.156 min
Delta R.T.: 0.000 min
Response: 9941508
Conc: 538.81 ng/ml



#31 AR-1260-1

R.T.: 6.042 min
Delta R.T.: 0.000 min
Response: 14106098
Conc: 683.15 ng/ml




#32 AR-1260-2

R.T.: 7.415 min
Delta R.T.: 0.000 min
Response: 12770832
Conc: 532.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM




#32 AR-1260-2

R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 17292349
Conc: 616.72 ng/ml




#33 AR-1260-3

R.T.: 7.773 min
Delta R.T.: 0.000 min
Response: 6069932
Conc: 321.05 ng/ml



#33 AR-1260-3

R.T.: 6.379 min
Delta R.T.: -0.001 min
Response: 12800334
Conc: 525.27 ng/ml




#34 AR-1260-4

R.T.: 7.993 min
Delta R.T.: -0.005 min
Response: 8530153
Conc: 449.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
VW-CONCRETE-WALL


Manual Integrations
APPROVED

Sohil
10/10/2018 3:51:43 PM




#34 AR-1260-4

R.T.: 6.848 min
Delta R.T.: 0.000 min
Response: 4233513
Conc: 219.21 ng/ml



#35 AR-1260-5

R.T.: 8.316 min
Delta R.T.: 0.001 min
Response: 8313366
Conc: 196.51 ng/ml



#35 AR-1260-5

R.T.: 7.087 min
Delta R.T.: -0.002 min
Response: 12494981
Conc: 243.71 ng/ml