

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100919\
 Data File : P0061822.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2019 12:25
 Operator : HP/AJ
 Sample : K5147-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OK-03-100719

Manual Integrations
 APPROVED

Ankita
 10/10/2019 9:32:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 13:04:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.511	1114497	862443	27.173	26.668
2) SA Decachlor...	10.004	8.369	846034	547261	18.167	16.903m
Target Compounds						
26) L6 AR-1254-1	6.363	5.336	191864	92556	84.099m	42.199m#
27) L6 AR-1254-2	6.580	5.479	235281	123935	65.119m	63.216m
28) L6 AR-1254-3	6.952	5.872	546608	195829	139.892m	61.509m#
29) L6 AR-1254-4	7.231	6.097	118008	82138	38.964m	40.485m
30) L6 AR-1254-5	7.648	6.507	234371	279757	74.631m	98.908 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100919\
Data File : PO061822.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2019 12:25
Operator : HP/AJ
Sample : K5147-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
OK-03-100719

Manual Integrations
APPROVED

Ankita
10/10/2019 9:32:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 13:04:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

