

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100919\
 Data File : P0061813.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2019 9:55
 Operator : HP/AJ
 Sample : PB123723BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB123723BS

Manual Integrations
 APPROVED

Ankita
 10/10/2019 9:31:21 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 10:29:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.354	3.510	911585	614596	22.226	19.005
2)	SA Decachlor...	10.009	8.369	1005744	645056	21.596	19.924m
Target Compounds							
3)	L1 AR-1016-1	5.516	4.559	404887	270807	220.797m	201.394m
4)	L1 AR-1016-2	5.538	4.576	514049	345889	198.872m	186.628m
5)	L1 AR-1016-3	5.600	4.747	326546	199907	203.273	199.564
6)	L1 AR-1016-4	5.699	4.789	276156	171398	206.558	197.790
7)	L1 AR-1016-5	5.992	4.995	281795	221148	205.476	193.745
31)	L7 AR-1260-1	7.114	6.002	523869	405369	200.674	198.730
32)	L7 AR-1260-2	7.370	6.187	630623	484935	197.812	201.910
33)	L7 AR-1260-3	7.727	6.336	493833	463251	205.604	205.605
34)	L7 AR-1260-4	7.952	6.799	563126	349174	210.747	189.130m
35)	L7 AR-1260-5	8.266	7.039	1016143	762404	205.359	195.808

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100919\
Data File : PO061813.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2019 9:55
Operator : HP/AJ
Sample : PB123723BS
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 10:29:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.354 min
Delta R.T.: -0.011 min
Response: 911585
Conc: 22.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM




#1 Tetrachloro-m-xylene

R.T.: 3.510 min
Delta R.T.: -0.016 min
Response: 614596
Conc: 19.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.009 min
Delta R.T.: -0.016 min
Response: 1005744
Conc: 21.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.369 min
Delta R.T.: -0.024 min
Response: 645056
Conc: 19.92 ng/ml m




#3 AR-1016-1

R.T.: 5.516 min
Delta R.T.: -0.013 min
Response: 404887
Conc: 220.80 ng/ml m

Instrument :
ECD_O
ClientSampleId :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM




#3 AR-1016-1

R.T.: 4.559 min
Delta R.T.: -0.019 min
Response: 270807
Conc: 201.39 ng/ml m




#4 AR-1016-2

R.T.: 5.538 min
Delta R.T.: -0.012 min
Response: 514049
Conc: 198.87 ng/ml m



#4 AR-1016-2

R.T.: 4.576 min
Delta R.T.: -0.019 min
Response: 345889
Conc: 186.63 ng/ml m




#5 AR-1016-3

R.T.: 5.600 min
Delta R.T.: -0.013 min
Response: 326546
Conc: 203.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM




#5 AR-1016-3

R.T.: 4.747 min
Delta R.T.: -0.018 min
Response: 199907
Conc: 199.56 ng/ml




#6 AR-1016-4

R.T.: 5.699 min
Delta R.T.: -0.012 min
Response: 276156
Conc: 206.56 ng/ml



#6 AR-1016-4

R.T.: 4.789 min
Delta R.T.: -0.019 min
Response: 171398
Conc: 197.79 ng/ml




#7 AR-1016-5

R.T.: 5.992 min
Delta R.T.: -0.011 min
Response: 281795
Conc: 205.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM




#7 AR-1016-5

R.T.: 4.995 min
Delta R.T.: -0.019 min
Response: 221148
Conc: 193.75 ng/ml




#31 AR-1260-1

R.T.: 7.114 min
Delta R.T.: -0.011 min
Response: 523869
Conc: 200.67 ng/ml



#31 AR-1260-1

R.T.: 6.002 min
Delta R.T.: -0.020 min
Response: 405369
Conc: 198.73 ng/ml




#32 AR-1260-2

R.T.: 7.370 min
Delta R.T.: -0.011 min
Response: 630623
Conc: 197.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM




#32 AR-1260-2

R.T.: 6.187 min
Delta R.T.: -0.020 min
Response: 484935
Conc: 201.91 ng/ml




#33 AR-1260-3

R.T.: 7.727 min
Delta R.T.: -0.011 min
Response: 493833
Conc: 205.60 ng/ml



#33 AR-1260-3

R.T.: 6.336 min
Delta R.T.: -0.020 min
Response: 463251
Conc: 205.60 ng/ml




#34 AR-1260-4

R.T.: 7.952 min
Delta R.T.: -0.011 min
Response: 563126
Conc: 210.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123723BS


Manual Integrations
APPROVED

Ankita
10/10/2019 9:31:21 AM




#34 AR-1260-4

R.T.: 6.799 min
Delta R.T.: -0.021 min
Response: 349174
Conc: 189.13 ng/ml m



#35 AR-1260-5

R.T.: 8.266 min
Delta R.T.: -0.012 min
Response: 1016143
Conc: 205.36 ng/ml



#35 AR-1260-5

R.T.: 7.039 min
Delta R.T.: -0.020 min
Response: 762404
Conc: 195.81 ng/ml