

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100919\
 Data File : P0061831.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2019 16:25
 Operator : HP/AJ
 Sample : PB123783BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB123783BL

Manual Integrations
 APPROVED

Ankita
 10/10/2019 9:32:21 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 01:11:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.366	3.515	856911	646985	20.893	20.006
2) SA Decachlor...	10.030	8.378	986110	634230	21.174	19.590m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100919\
Data File : PO061831.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2019 16:25
Operator : HP/AJ
Sample : PB123783BL
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
PB123783BL


Manual Integrations
APPROVED

Ankita
10/10/2019 9:32:21 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 01:11:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.366 min
Delta R.T.: 0.003 min
Response: 856911
Conc: 20.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB123783BL


Manual Integrations
APPROVED

Ankita
10/10/2019 9:32:21 AM




#1 Tetrachloro-m-xylene

R.T.: 3.515 min
Delta R.T.: -0.004 min
Response: 646985
Conc: 20.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.030 min
Delta R.T.: -0.006 min
Response: 986110
Conc: 21.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.378 min
Delta R.T.: -0.031 min
Response: 634230
Conc: 19.59 ng/ml m