

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101419\
 Data File : PP030482.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Oct 2019 21:08
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 01:12:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100919.M
 Quant Title : 8080.M
 QLast Update : Thu Oct 10 09:35:34 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 T	Dalapon	682.500	669.570	1.9	99	0.00
2 T	3,5-DICHLORO BENZOIC ACID	697.500	717.771	-2.9	103	0.00
3 T	4-Nitrophenol	682.500	748.633	-9.7	108	0.00
4 S	2,4-DCAA	750.000	771.052	-2.8	104	0.00
5 T	DICAMBA	705.000	727.132	-3.1	103	0.00
6 T	MCP P	70.500	71.232	-1.0	100	0.00
7 T	MCPA	69.750	65.708	5.8	95	0.00
8 T	DICHLORPROP	705.000	710.112	-0.7	102	0.00
9 T	2,4-D	705.000	627.091	11.1	89	0.00
10 T	Pentachlorophenol	712.500	750.842	-5.4	104	0.00
11 T	2,4,5-TP (SILVEX)	712.500	716.274	-0.5	100	0.00
12 T	2,4,5-T	712.500	562.785	21.0#	78	0.00
13 T	2,4-DB	712.500	738.102	-3.6	102	0.00
14 T	DINOSEB	705.000	644.438	8.6	89	0.00
15 T	Picloram	712.500	483.453	32.1#	66	0.00
16 T	DCPA	720.000	726.748	-0.9	99	0.00

Signal #2

1 T	Dalapon	682.500	665.387	2.5	97	0.00
2 T	3,5-DICHLORO BENZOIC ACID	697.500	724.982	-3.9	102	0.00
3 T	4-Nitrophenol	682.500	802.570	-17.6#	113	0.00
4 S	2,4-DCAA	750.000	779.160	-3.9	103	0.00
5 T	DICAMBA	705.000	727.779	-3.2	102	0.00
6 T	MCP P	70.500	75.544	-7.2	101	0.00
7 T	MCPA	69.750	66.289	5.0	93	0.00
8 T	DICHLORPROP	705.000	706.463	-0.2	100	0.00
9 T	2,4-D	705.000	636.312	9.7	89	0.00
10 T	Pentachlorophenol	712.500	736.891	-3.4	102	0.00
11 T	2,4,5-TP (SILVEX)	712.500	708.424	0.6	98	0.00
12 T	2,4,5-T	712.500	590.916	17.1#	81	0.00
13 T	2,4-DB	712.500	766.996	-7.6	106	0.00
14 T	DINOSEB	705.000	651.966	7.5	90	0.00
15 T	Picloram	712.500	445.168	37.5#	60	0.00
16 T	DCPA	720.000	715.453	0.6	98	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101419\
Data File : PP030482.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Oct 2019 21:08
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 01:12:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100919.M
Quant Title : 8080.M
QLast Update : Thu Oct 10 09:35:34 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0