

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102419\
 Data File : P0063123.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Oct 2019 16:55
 Operator : HP/AJ
 Sample : K5500-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 802-00-(0-2)

Manual Integrations
 APPROVED

Ankita
 10/25/2019 10:52:48 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 06:08:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.314	3.476	936993	707354	15.150m	16.495
2)	SA Decachlor...	9.933	8.321	889581	474834	15.899m	16.183
Target Compounds							
31)	L7 AR-1260-1	7.065	5.958	43733	76264	15.360	35.893 #
32)	L7 AR-1260-2	7.316	6.143	206616	38337	60.228m	15.107m#
33)	L7 AR-1260-3	7.678	6.290	52468	26613	19.828	11.786m#
34)	L7 AR-1260-4	7.902	6.754	77043	30133	27.544m	17.162 #
35)	L7 AR-1260-5	8.214	6.994	83209	69674	15.893	19.341

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102419\
Data File : PO063123.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Oct 2019 16:55
Operator : HP/AJ
Sample : K5500-03
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
802-00-(0-2)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:48 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 06:08:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.314 min
Delta R.T.: -0.011 min
Response: 936993
Conc: 15.15 ng/ml m

Instrument :
ECD_O
ClientSampleId :
802-00-(0-2)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:48 AM




#1 Tetrachloro-m-xylene

R.T.: 3.476 min
Delta R.T.: -0.005 min
Response: 707354
Conc: 16.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.933 min
Delta R.T.: -0.024 min
Response: 889581
Conc: 15.90 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.321 min
Delta R.T.: -0.017 min
Response: 474834
Conc: 16.18 ng/ml




#31 AR-1260-1

R.T.: 7.065 min
Delta R.T.: -0.015 min
Response: 43733
Conc: 15.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
802-00-(0-2)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:48 AM




#31 AR-1260-1

R.T.: 5.958 min
Delta R.T.: -0.010 min
Response: 76264
Conc: 35.89 ng/ml




#32 AR-1260-2

R.T.: 7.316 min
Delta R.T.: -0.021 min
Response: 206616
Conc: 60.23 ng/ml m



#32 AR-1260-2

R.T.: 6.143 min
Delta R.T.: -0.011 min
Response: 38337
Conc: 15.11 ng/ml m




#33 AR-1260-3

R.T.: 7.678 min
Delta R.T.: -0.015 min
Response: 52468
Conc: 19.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
802-00-(0-2)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:48 AM




#33 AR-1260-3

R.T.: 6.290 min
Delta R.T.: -0.013 min
Response: 26613
Conc: 11.79 ng/ml m




#34 AR-1260-4

R.T.: 7.902 min
Delta R.T.: -0.016 min
Response: 77043
Conc: 27.54 ng/ml m



#34 AR-1260-4

R.T.: 6.754 min
Delta R.T.: -0.012 min
Response: 30133
Conc: 17.16 ng/ml




#35 AR-1260-5

R.T.: 8.214 min
Delta R.T.: -0.016 min
Response: 83209
Conc: 15.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
802-00-(0-2)

Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:48 AM



#35 AR-1260-5

R.T.: 6.994 min
Delta R.T.: -0.012 min
Response: 69674
Conc: 19.34 ng/ml