

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102419\
 Data File : P0063127.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Oct 2019 18:00
 Operator : HP/AJ
 Sample : K5499-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 194-84-(0-1)

Manual Integrations
 APPROVED

Ankita
 10/25/2019 10:52:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 06:10:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.313	3.476	908766	688473	14.694	16.055
2)	SA Decachlor...	9.932	8.320	754604	407146	13.486m	13.876
Target Compounds							
31)	L7 AR-1260-1	7.064	5.958	549793	445770	193.105	209.801
32)	L7 AR-1260-2	7.316	6.143	2019148	423264	588.578m	166.796 #
33)	L7 AR-1260-3	7.676	6.291	380191	377285	143.679	167.080m
34)	L7 AR-1260-4	7.899	6.753	537587	223464	192.198	127.268 #
35)	L7 AR-1260-5	8.213	6.994	582524	559985	111.262	155.446 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102419\
Data File : PO063127.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Oct 2019 18:00
Operator : HP/AJ
Sample : K5499-03
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
194-84-(0-1)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 06:10:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.313 min
Delta R.T.: -0.012 min
Response: 908766
Conc: 14.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
194-84-(0-1)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:58 AM




#1 Tetrachloro-m-xylene

R.T.: 3.476 min
Delta R.T.: -0.005 min
Response: 688473
Conc: 16.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.932 min
Delta R.T.: -0.024 min
Response: 754604
Conc: 13.49 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.320 min
Delta R.T.: -0.018 min
Response: 407146
Conc: 13.88 ng/ml




#31 AR-1260-1

R.T.: 7.064 min
Delta R.T.: -0.016 min
Response: 549793
Conc: 193.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
194-84-(0-1)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:58 AM




#31 AR-1260-1

R.T.: 5.958 min
Delta R.T.: -0.010 min
Response: 445770
Conc: 209.80 ng/ml




#32 AR-1260-2

R.T.: 7.316 min
Delta R.T.: -0.020 min
Response: 2019148
Conc: 588.58 ng/ml m



#32 AR-1260-2

R.T.: 6.143 min
Delta R.T.: -0.011 min
Response: 423264
Conc: 166.80 ng/ml




#33 AR-1260-3

R.T.: 7.676 min
Delta R.T.: -0.017 min
Response: 380191
Conc: 143.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
194-84-(0-1)


Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:58 AM




#33 AR-1260-3

R.T.: 6.291 min
Delta R.T.: -0.012 min
Response: 377285
Conc: 167.08 ng/ml m




#34 AR-1260-4

R.T.: 7.899 min
Delta R.T.: -0.020 min
Response: 537587
Conc: 192.20 ng/ml



#34 AR-1260-4

R.T.: 6.753 min
Delta R.T.: -0.012 min
Response: 223464
Conc: 127.27 ng/ml




#35 AR-1260-5

R.T.: 8.213 min
Delta R.T.: -0.017 min
Response: 582524
Conc: 111.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
194-84-(0-1)

Manual Integrations
APPROVED

Ankita
10/25/2019 10:52:58 AM



#35 AR-1260-5

R.T.: 6.994 min
Delta R.T.: -0.012 min
Response: 559985
Conc: 155.45 ng/ml