Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102519\
Data File : PO063188.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Oct 2019 23:39
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 00:41:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.317	3.480	3891215	2863767	62.916	66.783	
2)	SA Decachlor...	9.936	8.323	2373475	1380310	42.419	47.043	
	Target Compounds							
3)	L1 AR-1016-1	5.477	4.522	1104943	791773	466.871	506.202	
4)	L1 AR-1016-2	5.498	4.539	1621562	1192817	476.061	512.839	
5)	L1 AR-1016-3	5.560	4.708	958570	618134	463.207	500.084	
6)	L1 AR-1016-4	5.658	4.751	817490	483491	482.322	492.003	
7)	L1 AR-1016-5	5.949	4.955	646306	585491	363.438	454.025	
8)	L2 AR-1221-1	4.521	3.685	107393	87354	154.328	171.802	
9)	L2 AR-1221-2	4.612	3.768	161957	131178	310.555	341.389	
10)	L2 AR-1221-3	4.688	3.840	592043	446891	362.639	379.753	
11)	L3 AR-1232-1	4.688	3.840	592043	446891	271.070	288.407	
12)	L3 AR-1232-2	5.210	4.539	805415	1192817	675.061	690.904	
13)	L3 AR-1232-3	5.498	4.708	1621562	618134	633.858	676.271	
14)	L3 AR-1232-4	5.658	4.790	817490	579695	635.944	712.575	
15)	L3 AR-1232-5	5.746	4.955	705167	585491	712.489	662.947	
16)	L4 AR-1242-1	5.477	4.522	1104943	791773	596.299	658.355	
17)	L4 AR-1242-2	5.498	4.539	1621562	1192817	611.285	669.653	
18)	L4 AR-1242-3	5.560	4.708	958570	618134	586.954	649.278	
19)	L4 AR-1242-4	5.658	4.790	817490	579695	605.566	614.061	
20)	L4 AR-1242-5	6.389	5.296	113292	423812	74.123	355.167	#
21)	L5 AR-1248-1	5.477	4.522	1104943	791773	770.940	806.491	
22)	L5 AR-1248-2	5.746	4.751	705167	483491	352.205	381.633	
23)	L5 AR-1248-3	5.949	4.790	646306	579695	292.634	436.698	#
24)	L5 AR-1248-4	6.350	4.955	142469	585491	55.768	370.008	#
25)	L5 AR-1248-5	6.389	5.336	113292	76510	46.334	49.714	
26)	L6 AR-1254-1	6.323	5.296	527910	423812	198.044	173.282	
27)	L6 AR-1254-2	6.539	5.441	537935	357595	131.573	168.847	#
28)	L6 AR-1254-3	6.905	5.847	292520	666585	69.912	204.589	#
29)	L6 AR-1254-4	7.189	6.057	211320	95485	70.128	49.438	#
30)	L6 AR-1254-5	7.605	6.464	1576365	1031381	503.005	390.498	
31)	L7 AR-1260-1	7.066	5.959	1174525	842596	412.530	396.565	
32)	L7 AR-1260-2	7.322	6.145	1209427	1023356	352.546	403.273	
33)	L7 AR-1260-3	7.679	6.293	1171706	881998	442.802	390.592	
34)	L7 AR-1260-4	7.904	6.756	1102413	737998	394.135	420.308	
35)	L7 AR-1260-5	8.215	6.996	2442369	1737621	466.492	482.346	
36)	L8 AR-1262-1	7.679	6.555	1171706	449606	285.442	377.260	#
37)	L8 AR-1262-2	8.215	6.996	2442369	1737621	378.777	413.541	
38)	L8 AR-1262-3	8.529	7.273	1466661	426666	333.531	242.588	#
39)	L8 AR-1262-4	8.600	7.336	456901	1241083	132.245	405.965	#
40)	L8 AR-1262-5	9.229	7.827	732371	446126	324.246	341.593	
41)	L9 AR-1268-1	8.529	7.273	1466661	426666	193.747	91.110	#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102519\
Data File : PO063188.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Oct 2019 23:39
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 00:41:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.600	7.336	456901	1241083	65.041	297.069 #
43) L9	AR-1268-3	0.000	7.538	0	26350	N.D.	7.562 #
44) L9	AR-1268-4	9.229	7.827	732371	446126	308.237	313.100
45) L9	AR-1268-5	9.620	8.095	211647	125682	11.743	13.576

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102519\
Data File : PO063188.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Oct 2019 23:39
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 00:41:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.317 min
Delta R.T.: -0.008 min
Response: 3891215
Conc: 62.92 ng/ml

Instrument :
ECD_O
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: -0.002 min
Response: 2863767
Conc: 66.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.936 min
Delta R.T.: -0.021 min
Response: 2373475
Conc: 42.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.323 min
Delta R.T.: -0.015 min
Response: 1380310
Conc: 47.04 ng/ml



#3 AR-1016-1

R.T.: 5.477 min
Delta R.T.: -0.010 min
Response: 1104943
Conc: 466.87 ng/ml

Instrument :
ECD_O
ClientSampleId :



#3 AR-1016-1

R.T.: 4.522 min
Delta R.T.: -0.006 min
Response: 791773
Conc: 506.20 ng/ml



#4 AR-1016-2

R.T.: 5.498 min
Delta R.T.: -0.011 min
Response: 1621562
Conc: 476.06 ng/ml



#4 AR-1016-2

R.T.: 4.539 min
Delta R.T.: -0.006 min
Response: 1192817
Conc: 512.84 ng/ml



#5 AR-1016-3

R.T.: 5.560 min
Delta R.T.: -0.011 min
Response: 958570
Conc: 463.21 ng/ml

Instrument :
ECD_O
ClientSampleId :



#5 AR-1016-3

R.T.: 4.708 min
Delta R.T.: -0.006 min
Response: 618134
Conc: 500.08 ng/ml



#6 AR-1016-4

R.T.: 5.658 min
Delta R.T.: -0.010 min
Response: 817490
Conc: 482.32 ng/ml



#6 AR-1016-4

R.T.: 4.751 min
Delta R.T.: -0.006 min
Response: 483491
Conc: 492.00 ng/ml



#7 AR-1016-5

R.T.: 5.949 min
Delta R.T.: -0.012 min
Response: 646306
Conc: 363.44 ng/ml

Instrument :
ECD_O
ClientSampleId :



#7 AR-1016-5

R.T.: 4.955 min
Delta R.T.: -0.007 min
Response: 585491
Conc: 454.03 ng/ml



#8 AR-1221-1

R.T.: 4.521 min
Delta R.T.: -0.014 min
Response: 107393
Conc: 154.33 ng/ml



#8 AR-1221-1

R.T.: 3.685 min
Delta R.T.: -0.004 min
Response: 87354
Conc: 171.80 ng/ml



#9 AR-1221-2

R.T.: 4.612 min
Delta R.T.: -0.008 min
Response: 161957
Conc: 310.56 ng/ml

Instrument :
ECD_O
ClientSampleId :



#9 AR-1221-2

R.T.: 3.768 min
Delta R.T.: -0.003 min
Response: 131178
Conc: 341.39 ng/ml



#10 AR-1221-3

R.T.: 4.688 min
Delta R.T.: -0.008 min
Response: 592043
Conc: 362.64 ng/ml



#10 AR-1221-3

R.T.: 3.840 min
Delta R.T.: -0.005 min
Response: 446891
Conc: 379.75 ng/ml



#11 AR-1232-1

R.T.: 4.688 min
Delta R.T.: -0.008 min
Response: 592043
Conc: 271.07 ng/ml

Instrument :
ECD_O
ClientSampleId :



#11 AR-1232-1

R.T.: 3.840 min
Delta R.T.: -0.004 min
Response: 446891
Conc: 288.41 ng/ml



#12 AR-1232-2

R.T.: 5.210 min
Delta R.T.: -0.011 min
Response: 805415
Conc: 675.06 ng/ml



#12 AR-1232-2

R.T.: 4.539 min
Delta R.T.: -0.006 min
Response: 1192817
Conc: 690.90 ng/ml



#13 AR-1232-3

R.T.: 5.498 min
Delta R.T.: -0.011 min
Response: 1621562
Conc: 633.86 ng/ml

Instrument :
ECD_O
ClientSampleId :



#13 AR-1232-3

R.T.: 4.708 min
Delta R.T.: -0.006 min
Response: 618134
Conc: 676.27 ng/ml



#14 AR-1232-4

R.T.: 5.658 min
Delta R.T.: -0.011 min
Response: 817490
Conc: 635.94 ng/ml



#14 AR-1232-4

R.T.: 4.790 min
Delta R.T.: -0.007 min
Response: 579695
Conc: 712.57 ng/ml



#15 AR-1232-5

R.T.: 5.746 min
Delta R.T.: -0.012 min
Response: 705167
Conc: 712.49 ng/ml

Instrument :
ECD_O
ClientSampleId :



#15 AR-1232-5

R.T.: 4.955 min
Delta R.T.: -0.007 min
Response: 585491
Conc: 662.95 ng/ml



#16 AR-1242-1

R.T.: 5.477 min
Delta R.T.: -0.010 min
Response: 1104943
Conc: 596.30 ng/ml



#16 AR-1242-1

R.T.: 4.522 min
Delta R.T.: -0.007 min
Response: 791773
Conc: 658.35 ng/ml



#17 AR-1242-2

R.T.: 5.498 min
Delta R.T.: -0.011 min
Response: 1621562
Conc: 611.29 ng/ml

Instrument :
ECD_O
ClientSampleId :



#17 AR-1242-2

R.T.: 4.539 min
Delta R.T.: -0.006 min
Response: 1192817
Conc: 669.65 ng/ml



#18 AR-1242-3

R.T.: 5.560 min
Delta R.T.: -0.011 min
Response: 958570
Conc: 586.95 ng/ml



#18 AR-1242-3

R.T.: 4.708 min
Delta R.T.: -0.007 min
Response: 618134
Conc: 649.28 ng/ml



#19 AR-1242-4

R.T.: 5.658 min
Delta R.T.: -0.011 min
Response: 817490
Conc: 605.57 ng/ml

Instrument :
ECD_O
ClientSampleId :



#19 AR-1242-4

R.T.: 4.790 min
Delta R.T.: -0.007 min
Response: 579695
Conc: 614.06 ng/ml



#20 AR-1242-5

R.T.: 6.389 min
Delta R.T.: -0.012 min
Response: 113292
Conc: 74.12 ng/ml



#20 AR-1242-5

R.T.: 5.296 min
Delta R.T.: -0.008 min
Response: 423812
Conc: 355.17 ng/ml



#21 AR-1248-1

R.T.: 5.477 min
Delta R.T.: -0.010 min
Response: 1104943
Conc: 770.94 ng/ml

Instrument :
ECD_O
ClientSampleId :



#21 AR-1248-1

R.T.: 4.522 min
Delta R.T.: -0.006 min
Response: 791773
Conc: 806.49 ng/ml



#22 AR-1248-2

R.T.: 5.746 min
Delta R.T.: -0.012 min
Response: 705167
Conc: 352.21 ng/ml



#22 AR-1248-2

R.T.: 4.751 min
Delta R.T.: -0.006 min
Response: 483491
Conc: 381.63 ng/ml



#23 AR-1248-3

R.T.: 5.949 min
Delta R.T.: -0.011 min
Response: 646306
Conc: 292.63 ng/ml

Instrument :
ECD_O
ClientSampleId :



#23 AR-1248-3

R.T.: 4.790 min
Delta R.T.: -0.006 min
Response: 579695
Conc: 436.70 ng/ml



#24 AR-1248-4

R.T.: 6.350 min
Delta R.T.: -0.013 min
Response: 142469
Conc: 55.77 ng/ml



#24 AR-1248-4

R.T.: 4.955 min
Delta R.T.: -0.007 min
Response: 585491
Conc: 370.01 ng/ml



#25 AR-1248-5

R.T.: 6.389 min
Delta R.T.: -0.012 min
Response: 113292
Conc: 46.33 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 5.336 min
Delta R.T.: -0.008 min
Response: 76510
Conc: 49.71 ng/ml



#26 AR-1254-1

R.T.: 6.323 min
Delta R.T.: -0.013 min
Response: 527910
Conc: 198.04 ng/ml



#26 AR-1254-1

R.T.: 5.296 min
Delta R.T.: -0.009 min
Response: 423812
Conc: 173.28 ng/ml



#27 AR-1254-2

R.T.: 6.539 min
Delta R.T.: -0.014 min
Response: 537935
Conc: 131.57 ng/ml

Instrument :
ECD_O
ClientSampleId :



#27 AR-1254-2

R.T.: 5.441 min
Delta R.T.: -0.008 min
Response: 357595
Conc: 168.85 ng/ml



#28 AR-1254-3

R.T.: 6.905 min
Delta R.T.: -0.013 min
Response: 292520
Conc: 69.91 ng/ml



#28 AR-1254-3

R.T.: 5.847 min
Delta R.T.: 0.005 min
Response: 666585
Conc: 204.59 ng/ml



#29 AR-1254-4

R.T.: 7.189 min
Delta R.T.: -0.014 min
Response: 211320
Conc: 70.13 ng/ml

Instrument :
ECD_O
ClientSampleId :



#29 AR-1254-4

R.T.: 6.057 min
Delta R.T.: -0.009 min
Response: 95485
Conc: 49.44 ng/ml



#30 AR-1254-5

R.T.: 7.605 min
Delta R.T.: -0.015 min
Response: 1576365
Conc: 503.00 ng/ml



#30 AR-1254-5

R.T.: 6.464 min
Delta R.T.: -0.011 min
Response: 1031381
Conc: 390.50 ng/ml



#31 AR-1260-1

R.T.: 7.066 min
Delta R.T.: -0.014 min
Response: 1174525
Conc: 412.53 ng/ml

Instrument :
ECD_O
ClientSampleId :



#31 AR-1260-1

R.T.: 5.959 min
Delta R.T.: -0.009 min
Response: 842596
Conc: 396.57 ng/ml



#32 AR-1260-2

R.T.: 7.322 min
Delta R.T.: -0.015 min
Response: 1209427
Conc: 352.55 ng/ml



#32 AR-1260-2

R.T.: 6.145 min
Delta R.T.: -0.009 min
Response: 1023356
Conc: 403.27 ng/ml



#33 AR-1260-3

R.T.: 7.679 min
Delta R.T.: -0.015 min
Response: 1171706
Conc: 442.80 ng/ml

Instrument :
ECD_O
ClientSampleId :



#33 AR-1260-3

R.T.: 6.293 min
Delta R.T.: -0.010 min
Response: 881998
Conc: 390.59 ng/ml



#34 AR-1260-4

R.T.: 7.904 min
Delta R.T.: -0.014 min
Response: 1102413
Conc: 394.13 ng/ml



#34 AR-1260-4

R.T.: 6.756 min
Delta R.T.: -0.010 min
Response: 737998
Conc: 420.31 ng/ml



#35 AR-1260-5

R.T.: 8.215 min
Delta R.T.: -0.015 min
Response: 2442369
Conc: 466.49 ng/ml

Instrument :
ECD_O
ClientSampleId :



#35 AR-1260-5

R.T.: 6.996 min
Delta R.T.: -0.010 min
Response: 1737621
Conc: 482.35 ng/ml



#36 AR-1262-1

R.T.: 7.679 min
Delta R.T.: -0.014 min
Response: 1171706
Conc: 285.44 ng/ml



#36 AR-1262-1

R.T.: 6.555 min
Delta R.T.: -0.010 min
Response: 449606
Conc: 377.26 ng/ml



#37 AR-1262-2

R.T.: 8.215 min
Delta R.T.: -0.014 min
Response: 2442369
Conc: 378.78 ng/ml

Instrument :
ECD_O
ClientSampleId :



#37 AR-1262-2

R.T.: 6.996 min
Delta R.T.: -0.011 min
Response: 1737621
Conc: 413.54 ng/ml



#38 AR-1262-3

R.T.: 8.529 min
Delta R.T.: 0.000 min
Response: 1466661
Conc: 333.53 ng/ml



#38 AR-1262-3

R.T.: 7.273 min
Delta R.T.: -0.012 min
Response: 426666
Conc: 242.59 ng/ml



#39 AR-1262-4

R.T.: 8.600 min
Delta R.T.: -0.016 min
Response: 456901
Conc: 132.24 ng/ml

Instrument :
ECD_O
ClientSampleId :



#39 AR-1262-4

R.T.: 7.336 min
Delta R.T.: -0.013 min
Response: 1241083
Conc: 405.97 ng/ml



#40 AR-1262-5

R.T.: 9.229 min
Delta R.T.: -0.016 min
Response: 732371
Conc: 324.25 ng/ml



#40 AR-1262-5

R.T.: 7.827 min
Delta R.T.: -0.012 min
Response: 446126
Conc: 341.59 ng/ml



#41 AR-1268-1

R.T.: 8.529 min
Delta R.T.: 0.002 min
Response: 1466661
Conc: 193.75 ng/ml

Instrument :
ECD_O
ClientSampleId :



#41 AR-1268-1

R.T.: 7.273 min
Delta R.T.: -0.012 min
Response: 426666
Conc: 91.11 ng/ml



#42 AR-1268-2

R.T.: 8.600 min
Delta R.T.: -0.018 min
Response: 456901
Conc: 65.04 ng/ml



#42 AR-1268-2

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 1241083
Conc: 297.07 ng/ml



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.834 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#43 AR-1268-3

R.T.: 7.538 min
Delta R.T.: -0.013 min
Response: 26350
Conc: 7.56 ng/ml



#44 AR-1268-4

R.T.: 9.229 min
Delta R.T.: -0.016 min
Response: 732371
Conc: 308.24 ng/ml



#44 AR-1268-4

R.T.: 7.827 min
Delta R.T.: -0.011 min
Response: 446126
Conc: 313.10 ng/ml



#45 AR-1268-5

R.T.: 9.620 min
Delta R.T.: -0.018 min
Response: 211647
Conc: 11.74 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.095 min
Delta R.T.: -0.013 min
Response: 125682
Conc: 13.58 ng/ml