

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102519\
 Data File : P0063185.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Oct 2019 20:38
 Operator : HP/AJ
 Sample : K5499-01RE
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 193-26-(0-1)RE

Manual Integrations
 APPROVED

Ankita
 10/28/2019 10:52:16 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 23:22:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.317	3.480	1274294	1076086	20.604m	25.094
2)	SA Decachlor...	9.936	8.325	1128431	818406	20.168m	27.893 #
Target Compounds							
31)	L7 AR-1260-1	7.067	5.959	453346	401700	159.229	189.059
32)	L7 AR-1260-2	7.319	6.145	2155101	306941	628.208m	120.956 #
33)	L7 AR-1260-3	7.680	6.293	373807	319722	141.266	141.589
34)	L7 AR-1260-4	7.902	6.756	418431	212763	149.597	121.174
35)	L7 AR-1260-5	8.216	6.996	655766	556790	125.251m	154.559

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102519\
Data File : PO063185.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Oct 2019 20:38
Operator : HP/AJ
Sample : K5499-01RE
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
193-26-(0-1)RE


Manual Integrations
APPROVED

Ankita
10/28/2019 10:52:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 23:22:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.317 min
Delta R.T.: -0.008 min
Response: 1274294
Conc: 20.60 ng/ml m

Instrument :
ECD_O
ClientSampleId :
193-26-(0-1)RE


Manual Integrations
APPROVED

Ankita
10/28/2019 10:52:16 AM




#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: -0.001 min
Response: 1076086
Conc: 25.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.936 min
Delta R.T.: -0.020 min
Response: 1128431
Conc: 20.17 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.325 min
Delta R.T.: -0.013 min
Response: 818406
Conc: 27.89 ng/ml




#31 AR-1260-1

R.T.: 7.067 min
Delta R.T.: -0.013 min
Response: 453346
Conc: 159.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
193-26-(0-1)RE


Manual Integrations
APPROVED

Ankita
10/28/2019 10:52:16 AM




#31 AR-1260-1

R.T.: 5.959 min
Delta R.T.: -0.009 min
Response: 401700
Conc: 189.06 ng/ml




#32 AR-1260-2

R.T.: 7.319 min
Delta R.T.: -0.018 min
Response: 2155101
Conc: 628.21 ng/ml m



#32 AR-1260-2

R.T.: 6.145 min
Delta R.T.: -0.009 min
Response: 306941
Conc: 120.96 ng/ml




#33 AR-1260-3

R.T.: 7.680 min
Delta R.T.: -0.014 min
Response: 373807
Conc: 141.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
193-26-(0-1)RE


Manual Integrations
APPROVED

Ankita
10/28/2019 10:52:16 AM




#33 AR-1260-3

R.T.: 6.293 min
Delta R.T.: -0.010 min
Response: 319722
Conc: 141.59 ng/ml




#34 AR-1260-4

R.T.: 7.902 min
Delta R.T.: -0.016 min
Response: 418431
Conc: 149.60 ng/ml



#34 AR-1260-4

R.T.: 6.756 min
Delta R.T.: -0.010 min
Response: 212763
Conc: 121.17 ng/ml




#35 AR-1260-5

R.T.: 8.216 min
Delta R.T.: -0.014 min
Response: 655766
Conc: 125.25 ng/ml m

Instrument :
ECD_O
ClientSampleId :
193-26-(0-1)RE

Manual Integrations
APPROVED

Ankita
10/28/2019 10:52:16 AM



#35 AR-1260-5

R.T.: 6.996 min
Delta R.T.: -0.010 min
Response: 556790
Conc: 154.56 ng/ml