

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102919\
Data File : P0063234.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Oct 2019 16:21
Operator : HP/AJ
Sample : PB124375BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB124375BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 09:05:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.318	3.479	901001	874025	15.460	16.004
2)	SA Decachlor...	9.937	8.322	1004992	723844	17.272	16.931

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102919\
Data File : P0063234.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Oct 2019 16:21
Operator : HP/AJ
Sample : PB124375BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB124375BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 09:05:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.318 min
Delta R.T.: 0.002 min
Response: 901001
Conc: 15.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124375BL




#1 Tetrachloro-m-xylene

R.T.: 3.479 min
Delta R.T.: 0.000 min
Response: 874025
Conc: 16.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.937 min
Delta R.T.: 0.004 min
Response: 1004992
Conc: 17.27 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.322 min
Delta R.T.: 0.001 min
Response: 723844
Conc: 16.93 ng/ml