

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0103119\
 Data File : P0063373.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 31 Oct 2019 19:55
 Operator : HP/AJ
 Sample : K5620-03
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-25-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 01:18:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.316	3.481	1358514	1358694	23.310	24.878
2) SA Decachlor...	9.936	8.320	946461	627794	16.266	14.684
Target Compounds						
38) L8 AR-1262-3	0.000	7.209	0	16339	N.D.	6.096 #
40) L8 AR-1262-5	0.000	7.786	0	32879	N.D.	16.755 #
41) L9 AR-1268-1	0.000	7.209	0	16339	N.D.	2.050 #
44) L9 AR-1268-4	0.000	7.786	0	32879	N.D.	15.233 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103119\
Data File : P0063373.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 31 Oct 2019 19:55
Operator : HP/AJ
Sample : K5620-03
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-25-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 01:18:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: 0.000 min
Response: 1358514
Conc: 23.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-25-S




#1 Tetrachloro-m-xylene

R.T.: 3.481 min
Delta R.T.: 0.003 min
Response: 1358694
Conc: 24.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.936 min
Delta R.T.: 0.002 min
Response: 946461
Conc: 16.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.320 min
Delta R.T.: 0.000 min
Response: 627794
Conc: 14.68 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 8.512 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
TP-25-S




#38 AR-1262-3

R.T.: 7.209 min
Delta R.T.: -0.063 min
Response: 16339
Conc: 6.10 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 9.228 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.786 min
Delta R.T.: -0.039 min
Response: 32879
Conc: 16.75 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.512 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
TP-25-S




#41 AR-1268-1

R.T.: 7.209 min
Delta R.T.: -0.063 min
Response: 16339
Conc: 2.05 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.230 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 7.786 min
Delta R.T.: -0.039 min
Response: 32879
Conc: 15.23 ng/ml