

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103119\
 Data File : PO063389.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Nov 2019 1:10
 Operator : HP/AJ
 Sample : K5625-03 5X
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 03-A-03-B-03-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 01:29:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.316	3.481	279674	303477	4.799	5.557
2)	SA Decachlor...	9.935	8.326	98551	438532	1.694	10.258 #
Target Compounds							
3)	L1 AR-1016-1	5.499	4.522	72060	66669	32.962	35.585
4)	L1 AR-1016-2	5.499	4.543	72060	41192	23.002	15.102 #
5)	L1 AR-1016-3	5.499	4.728	72060	257879	38.463	185.808 #
6)	L1 AR-1016-4	5.696	4.753	-62557	335806	N.D.	300.667 #
7)	L1 AR-1016-5	5.951	4.968	148584	460857	97.535	305.909 #
8)	L2 AR-1221-1	0.000	3.710	0	31779	N.D.	61.460 #
9)	L2 AR-1221-2	0.000	3.710	0	31779	N.D.	82.310 #
10)	L2 AR-1221-3	0.000	3.840	0	37246	N.D.	30.291 #
11)	L3 AR-1232-1	0.000	3.840	0	37246	N.D.	21.822 #
12)	L3 AR-1232-2	5.171	4.543	133464	41192	134.824	20.941 #
13)	L3 AR-1232-3	5.499	4.728	72060	257879	31.822	266.558 #
14)	L3 AR-1232-4	5.696	4.780	-62557	370233	N.D.	445.910 #
15)	L3 AR-1232-5	5.760	4.968	124548	460857	155.323	492.391 #
16)	L4 AR-1242-1	5.499	4.522	72060	66669	44.925	48.245
17)	L4 AR-1242-2	5.499	4.543	72060	41192	31.254	20.077 #
18)	L4 AR-1242-3	5.499	4.728	72060	257879	51.863	246.070 #
19)	L4 AR-1242-4	5.696	4.780	-62557	370233	N.D.	372.986 #
20)	L4 AR-1242-5	6.400	5.296	4016466	1251827	2790.582	896.064 #
21)	L5 AR-1248-1	5.499	4.522	72060	66669	58.446	60.996
22)	L5 AR-1248-2	5.744	4.753	349276	335806	201.633	240.136
23)	L5 AR-1248-3	5.951	4.780	148584	370233	76.646	256.005 #
24)	L5 AR-1248-4	6.340	4.968	7186925	460857	2887.519	255.932 #
25)	L5 AR-1248-5	6.400	5.330	4016466	216438	1720.451	117.085 #
26)	L6 AR-1254-1	6.340	5.296	7186925	1251827	2826.810	415.573 #
27)	L6 AR-1254-2	6.541	5.441	2328181	965862	575.568	358.124 #
28)	L6 AR-1254-3	6.907	5.829	1830730	5204554	430.111	1212.276 #
29)	L6 AR-1254-4	7.182	6.047	290359	534427	85.783	187.541 #
30)	L6 AR-1254-5	7.608	6.470	1865418	1456710	562.380	390.786 #
31)	L7 AR-1260-1	7.069	5.951	6767333	315298	2182.202	114.162 #
32)	L7 AR-1260-2	7.352	6.149	22310356	3659826	5526.537	953.856 #
33)	L7 AR-1260-3	7.670	6.283	3689024	4992078	1148.700	1591.032 #
34)	L7 AR-1260-4	7.898	6.752	21788527	13708080	6825.209	5414.143
35)	L7 AR-1260-5	8.200	7.001	909507	555500	155.079	91.134 #
36)	L8 AR-1262-1	7.670	6.555	3689024	6121830	772.816	3364.441 #
37)	L8 AR-1262-2	8.200	7.001	909507	555500	130.453	77.942 #
38)	L8 AR-1262-3	8.517	7.263	3316426	293961	717.550	109.681 #
39)	L8 AR-1262-4	8.588	7.331	6118545	2657563	1699.369	579.506 #
40)	L8 AR-1262-5	9.218	7.780	31941972	728601	13956.110	371.292 #
41)	L9 AR-1268-1	8.517	7.263	3316426	293961	395.994	36.890 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103119\
 Data File : PO063389.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Nov 2019 1:10
 Operator : HP/AJ
 Sample : K5625-03 5X
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 03-A-03-B-03-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 01:29:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.588	7.342	6118545	3941302	835.075	598.572 #
43)	L9 AR-1268-3	8.800	7.541	1502033	1711306	238.907	306.237 #
44)	L9 AR-1268-4	9.218	7.780	31941972	728601	12977.588	337.574 #
45)	L9 AR-1268-5	9.625	8.103	1023753	1660596	57.011	110.911 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103119\
Data File : PO063389.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Nov 2019 1:10
Operator : HP/AJ
Sample : K5625-03 5X
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 01:29:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: 0.000 min
Response: 279674
Conc: 4.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#1 Tetrachloro-m-xylene

R.T.: 3.481 min
Delta R.T.: 0.003 min
Response: 303477
Conc: 5.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.935 min
Delta R.T.: 0.001 min
Response: 98551
Conc: 1.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.326 min
Delta R.T.: 0.005 min
Response: 438532
Conc: 10.26 ng/ml



#3 AR-1016-1


R.T.: 5.499 min
Delta R.T.: 0.022 min
Response: 72060
Conc: 32.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#3 AR-1016-1

R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 66669
Conc: 35.59 ng/ml




#4 AR-1016-2

R.T.: 5.499 min
Delta R.T.: 0.000 min
Response: 72060
Conc: 23.00 ng/ml



#4 AR-1016-2


R.T.: 4.543 min
Delta R.T.: 0.004 min
Response: 41192
Conc: 15.10 ng/ml



#5 AR-1016-3


R.T.: 5.499 min
Delta R.T.: -0.061 min
Response: 72060
Conc: 38.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#5 AR-1016-3

R.T.: 4.728 min
Delta R.T.: 0.020 min
Response: 257879
Conc: 185.81 ng/ml




#6 AR-1016-4

R.T.: 5.696 min
Delta R.T.: 0.038 min
Response: -62557
Conc: N.D.



#6 AR-1016-4


R.T.: 4.753 min
Delta R.T.: 0.002 min
Response: 335806
Conc: 300.67 ng/ml



#7 AR-1016-5


R.T.: 5.951 min
Delta R.T.: 0.002 min
Response: 148584
Conc: 97.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#7 AR-1016-5

R.T.: 4.968 min
Delta R.T.: 0.013 min
Response: 460857
Conc: 305.91 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.526 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.710 min
Delta R.T.: 0.025 min
Response: 31779
Conc: 61.46 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.611 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#9 AR-1221-2

R.T.: 3.710 min
Delta R.T.: -0.057 min
Response: 31779
Conc: 82.31 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T.: 4.687 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 3.840 min
Delta R.T.: 0.000 min
Response: 37246
Conc: 30.29 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 4.693 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#11 AR-1232-1

R.T.: 3.840 min
Delta R.T.: 0.000 min
Response: 37246
Conc: 21.82 ng/ml




#12 AR-1232-2

R.T.: 5.171 min
Delta R.T.: -0.045 min
Response: 133464
Conc: 134.82 ng/ml



#12 AR-1232-2


R.T.: 4.543 min
Delta R.T.: 0.003 min
Response: 41192
Conc: 20.94 ng/ml



#13 AR-1232-3


R.T.: 5.499 min
Delta R.T.: -0.006 min
Response: 72060
Conc: 31.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#13 AR-1232-3

R.T.: 4.728 min
Delta R.T.: 0.018 min
Response: 257879
Conc: 266.56 ng/ml




#14 AR-1232-4

R.T.: 5.696 min
Delta R.T.: 0.032 min
Response: -62557
Conc: N.D.



#14 AR-1232-4


R.T.: 4.780 min
Delta R.T.: -0.012 min
Response: 370233
Conc: 445.91 ng/ml



#15 AR-1232-5


R.T.: 5.760 min
Delta R.T.: 0.007 min
Response: 124548
Conc: 155.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#15 AR-1232-5

R.T.: 4.968 min
Delta R.T.: 0.011 min
Response: 460857
Conc: 492.39 ng/ml




#16 AR-1242-1

R.T.: 5.499 min
Delta R.T.: 0.022 min
Response: 72060
Conc: 44.93 ng/ml



#16 AR-1242-1


R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 66669
Conc: 48.25 ng/ml



#17 AR-1242-2


R.T.: 5.499 min
Delta R.T.: 0.000 min
Response: 72060
Conc: 31.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#17 AR-1242-2

R.T.: 4.543 min
Delta R.T.: 0.004 min
Response: 41192
Conc: 20.08 ng/ml




#18 AR-1242-3

R.T.: 5.499 min
Delta R.T.: -0.061 min
Response: 72060
Conc: 51.86 ng/ml



#18 AR-1242-3


R.T.: 4.728 min
Delta R.T.: 0.019 min
Response: 257879
Conc: 246.07 ng/ml



#19 AR-1242-4


R.T.: 5.696 min
Delta R.T.: 0.038 min
Response: -62557
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#19 AR-1242-4

R.T.: 4.780 min
Delta R.T.: -0.010 min
Response: 370233
Conc: 372.99 ng/ml




#20 AR-1242-5

R.T.: 6.400 min
Delta R.T.: 0.011 min
Response: 4016466
Conc: 2790.58 ng/ml



#20 AR-1242-5


R.T.: 5.296 min
Delta R.T.: -0.001 min
Response: 1251827
Conc: 896.06 ng/ml



#21 AR-1248-1


R.T.: 5.499 min
Delta R.T.: 0.023 min
Response: 72060
Conc: 58.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#21 AR-1248-1

R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 66669
Conc: 61.00 ng/ml




#22 AR-1248-2

R.T.: 5.744 min
Delta R.T.: -0.003 min
Response: 349276
Conc: 201.63 ng/ml



#22 AR-1248-2


R.T.: 4.753 min
Delta R.T.: 0.002 min
Response: 335806
Conc: 240.14 ng/ml



#23 AR-1248-3


R.T.: 5.951 min
Delta R.T.: 0.002 min
Response: 148584
Conc: 76.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#23 AR-1248-3

R.T.: 4.780 min
Delta R.T.: -0.010 min
Response: 370233
Conc: 256.00 ng/ml




#24 AR-1248-4

R.T.: 6.340 min
Delta R.T.: -0.011 min
Response: 7186925
Conc: 2887.52 ng/ml



#24 AR-1248-4


R.T.: 4.968 min
Delta R.T.: 0.013 min
Response: 460857
Conc: 255.93 ng/ml



#25 AR-1248-5


R.T.: 6.400 min
Delta R.T.: 0.011 min
Response: 4016466
Conc: 1720.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#25 AR-1248-5

R.T.: 5.330 min
Delta R.T.: -0.006 min
Response: 216438
Conc: 117.09 ng/ml




#26 AR-1254-1

R.T.: 6.340 min
Delta R.T.: 0.017 min
Response: 7186925
Conc: 2826.81 ng/ml



#26 AR-1254-1


R.T.: 5.296 min
Delta R.T.: 0.000 min
Response: 1251827
Conc: 415.57 ng/ml



#27 AR-1254-2


R.T.: 6.541 min
Delta R.T.: 0.002 min
Response: 2328181
Conc: 575.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#27 AR-1254-2

R.T.: 5.441 min
Delta R.T.: 0.000 min
Response: 965862
Conc: 358.12 ng/ml




#28 AR-1254-3

R.T.: 6.907 min
Delta R.T.: 0.003 min
Response: 1830730
Conc: 430.11 ng/ml



#28 AR-1254-3


R.T.: 5.829 min
Delta R.T.: -0.003 min
Response: 5204554
Conc: 1212.28 ng/ml



#29 AR-1254-4


R.T.: 7.182 min
Delta R.T.: -0.007 min
Response: 290359
Conc: 85.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#29 AR-1254-4

R.T.: 6.047 min
Delta R.T.: -0.009 min
Response: 534427
Conc: 187.54 ng/ml




#30 AR-1254-5

R.T.: 7.608 min
Delta R.T.: 0.003 min
Response: 1865418
Conc: 562.38 ng/ml



#30 AR-1254-5


R.T.: 6.470 min
Delta R.T.: 0.005 min
Response: 1456710
Conc: 390.79 ng/ml



#31 AR-1260-1


R.T.: 7.069 min
Delta R.T.: 0.003 min
Response: 6767333
Conc: 2182.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#31 AR-1260-1

R.T.: 5.951 min
Delta R.T.: -0.007 min
Response: 315298
Conc: 114.16 ng/ml




#32 AR-1260-2

R.T.: 7.352 min
Delta R.T.: 0.030 min
Response: 22310356
Conc: 5526.54 ng/ml



#32 AR-1260-2


R.T.: 6.149 min
Delta R.T.: 0.005 min
Response: 3659826
Conc: 953.86 ng/ml



#33 AR-1260-3


R.T.: 7.670 min
Delta R.T.: -0.008 min
Response: 3689024
Conc: 1148.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#33 AR-1260-3

R.T.: 6.283 min
Delta R.T.: -0.010 min
Response: 4992078
Conc: 1591.03 ng/ml




#34 AR-1260-4

R.T.: 7.898 min
Delta R.T.: -0.006 min
Response: 21788527
Conc: 6825.21 ng/ml



#34 AR-1260-4


R.T.: 6.752 min
Delta R.T.: -0.003 min
Response: 13708080
Conc: 5414.14 ng/ml



#35 AR-1260-5


R.T.: 8.200 min
Delta R.T.: -0.014 min
Response: 909507
Conc: 155.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#35 AR-1260-5

R.T.: 7.001 min
Delta R.T.: 0.006 min
Response: 555500
Conc: 91.13 ng/ml




#36 AR-1262-1

R.T.: 7.670 min
Delta R.T.: -0.008 min
Response: 3689024
Conc: 772.82 ng/ml



#36 AR-1262-1


R.T.: 6.555 min
Delta R.T.: 0.001 min
Response: 6121830
Conc: 3364.44 ng/ml



#37 AR-1262-2


R.T.: 8.200 min
Delta R.T.: -0.014 min
Response: 909507
Conc: 130.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#37 AR-1262-2

R.T.: 7.001 min
Delta R.T.: 0.006 min
Response: 555500
Conc: 77.94 ng/ml




#38 AR-1262-3

R.T.: 8.517 min
Delta R.T.: 0.005 min
Response: 3316426
Conc: 717.55 ng/ml



#38 AR-1262-3


R.T.: 7.263 min
Delta R.T.: -0.009 min
Response: 293961
Conc: 109.68 ng/ml



#39 AR-1262-4


R.T.: 8.588 min
Delta R.T.: -0.011 min
Response: 6118545
Conc: 1699.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#39 AR-1262-4

R.T.: 7.331 min
Delta R.T.: -0.005 min
Response: 2657563
Conc: 579.51 ng/ml




#40 AR-1262-5

R.T.: 9.218 min
Delta R.T.: -0.010 min
Response: 31941972
Conc: 13956.11 ng/ml



#40 AR-1262-5


R.T.: 7.780 min
Delta R.T.: -0.045 min
Response: 728601
Conc: 371.29 ng/ml



#41 AR-1268-1


R.T.: 8.517 min
Delta R.T.: 0.005 min
Response: 3316426
Conc: 395.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#41 AR-1268-1

R.T.: 7.263 min
Delta R.T.: -0.009 min
Response: 293961
Conc: 36.89 ng/ml




#42 AR-1268-2

R.T.: 8.588 min
Delta R.T.: -0.015 min
Response: 6118545
Conc: 835.08 ng/ml



#42 AR-1268-2


R.T.: 7.342 min
Delta R.T.: 0.005 min
Response: 3941302
Conc: 598.57 ng/ml



#43 AR-1268-3


R.T.: 8.800 min
Delta R.T.: -0.018 min
Response: 1502033
Conc: 238.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C




#43 AR-1268-3

R.T.: 7.541 min
Delta R.T.: 0.004 min
Response: 1711306
Conc: 306.24 ng/ml




#44 AR-1268-4

R.T.: 9.218 min
Delta R.T.: -0.012 min
Response: 31941972
Conc: 12977.59 ng/ml



#44 AR-1268-4


R.T.: 7.780 min
Delta R.T.: -0.045 min
Response: 728601
Conc: 337.57 ng/ml



#45 AR-1268-5

R.T.: 9.625 min
Delta R.T.: 0.006 min
Response: 1023753
Conc: 57.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
03-A-03-B-03-C



#45 AR-1268-5

R.T.: 8.103 min
Delta R.T.: 0.010 min
Response: 1660596
Conc: 110.91 ng/ml