

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0103119\
 Data File : P0063363.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 31 Oct 2019 17:09
 Operator : HP/AJ
 Sample : PB124471BS
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB124471BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 01:14:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.316	3.480	1481936	1441889	25.427	26.402
2) SA Decachlor...	9.935	8.321	1283915	913047	22.066	21.357
Target Compounds						
3) L1 AR-1016-1	5.474	4.522	469222	366359	214.632	195.548
4) L1 AR-1016-2	5.497	4.538	693959	669252	221.514	245.369
5) L1 AR-1016-3	5.559	4.708	422468	298041	225.500	214.746
6) L1 AR-1016-4	5.656	4.750	338804	244122	214.557	218.577
7) L1 AR-1016-5	5.947	4.955	337941	299525	221.835	198.819
31) L7 AR-1260-1	7.064	5.958	567419	531094	182.970	192.296
32) L7 AR-1260-2	7.320	6.144	729640	774460	180.740	201.847
33) L7 AR-1260-3	7.678	6.292	602198	630484	187.514	200.942
34) L7 AR-1260-4	7.903	6.754	563668	565354	176.568	223.292 #
35) L7 AR-1260-5	8.214	6.995	1153077	1348543	196.609	221.239



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103119\
Data File : PO063363.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 31 Oct 2019 17:09
Operator : HP/AJ
Sample : PB124471BS
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
PB124471BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 01:14:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: 0.000 min
Response: 1481936
Conc: 25.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124471BS




#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: 0.002 min
Response: 1441889
Conc: 26.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.935 min
Delta R.T.: 0.002 min
Response: 1283915
Conc: 22.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.321 min
Delta R.T.: 0.000 min
Response: 913047
Conc: 21.36 ng/ml



#3 AR-1016-1


R.T.: 5.474 min
Delta R.T.: -0.002 min
Response: 469222
Conc: 214.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124471BS




#3 AR-1016-1

R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 366359
Conc: 195.55 ng/ml




#4 AR-1016-2

R.T.: 5.497 min
Delta R.T.: -0.002 min
Response: 693959
Conc: 221.51 ng/ml



#4 AR-1016-2


R.T.: 4.538 min
Delta R.T.: 0.000 min
Response: 669252
Conc: 245.37 ng/ml



#5 AR-1016-3


R.T.: 5.559 min
Delta R.T.: -0.001 min
Response: 422468
Conc: 225.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124471BS




#5 AR-1016-3

R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 298041
Conc: 214.75 ng/ml




#6 AR-1016-4

R.T.: 5.656 min
Delta R.T.: -0.001 min
Response: 338804
Conc: 214.56 ng/ml



#6 AR-1016-4


R.T.: 4.750 min
Delta R.T.: 0.000 min
Response: 244122
Conc: 218.58 ng/ml



#7 AR-1016-5


R.T.: 5.947 min
Delta R.T.: -0.002 min
Response: 337941
Conc: 221.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124471BS




#7 AR-1016-5

R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 299525
Conc: 198.82 ng/ml




#31 AR-1260-1

R.T.: 7.064 min
Delta R.T.: -0.002 min
Response: 567419
Conc: 182.97 ng/ml



#31 AR-1260-1


R.T.: 5.958 min
Delta R.T.: 0.000 min
Response: 531094
Conc: 192.30 ng/ml



#32 AR-1260-2


R.T.: 7.320 min
Delta R.T.: -0.002 min
Response: 729640
Conc: 180.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124471BS




#32 AR-1260-2

R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 774460
Conc: 201.85 ng/ml




#33 AR-1260-3

R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 602198
Conc: 187.51 ng/ml



#33 AR-1260-3


R.T.: 6.292 min
Delta R.T.: 0.000 min
Response: 630484
Conc: 200.94 ng/ml



#34 AR-1260-4


R.T.: 7.903 min
Delta R.T.: 0.000 min
Response: 563668
Conc: 176.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB124471BS




#34 AR-1260-4

R.T.: 6.754 min
Delta R.T.: 0.000 min
Response: 565354
Conc: 223.29 ng/ml



#35 AR-1260-5

R.T.: 8.214 min
Delta R.T.: 0.000 min
Response: 1153077
Conc: 196.61 ng/ml



#35 AR-1260-5

R.T.: 6.995 min
Delta R.T.: 0.000 min
Response: 1348543
Conc: 221.24 ng/ml