

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0110119\
 Data File : P0063431.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Nov 2019 22:21
 Operator : HP/AJ
 Sample : K5633-01
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S7-01-A

Manual Integrations
 APPROVED

Ankita
 11/4/2019 11:51:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 23:30:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.313	3.479	986159	994974	16.921	18.218
2)	SA Decachlor...	9.923	8.314	570577	448040	9.806m	10.480

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110119\
Data File : PO063431.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Nov 2019 22:21
Operator : HP/AJ
Sample : K5633-01
Misc :
ALS Vial : 36 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
S7-01-A


Manual Integrations
APPROVED

Ankita
11/4/2019 11:51:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 23:30:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.313 min
Delta R.T.: -0.003 min
Response: 986159
Conc: 16.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-01-A


Manual Integrations
APPROVED

Ankita
11/4/2019 11:51:14 AM




#1 Tetrachloro-m-xylene

R.T.: 3.479 min
Delta R.T.: 0.000 min
Response: 994974
Conc: 18.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.923 min
Delta R.T.: 0.001 min
Response: 570577
Conc: 9.81 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.314 min
Delta R.T.: -0.007 min
Response: 448040
Conc: 10.48 ng/ml