

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110219\
 Data File : PO063467.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2019 13:23
 Operator : HP/AJ
 Sample : K5633-03DL 2X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 S7-03-ADL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 00:31:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.310	3.477	421540	424573	7.233	7.774
2)	SA Decachlor...	0.000	8.313	0	134339	N.D.	3.142 #
Target Compounds							
3)	L1 AR-1016-1	5.407	4.542	718339	118711	328.583	63.363 #
4)	L1 AR-1016-2	0.000	4.542	0	118711	N.D.	43.523 #
5)	L1 AR-1016-3	0.000	4.701	0	96774	N.D.	69.728 #
6)	L1 AR-1016-4	5.623	4.701	719240	96774	455.480	86.647 #
7)	L1 AR-1016-5	5.946	4.957	1372371	27363	900.866	18.163 #
10)	L2 AR-1221-3	0.000	3.889	0	83835	N.D.	68.180 #
11)	L3 AR-1232-1	0.000	3.889	0	83835	N.D.	49.119 #
12)	L3 AR-1232-2	0.000	4.542	0	118711	N.D.	60.350 #
13)	L3 AR-1232-3	0.000	4.701	0	96774	N.D.	100.031 #
14)	L3 AR-1232-4	5.623	4.807	719240	1364627	651.234	1643.558 #
15)	L3 AR-1232-5	0.000	4.957	0	27363	N.D.	29.236 #
16)	L4 AR-1242-1	5.407	4.542	718339	118711	447.842	85.906 #
17)	L4 AR-1242-2	0.000	4.542	0	118711	N.D.	57.861 #
18)	L4 AR-1242-3	5.623	4.701	719240	96774	517.654	92.343 #
19)	L4 AR-1242-4	5.623	4.807	719240	1364627	615.602	1374.774 #
20)	L4 AR-1242-5	6.349	5.295	653199	1181443	453.833	845.683 #
21)	L5 AR-1248-1	5.407	4.542	718339	118711	582.624	108.611 #
22)	L5 AR-1248-2	0.000	4.701	0	96774	N.D.	69.203 #
23)	L5 AR-1248-3	5.946	4.807	1372371	1364627	707.929	943.597 #
24)	L5 AR-1248-4	6.349	4.957	653199	27363	262.438	15.196 #
25)	L5 AR-1248-5	6.349	5.336	653199	133410	279.797	72.170 #
26)	L6 AR-1254-1	6.323	5.295	947455	1181443	372.659	392.207
27)	L6 AR-1254-2	6.536	5.435	1021515	596514	252.537	221.177
28)	L6 AR-1254-3	6.899	5.841	605496	1507715	142.255	351.186 #
29)	L6 AR-1254-4	7.179	6.080	450984	683881	133.238	239.988 #
30)	L6 AR-1254-5	7.595	6.457	3409825	2675689	1027.982	717.796 #
31)	L7 AR-1260-1	7.058	5.952	2073580	2001133	668.649	724.563
32)	L7 AR-1260-2	7.312	6.139	3474814	3185937	860.752	830.347
33)	L7 AR-1260-3	7.670	6.286	2742018	2409352	853.818	767.888
34)	L7 AR-1260-4	7.895	6.747	2346239	2184435	734.954	862.765
35)	L7 AR-1260-5	8.206	6.989	5093343	5405978	868.458	886.893
36)	L8 AR-1262-1	7.670	6.547	2742018	1322159	574.427	726.634 #
37)	L8 AR-1262-2	8.206	6.989	5093343	5405978	730.554	758.512
38)	L8 AR-1262-3	8.518	7.265	2981696	1285465	645.127	479.627 #
39)	L8 AR-1262-4	8.574	7.328	2030482	3367798	563.948	734.379 #
40)	L8 AR-1262-5	9.217	7.818	1211603	1036958	529.375	528.429
41)	L9 AR-1268-1	8.518	7.265	2981696	1285465	356.026	161.316 #
42)	L9 AR-1268-2	8.574	7.328	2030482	3367798	277.126	511.473 #
43)	L9 AR-1268-3	0.000	7.529	0	78661	N.D.	14.076 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110219\
 Data File : P0063467.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2019 13:23
 Operator : HP/AJ
 Sample : K5633-03DL 2X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S7-03-ADL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 00:31:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	9.217	7.818	1211603	1036958	492.258	480.441
45) L9	AR-1268-5	9.606	8.086	380189	275207	21.172	18.381



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110219\
Data File : P0063467.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Nov 2019 13:23
Operator : HP/AJ
Sample : K5633-03DL 2X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S7-03-ADL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:31:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.310 min
Delta R.T.: -0.006 min
Response: 421540
Conc: 7.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#1 Tetrachloro-m-xylene

R.T.: 3.477 min
Delta R.T.: 0.000 min
Response: 424573
Conc: 7.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.922 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 8.313 min
Delta R.T.: -0.008 min
Response: 134339
Conc: 3.14 ng/ml



#3 AR-1016-1


R.T.: 5.407 min
Delta R.T.: -0.063 min
Response: 718339
Conc: 328.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#3 AR-1016-1

R.T.: 4.542 min
Delta R.T.: 0.020 min
Response: 118711
Conc: 63.36 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T.: 5.490 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 4.542 min
Delta R.T.: 0.003 min
Response: 118711
Conc: 43.52 ng/ml



#5 AR-1016-3


R.T.: 0.000 min
Exp R.T.: 5.550 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#5 AR-1016-3

R.T.: 4.701 min
Delta R.T.: -0.007 min
Response: 96774
Conc: 69.73 ng/ml




#6 AR-1016-4

R.T.: 5.623 min
Delta R.T.: -0.027 min
Response: 719240
Conc: 455.48 ng/ml



#6 AR-1016-4


R.T.: 4.701 min
Delta R.T.: -0.049 min
Response: 96774
Conc: 86.65 ng/ml



#7 AR-1016-5


R.T.: 5.946 min
Delta R.T.: 0.006 min
Response: 1372371
Conc: 900.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#7 AR-1016-5

R.T.: 4.957 min
Delta R.T.: 0.002 min
Response: 27363
Conc: 18.16 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.687 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 3.889 min
Delta R.T.: 0.049 min
Response: 83835
Conc: 68.18 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 4.693 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#11 AR-1232-1

R.T.: 3.889 min
Delta R.T.: 0.048 min
Response: 83835
Conc: 49.12 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T.: 5.216 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 4.542 min
Delta R.T.: 0.002 min
Response: 118711
Conc: 60.35 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 5.504 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#13 AR-1232-3

R.T.: 4.701 min
Delta R.T.: -0.008 min
Response: 96774
Conc: 100.03 ng/ml




#14 AR-1232-4

R.T.: 5.623 min
Delta R.T.: -0.041 min
Response: 719240
Conc: 651.23 ng/ml



#14 AR-1232-4


R.T.: 4.807 min
Delta R.T.: 0.016 min
Response: 1364627
Conc: 1643.56 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 5.753 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#15 AR-1232-5

R.T.: 4.957 min
Delta R.T.: 0.000 min
Response: 27363
Conc: 29.24 ng/ml




#16 AR-1242-1

R.T.: 5.407 min
Delta R.T.: -0.070 min
Response: 718339
Conc: 447.84 ng/ml



#16 AR-1242-1


R.T.: 4.542 min
Delta R.T.: 0.020 min
Response: 118711
Conc: 85.91 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T.: 5.499 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#17 AR-1242-2

R.T.: 4.542 min
Delta R.T.: 0.003 min
Response: 118711
Conc: 57.86 ng/ml




#18 AR-1242-3

R.T.: 5.623 min
Delta R.T.: 0.063 min
Response: 719240
Conc: 517.65 ng/ml



#18 AR-1242-3


R.T.: 4.701 min
Delta R.T.: -0.007 min
Response: 96774
Conc: 92.34 ng/ml



#19 AR-1242-4


R.T.: 5.623 min
Delta R.T.: -0.035 min
Response: 719240
Conc: 615.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#19 AR-1242-4

R.T.: 4.807 min
Delta R.T.: 0.017 min
Response: 1364627
Conc: 1374.77 ng/ml




#20 AR-1242-5

R.T.: 6.349 min
Delta R.T.: -0.040 min
Response: 653199
Conc: 453.83 ng/ml



#20 AR-1242-5


R.T.: 5.295 min
Delta R.T.: -0.002 min
Response: 1181443
Conc: 845.68 ng/ml



#21 AR-1248-1


R.T.: 5.407 min
Delta R.T.: -0.069 min
Response: 718339
Conc: 582.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#21 AR-1248-1

R.T.: 4.542 min
Delta R.T.: 0.020 min
Response: 118711
Conc: 108.61 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T. : 5.746 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 4.701 min
Delta R.T.: -0.049 min
Response: 96774
Conc: 69.20 ng/ml



#23 AR-1248-3


R.T.: 5.946 min
Delta R.T.: -0.003 min
Response: 1372371
Conc: 707.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#23 AR-1248-3

R.T.: 4.807 min
Delta R.T.: 0.017 min
Response: 1364627
Conc: 943.60 ng/ml




#24 AR-1248-4

R.T.: 6.349 min
Delta R.T.: -0.002 min
Response: 653199
Conc: 262.44 ng/ml



#24 AR-1248-4


R.T.: 4.957 min
Delta R.T.: 0.002 min
Response: 27363
Conc: 15.20 ng/ml



#25 AR-1248-5


R.T.: 6.349 min
Delta R.T.: -0.040 min
Response: 653199
Conc: 279.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#25 AR-1248-5

R.T.: 5.336 min
Delta R.T.: 0.000 min
Response: 133410
Conc: 72.17 ng/ml




#26 AR-1254-1

R.T.: 6.323 min
Delta R.T.: 0.000 min
Response: 947455
Conc: 372.66 ng/ml



#26 AR-1254-1


R.T.: 5.295 min
Delta R.T.: 0.000 min
Response: 1181443
Conc: 392.21 ng/ml



#27 AR-1254-2


R.T.: 6.536 min
Delta R.T.: -0.003 min
Response: 1021515
Conc: 252.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#27 AR-1254-2

R.T.: 5.435 min
Delta R.T.: -0.006 min
Response: 596514
Conc: 221.18 ng/ml




#28 AR-1254-3

R.T.: 6.899 min
Delta R.T.: -0.005 min
Response: 605496
Conc: 142.26 ng/ml



#28 AR-1254-3


R.T.: 5.841 min
Delta R.T.: 0.008 min
Response: 1507715
Conc: 351.19 ng/ml



#29 AR-1254-4


R.T.: 7.179 min
Delta R.T.: -0.010 min
Response: 450984
Conc: 133.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#29 AR-1254-4

R.T.: 6.080 min
Delta R.T.: 0.024 min
Response: 683881
Conc: 239.99 ng/ml




#30 AR-1254-5

R.T.: 7.595 min
Delta R.T.: -0.011 min
Response: 3409825
Conc: 1027.98 ng/ml



#30 AR-1254-5


R.T.: 6.457 min
Delta R.T.: -0.007 min
Response: 2675689
Conc: 717.80 ng/ml



#31 AR-1260-1


R.T.: 7.058 min
Delta R.T.: -0.002 min
Response: 2073580
Conc: 668.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#31 AR-1260-1

R.T.: 5.952 min
Delta R.T.: -0.006 min
Response: 2001133
Conc: 724.56 ng/ml




#32 AR-1260-2

R.T.: 7.312 min
Delta R.T.: 0.002 min
Response: 3474814
Conc: 860.75 ng/ml



#32 AR-1260-2


R.T.: 6.139 min
Delta R.T.: -0.006 min
Response: 3185937
Conc: 830.35 ng/ml



#33 AR-1260-3


R.T.: 7.670 min
Delta R.T.: 0.000 min
Response: 2742018
Conc: 853.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#33 AR-1260-3

R.T.: 6.286 min
Delta R.T.: -0.007 min
Response: 2409352
Conc: 767.89 ng/ml




#34 AR-1260-4

R.T.: 7.895 min
Delta R.T.: -0.005 min
Response: 2346239
Conc: 734.95 ng/ml



#34 AR-1260-4


R.T.: 6.747 min
Delta R.T.: -0.008 min
Response: 2184435
Conc: 862.76 ng/ml



#35 AR-1260-5


R.T.: 8.206 min
Delta R.T.: -0.004 min
Response: 5093343
Conc: 868.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#35 AR-1260-5

R.T.: 6.989 min
Delta R.T.: -0.007 min
Response: 5405978
Conc: 886.89 ng/ml




#36 AR-1262-1

R.T.: 7.670 min
Delta R.T.: -0.009 min
Response: 2742018
Conc: 574.43 ng/ml



#36 AR-1262-1


R.T.: 6.547 min
Delta R.T.: -0.007 min
Response: 1322159
Conc: 726.63 ng/ml



#37 AR-1262-2


R.T.: 8.206 min
Delta R.T.: -0.008 min
Response: 5093343
Conc: 730.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#37 AR-1262-2

R.T.: 6.989 min
Delta R.T.: -0.007 min
Response: 5405978
Conc: 758.51 ng/ml




#38 AR-1262-3

R.T.: 8.518 min
Delta R.T.: 0.006 min
Response: 2981696
Conc: 645.13 ng/ml



#38 AR-1262-3


R.T.: 7.265 min
Delta R.T.: -0.007 min
Response: 1285465
Conc: 479.63 ng/ml



#39 AR-1262-4


R.T.: 8.574 min
Delta R.T.: -0.026 min
Response: 2030482
Conc: 563.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#39 AR-1262-4

R.T.: 7.328 min
Delta R.T.: -0.008 min
Response: 3367798
Conc: 734.38 ng/ml




#40 AR-1262-5

R.T.: 9.217 min
Delta R.T.: -0.011 min
Response: 1211603
Conc: 529.37 ng/ml



#40 AR-1262-5


R.T.: 7.818 min
Delta R.T.: -0.007 min
Response: 1036958
Conc: 528.43 ng/ml



#41 AR-1268-1


R.T.: 8.518 min
Delta R.T.: 0.005 min
Response: 2981696
Conc: 356.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#41 AR-1268-1

R.T.: 7.265 min
Delta R.T.: -0.007 min
Response: 1285465
Conc: 161.32 ng/ml




#42 AR-1268-2

R.T.: 8.574 min
Delta R.T.: -0.030 min
Response: 2030482
Conc: 277.13 ng/ml



#42 AR-1268-2


R.T.: 7.328 min
Delta R.T.: -0.009 min
Response: 3367798
Conc: 511.47 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.818 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL




#43 AR-1268-3

R.T.: 7.529 min
Delta R.T.: -0.008 min
Response: 78661
Conc: 14.08 ng/ml




#44 AR-1268-4

R.T.: 9.217 min
Delta R.T.: -0.012 min
Response: 1211603
Conc: 492.26 ng/ml



#44 AR-1268-4


R.T.: 7.818 min
Delta R.T.: -0.007 min
Response: 1036958
Conc: 480.44 ng/ml



#45 AR-1268-5

R.T.: 9.606 min
Delta R.T.: -0.014 min
Response: 380189
Conc: 21.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
S7-03-ADL



#45 AR-1268-5

R.T.: 8.086 min
Delta R.T.: -0.007 min
Response: 275207
Conc: 18.38 ng/ml