

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110219\
 Data File : PO063494.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2019 21:23
 Operator : HP/AJ
 Sample : K5642-05MS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S1-05-AMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 00:41:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.310	3.478	908736	909143	15.592	16.647
2)	SA Decachlor...	9.922	8.312	493703	332810	8.485	7.785
Target Compounds							
3)	L1 AR-1016-1	5.469	4.518	380295	300436	173.955	160.361
4)	L1 AR-1016-2	5.490	4.535	639289	612464	204.063	224.549
5)	L1 AR-1016-3	5.553	4.704	376454	366416	200.939	264.011 #
6)	L1 AR-1016-4	5.624	4.746	1290014	196524	816.939	175.959 #
7)	L1 AR-1016-5	5.899	4.953	2278815	100580	1495.884	66.763 #
8)	L2 AR-1221-1	0.000	3.682	0	37399	N.D.	72.329 #
9)	L2 AR-1221-2	4.681	3.765	228949	43628	529.065	112.999 #
10)	L2 AR-1221-3	4.681	3.837	228949	240965	165.709	195.967
11)	L3 AR-1232-1	4.681	3.837	228949	240965	119.985	141.181
12)	L3 AR-1232-2	5.202	4.535	432565	612464	436.974	311.360 #
13)	L3 AR-1232-3	5.490	4.704	639289	366416	282.309	378.748 #
14)	L3 AR-1232-4	5.624	4.786	1290014	226715	1168.039	273.056 #
15)	L3 AR-1232-5	5.739	4.953	246794	100580	307.775	107.462 #
16)	L4 AR-1242-1	5.469	4.518	380295	300436	237.091	217.413
17)	L4 AR-1242-2	5.490	4.535	639289	612464	277.275	298.519
18)	L4 AR-1242-3	5.553	4.704	376454	366416	270.942	349.637 #
19)	L4 AR-1242-4	5.624	4.786	1290014	226715	1104.130	228.401 #
20)	L4 AR-1242-5	6.348	5.296	1685579	1487834	1171.116	1064.999
21)	L5 AR-1248-1	5.469	4.518	380295	300436	308.446	274.873
22)	L5 AR-1248-2	5.739	4.746	246794	196524	142.471	140.534
23)	L5 AR-1248-3	5.899	4.786	2278815	226715	1175.513	156.766 #
24)	L5 AR-1248-4	6.348	4.953	1685579	100580	677.222	55.856 #
25)	L5 AR-1248-5	6.348	5.335	1685579	194370	722.017	105.147 #
26)	L6 AR-1254-1	6.326	5.296	1182053	1487834	464.933	493.921
27)	L6 AR-1254-2	6.541	5.433	1183526	324559	292.589	120.341 #
28)	L6 AR-1254-3	6.904	5.841	554464	452922	130.266	105.497
29)	L6 AR-1254-4	7.155	6.063	210453	169025	62.176	59.314
30)	L6 AR-1254-5	7.596	6.456	772387	628169	232.856	168.516 #
31)	L7 AR-1260-1	7.057	5.952	477265	498824	153.899	180.612
32)	L7 AR-1260-2	7.311	6.138	1538747	724869	381.166	188.922 #
33)	L7 AR-1260-3	7.670	6.286	414947	557449	129.207	177.665 #
34)	L7 AR-1260-4	7.894	6.747	414252	357499	129.764	141.198
35)	L7 AR-1260-5	8.205	6.988	755108	844177	128.752	138.494
36)	L8 AR-1262-1	7.670	6.547	414947	243899	86.927	134.042 #
37)	L8 AR-1262-2	8.205	6.988	755108	844177	108.307	118.446
38)	L8 AR-1262-3	8.519	7.265	422001	155948	91.305	58.187 #
39)	L8 AR-1262-4	0.000	7.327	0	499119	N.D.	108.837 #
40)	L8 AR-1262-5	0.000	7.818	0	125234	N.D.	63.819 #
41)	L9 AR-1268-1	8.519	7.265	422001	155948	50.389	19.570 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110219\
 Data File : P0063494.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2019 21:23
 Operator : HP/AJ
 Sample : K5642-05MS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S1-05-AMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 00:41:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	0.000	7.327	0	499119	N.D.	75.802 #
44) L9	AR-1268-4	0.000	7.818	0	125234	N.D.	58.023 #
45) L9	AR-1268-5	0.000	8.085	0	49292	N.D.	3.292 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110219\
Data File : PO063494.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Nov 2019 21:23
Operator : HP/AJ
Sample : K5642-05MS
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:41:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.310 min
Delta R.T.: -0.006 min
Response: 908736
Conc: 15.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#1 Tetrachloro-m-xylene

R.T.: 3.478 min
Delta R.T.: 0.000 min
Response: 909143
Conc: 16.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.922 min
Delta R.T.: 0.000 min
Response: 493703
Conc: 8.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.312 min
Delta R.T.: -0.009 min
Response: 332810
Conc: 7.78 ng/ml



#3 AR-1016-1


R.T.: 5.469 min
Delta R.T.: 0.000 min
Response: 380295
Conc: 173.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#3 AR-1016-1

R.T.: 4.518 min
Delta R.T.: -0.004 min
Response: 300436
Conc: 160.36 ng/ml




#4 AR-1016-2

R.T.: 5.490 min
Delta R.T.: 0.000 min
Response: 639289
Conc: 204.06 ng/ml



#4 AR-1016-2


R.T.: 4.535 min
Delta R.T.: -0.003 min
Response: 612464
Conc: 224.55 ng/ml



#5 AR-1016-3


R.T.: 5.553 min
Delta R.T.: 0.003 min
Response: 376454
Conc: 200.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#5 AR-1016-3

R.T.: 4.704 min
Delta R.T.: -0.005 min
Response: 366416
Conc: 264.01 ng/ml




#6 AR-1016-4

R.T.: 5.624 min
Delta R.T.: -0.026 min
Response: 1290014
Conc: 816.94 ng/ml



#6 AR-1016-4


R.T.: 4.746 min
Delta R.T.: -0.004 min
Response: 196524
Conc: 175.96 ng/ml



#7 AR-1016-5


R.T.: 5.899 min
Delta R.T.: -0.041 min
Response: 2278815
Conc: 1495.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#7 AR-1016-5

R.T.: 4.953 min
Delta R.T.: -0.003 min
Response: 100580
Conc: 66.76 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.526 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.682 min
Delta R.T.: -0.003 min
Response: 37399
Conc: 72.33 ng/ml



#9 AR-1221-2


R.T.: 4.681 min
Delta R.T.: 0.069 min
Response: 228949
Conc: 529.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#9 AR-1221-2

R.T.: 3.765 min
Delta R.T.: -0.002 min
Response: 43628
Conc: 113.00 ng/ml




#10 AR-1221-3

R.T.: 4.681 min
Delta R.T.: -0.006 min
Response: 228949
Conc: 165.71 ng/ml



#10 AR-1221-3


R.T.: 3.837 min
Delta R.T.: -0.003 min
Response: 240965
Conc: 195.97 ng/ml



#11 AR-1232-1


R.T.: 4.681 min
Delta R.T.: -0.012 min
Response: 228949
Conc: 119.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#11 AR-1232-1

R.T.: 3.837 min
Delta R.T.: -0.004 min
Response: 240965
Conc: 141.18 ng/ml




#12 AR-1232-2

R.T.: 5.202 min
Delta R.T.: -0.014 min
Response: 432565
Conc: 436.97 ng/ml



#12 AR-1232-2


R.T.: 4.535 min
Delta R.T.: -0.005 min
Response: 612464
Conc: 311.36 ng/ml



#13 AR-1232-3


R.T.: 5.490 min
Delta R.T.: -0.014 min
Response: 639289
Conc: 282.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#13 AR-1232-3

R.T.: 4.704 min
Delta R.T.: -0.006 min
Response: 366416
Conc: 378.75 ng/ml




#14 AR-1232-4

R.T.: 5.624 min
Delta R.T.: -0.040 min
Response: 1290014
Conc: 1168.04 ng/ml



#14 AR-1232-4


R.T.: 4.786 min
Delta R.T.: -0.006 min
Response: 226715
Conc: 273.06 ng/ml



#15 AR-1232-5


R.T.: 5.739 min
Delta R.T.: -0.013 min
Response: 246794
Conc: 307.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#15 AR-1232-5

R.T.: 4.953 min
Delta R.T.: -0.004 min
Response: 100580
Conc: 107.46 ng/ml




#16 AR-1242-1

R.T.: 5.469 min
Delta R.T.: -0.008 min
Response: 380295
Conc: 237.09 ng/ml



#16 AR-1242-1


R.T.: 4.518 min
Delta R.T.: -0.004 min
Response: 300436
Conc: 217.41 ng/ml



#17 AR-1242-2


R.T.: 5.490 min
Delta R.T.: -0.009 min
Response: 639289
Conc: 277.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#17 AR-1242-2

R.T.: 4.535 min
Delta R.T.: -0.003 min
Response: 612464
Conc: 298.52 ng/ml




#18 AR-1242-3

R.T.: 5.553 min
Delta R.T.: -0.008 min
Response: 376454
Conc: 270.94 ng/ml



#18 AR-1242-3


R.T.: 4.704 min
Delta R.T.: -0.005 min
Response: 366416
Conc: 349.64 ng/ml



#19 AR-1242-4


R.T.: 5.624 min
Delta R.T.: -0.035 min
Response: 1290014
Conc: 1104.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#19 AR-1242-4

R.T.: 4.786 min
Delta R.T.: -0.004 min
Response: 226715
Conc: 228.40 ng/ml




#20 AR-1242-5

R.T.: 6.348 min
Delta R.T.: -0.041 min
Response: 1685579
Conc: 1171.12 ng/ml



#20 AR-1242-5


R.T.: 5.296 min
Delta R.T.: 0.000 min
Response: 1487834
Conc: 1065.00 ng/ml



#21 AR-1248-1


R.T.: 5.469 min
Delta R.T.: -0.007 min
Response: 380295
Conc: 308.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#21 AR-1248-1

R.T.: 4.518 min
Delta R.T.: -0.004 min
Response: 300436
Conc: 274.87 ng/ml




#22 AR-1248-2

R.T.: 5.739 min
Delta R.T.: -0.007 min
Response: 246794
Conc: 142.47 ng/ml



#22 AR-1248-2


R.T.: 4.746 min
Delta R.T.: -0.005 min
Response: 196524
Conc: 140.53 ng/ml



#23 AR-1248-3


R.T.: 5.899 min
Delta R.T.: -0.050 min
Response: 2278815
Conc: 1175.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#23 AR-1248-3

R.T.: 4.786 min
Delta R.T.: -0.004 min
Response: 226715
Conc: 156.77 ng/ml




#24 AR-1248-4

R.T.: 6.348 min
Delta R.T.: -0.003 min
Response: 1685579
Conc: 677.22 ng/ml



#24 AR-1248-4


R.T.: 4.953 min
Delta R.T.: -0.003 min
Response: 100580
Conc: 55.86 ng/ml



#25 AR-1248-5


R.T.: 6.348 min
Delta R.T.: -0.040 min
Response: 1685579
Conc: 722.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#25 AR-1248-5

R.T.: 5.335 min
Delta R.T.: 0.000 min
Response: 194370
Conc: 105.15 ng/ml




#26 AR-1254-1

R.T.: 6.326 min
Delta R.T.: 0.003 min
Response: 1182053
Conc: 464.93 ng/ml



#26 AR-1254-1


R.T.: 5.296 min
Delta R.T.: 0.000 min
Response: 1487834
Conc: 493.92 ng/ml



#27 AR-1254-2


R.T.: 6.541 min
Delta R.T.: 0.002 min
Response: 1183526
Conc: 292.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#27 AR-1254-2

R.T.: 5.433 min
Delta R.T.: -0.007 min
Response: 324559
Conc: 120.34 ng/ml




#28 AR-1254-3

R.T.: 6.904 min
Delta R.T.: 0.000 min
Response: 554464
Conc: 130.27 ng/ml



#28 AR-1254-3


R.T.: 5.841 min
Delta R.T.: 0.009 min
Response: 452922
Conc: 105.50 ng/ml



#29 AR-1254-4


R.T.: 7.155 min
Delta R.T.: -0.033 min
Response: 210453
Conc: 62.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#29 AR-1254-4

R.T.: 6.063 min
Delta R.T.: 0.007 min
Response: 169025
Conc: 59.31 ng/ml




#30 AR-1254-5

R.T.: 7.596 min
Delta R.T.: -0.010 min
Response: 772387
Conc: 232.86 ng/ml



#30 AR-1254-5


R.T.: 6.456 min
Delta R.T.: -0.008 min
Response: 628169
Conc: 168.52 ng/ml



#31 AR-1260-1


R.T.: 7.057 min
Delta R.T.: -0.003 min
Response: 477265
Conc: 153.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#31 AR-1260-1

R.T.: 5.952 min
Delta R.T.: -0.007 min
Response: 498824
Conc: 180.61 ng/ml




#32 AR-1260-2

R.T.: 7.311 min
Delta R.T.: 0.000 min
Response: 1538747
Conc: 381.17 ng/ml



#32 AR-1260-2


R.T.: 6.138 min
Delta R.T.: -0.007 min
Response: 724869
Conc: 188.92 ng/ml



#33 AR-1260-3


R.T.: 7.670 min
Delta R.T.: 0.000 min
Response: 414947
Conc: 129.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#33 AR-1260-3

R.T.: 6.286 min
Delta R.T.: -0.007 min
Response: 557449
Conc: 177.67 ng/ml




#34 AR-1260-4

R.T.: 7.894 min
Delta R.T.: -0.006 min
Response: 414252
Conc: 129.76 ng/ml



#34 AR-1260-4


R.T.: 6.747 min
Delta R.T.: -0.008 min
Response: 357499
Conc: 141.20 ng/ml



#35 AR-1260-5


R.T.: 8.205 min
Delta R.T.: -0.005 min
Response: 755108
Conc: 128.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#35 AR-1260-5

R.T.: 6.988 min
Delta R.T.: -0.007 min
Response: 844177
Conc: 138.49 ng/ml




#36 AR-1262-1

R.T.: 7.670 min
Delta R.T.: -0.009 min
Response: 414947
Conc: 86.93 ng/ml



#36 AR-1262-1


R.T.: 6.547 min
Delta R.T.: -0.007 min
Response: 243899
Conc: 134.04 ng/ml



#37 AR-1262-2


R.T.: 8.205 min
Delta R.T.: -0.009 min
Response: 755108
Conc: 108.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#37 AR-1262-2

R.T.: 6.988 min
Delta R.T.: -0.007 min
Response: 844177
Conc: 118.45 ng/ml




#38 AR-1262-3

R.T.: 8.519 min
Delta R.T.: 0.008 min
Response: 422001
Conc: 91.31 ng/ml



#38 AR-1262-3


R.T.: 7.265 min
Delta R.T.: -0.007 min
Response: 155948
Conc: 58.19 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.600 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#39 AR-1262-4

R.T.: 7.327 min
Delta R.T.: -0.008 min
Response: 499119
Conc: 108.84 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 9.228 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.818 min
Delta R.T.: -0.007 min
Response: 125234
Conc: 63.82 ng/ml



#41 AR-1268-1


R.T.: 8.519 min
Delta R.T.: 0.007 min
Response: 422001
Conc: 50.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#41 AR-1268-1

R.T.: 7.265 min
Delta R.T.: -0.007 min
Response: 155948
Conc: 19.57 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.603 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.327 min
Delta R.T.: -0.009 min
Response: 499119
Conc: 75.80 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 9.230 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
S1-05-AMS




#44 AR-1268-4

R.T.: 7.818 min
Delta R.T.: -0.007 min
Response: 125234
Conc: 58.02 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.620 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.085 min
Delta R.T.: -0.008 min
Response: 49292
Conc: 3.29 ng/ml