

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO111218\
 Data File : PO051440.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Nov 2018 3:02
 Operator : SM/SJ
 Sample : J5164-09
 Misc : AR1232 MDL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 MDL-WATER-03-QT4-2018

Manual Integrations
 APPROVED

Sohil
 11/14/2018 5:58:35 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 06:27:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 13 02:51:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.331	3.332	13559198	4513295	24.051	20.226
2) SA Decachlor...	9.961	8.061	8137582	5121659	21.930	23.108
Target Compounds						
11) L3 AR-1232-1	4.703	3.684	427656	98465	42.861	23.298m#
12) L3 AR-1232-2	5.226	4.360	170030	166397	32.442m	29.001m
13) L3 AR-1232-3	5.515	4.521	399472	78144	36.112m	30.470m
14) L3 AR-1232-4	5.675	4.602	120195	59800	22.891m	26.997m
15) L3 AR-1232-5	5.761	4.766	125407	59611	33.463m	23.265m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111218\
Data File : PO051440.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Nov 2018 3:02
Operator : SM/SJ
Sample : J5164-09
Misc : AR1232 MDL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MDL-WATER-03-QT4-2018

Manual Integrations
APPROVED

Sohil
11/14/2018 5:58:35 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 06:27:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110918.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 13 02:51:44 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

