

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112818\
 Data File : PO051898.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Nov 2018 12:07
 Operator : SM/SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 28 14:06:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 21 00:55:28 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.314	3.309	14893522	5001667	19.784	20.896
2) SA Decachlor...	9.928	8.029	10265696	4117959	22.130	19.049
Target Compounds						
34) L7 AR-1260-4	7.969	0.000	101985	0	3.708	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112818\
Data File : PO051898.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Nov 2018 12:07
Operator : SM/SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 28 14:06:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112018.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 21 00:55:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.314 min
Delta R.T.: -0.005 min
Response: 14893522
Conc: 19.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.309 min
Delta R.T.: -0.012 min
Response: 5001667
Conc: 20.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.928 min
Delta R.T.: -0.012 min
Response: 10265696
Conc: 22.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.029 min
Delta R.T.: -0.016 min
Response: 4117959
Conc: 19.05 ng/ml



#34 AR-1260-4

R.T.: 7.969 min
Delta R.T.: 0.061 min
Response: 101985
Conc: 3.71 ng/ml

Instrument :
ECD_O
ClientSampleId :



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 6.512 min
Response: 0
Conc: N.D.