

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120118\
 Data File : P0052109.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Dec 2018 13:41
 Operator : SM/SJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 06:23:43 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 03 06:23:19 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.281	3.291	54535480	20086660	48.534	47.485
2)	SA Decachlor...	9.869	8.002	41029925	17724230	44.606	45.567
Target Compounds							
26)	L6 AR-1254-1	6.287	5.049	16729677	11294375	500.000	500.000
27)	L6 AR-1254-2	6.504	5.190	26351151	9827525	500.000	500.000
28)	L6 AR-1254-3	6.869	5.570	28170868	15414610	500.000	500.000
29)	L6 AR-1254-4	7.153	5.792	20870379	9299540	500.000	500.000
30)	L6 AR-1254-5	7.569	6.190	22092672	13218155	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120118\
Data File : PO052109.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Dec 2018 13:41
Operator : SM/SJ
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 06:23:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 03 06:23:19 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

