

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0120318.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 04 03:10:39 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0052145.D 1000 =P0052143.D 750 =P0052144.D
 250 =P0052146.D 50 =P0052147.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	8.290	9.689	8.495	8.161	8.412	8.609	E5	7.16
2) SA	Decachlorobiphenyl	4.992	4.922	4.916	4.895	5.298	5.004	E5	3.36
3) L1	AR-1016-1	2.201	2.438	2.220	2.218	2.357	2.287	E4	4.60
4) L1	AR-1016-2	3.407	3.607	3.364	3.360	3.664	3.481	E4	4.14
5) L1	AR-1016-3	2.048	2.173	2.005	2.023	2.139	2.078	E4	3.56
6) L1	AR-1016-4	1.618	1.798	1.613	1.631	1.673	1.667	E4	4.63
7) L1	AR-1016-5	1.557	1.696	1.557	1.555	1.634	1.600	E4	3.98
8) L2	AR-1221-1	7.689					7.689	E3	0.00
9) L2	AR-1221-2	5.607					5.607	E3	0.00
10) L2	AR-1221-3	1.767					1.767	E4	0.00
11) L3	AR-1232-1	1.471					1.471	E4	0.00
12) L3	AR-1232-2	7.401					7.401	E3	0.00
13) L3	AR-1232-3	1.532					1.532	E4	0.00
14) L3	AR-1232-4	7.200					7.200	E3	0.00
15) L3	AR-1232-5	4.680					4.680	E3	0.00
16) L4	AR-1242-1	1.846					1.846	E4	0.00
17) L4	AR-1242-2	2.747					2.747	E4	0.00
18) L4	AR-1242-3	1.580					1.580	E4	0.00
19) L4	AR-1242-4	1.349					1.349	E4	0.00
20) L4	AR-1242-5	1.410					1.410	E4	0.00
21) L5	AR-1248-1	1.325					1.325	E4	0.00
22) L5	AR-1248-2	1.868					1.868	E4	0.00
23) L5	AR-1248-3	2.109					2.109	E4	0.00
24) L5	AR-1248-4	2.296					2.296	E4	0.00
25) L5	AR-1248-5	2.219					2.219	E4	0.00
26) L6	AR-1254-1	2.584					2.584	E4	0.00
27) L6	AR-1254-2	3.989					3.989	E4	0.00
28) L6	AR-1254-3	4.197					4.197	E4	0.00
29) L6	AR-1254-4	2.864					2.864	E4	0.00
30) L6	AR-1254-5	2.945					2.945	E4	0.00
31) L7	AR-1260-1	2.854	2.881	2.767	2.954	3.303	2.952	E4	7.02
32) L7	AR-1260-2	3.427	3.579	3.435	3.408	3.765	3.523	E4	4.30
33) L7	AR-1260-3	2.352	2.444	2.325	2.356	2.616	2.419	E4	4.93
34) L7	AR-1260-4	2.419	2.437	2.395	2.416	2.568	2.447	E4	2.83
35) L7	AR-1260-5	4.881	4.931	4.900	4.833	5.022	4.913	E4	1.43
36) L8	AR-1262-1	4.072					4.072	E4	0.00
37) L8	AR-1262-2	6.487					6.487	E4	0.00
38) L8	AR-1262-3	4.143					4.143	E4	0.00
39) L8	AR-1262-4	3.322					3.322	E4	0.00
40) L8	AR-1262-5	2.097					2.097	E4	0.00
41) L9	AR-1268-1	9.326					9.326	E4	0.00
42) L9	AR-1268-2	8.249					8.249	E4	0.00
43) L9	AR-1268-3	7.590					7.590	E4	0.00
44) L9	AR-1268-4	2.602					2.602	E4	0.00
45) L9	AR-1268-5	2.053					2.053	E5	0.00

Signal #2 Calibration Files

500 =P0052145.D 1000 =P0052143.D 750 =P0052144.D
 250 =P0052146.D 50 =P0052147.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.268	2.344	2.293	2.185	2.061	2.230	E5	4.96
2) SA	Decachlorobiphenyl	2.045	1.980	1.990	2.063	2.346	2.085	E5	7.21
3) L1	AR-1016-1	6.747	6.612	6.635	7.031	8.838	7.173	E3	13.18
4) L1	AR-1016-2	1.221	1.297	1.264	1.258	1.512	1.310	E4	8.84
5) L1	AR-1016-3	5.678	5.830	5.681	5.754	5.635	5.716	E3	1.34

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0120318.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 04 03:10:39 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0052145.D 1000 =P0052143.D 750 =P0052144.D
 250 =P0052146.D 50 =P0052147.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	4.918	4.892	4.789	5.117	5.995	5.142	E3	9.55
7)	L1 AR-1016-5	6.376	6.416	6.345	6.786	8.639	6.912	E3	14.20
8)	L2 AR-1221-1	1.881					1.881	E3	0.00
9)	L2 AR-1221-2	1.629					1.629	E3	0.00
10)	L2 AR-1221-3	4.913					4.913	E3	0.00
11)	L3 AR-1232-1	4.239					4.239	E3	0.00
12)	L3 AR-1232-2	5.609					5.609	E3	0.00
13)	L3 AR-1232-3	2.424					2.424	E3	0.00
14)	L3 AR-1232-4	2.124					2.124	E3	0.00
15)	L3 AR-1232-5	2.507					2.507	E3	0.00
16)	L4 AR-1242-1	5.533					5.533	E3	0.00
17)	L4 AR-1242-2	1.052					1.052	E4	0.00
18)	L4 AR-1242-3	4.772					4.772	E3	0.00
19)	L4 AR-1242-4	4.571					4.571	E3	0.00
20)	L4 AR-1242-5	6.385					6.385	E3	0.00
21)	L5 AR-1248-1	4.493					4.493	E3	0.00
22)	L5 AR-1248-2	6.545					6.545	E3	0.00
23)	L5 AR-1248-3	6.497					6.497	E3	0.00
24)	L5 AR-1248-4	8.253					8.253	E3	0.00
25)	L5 AR-1248-5	8.118					8.118	E3	0.00
26)	L6 AR-1254-1	1.396					1.396	E4	0.00
27)	L6 AR-1254-2	1.258					1.258	E4	0.00
28)	L6 AR-1254-3	1.961					1.961	E4	0.00
29)	L6 AR-1254-4	1.132					1.132	E4	0.00
30)	L6 AR-1254-5	1.685					1.685	E4	0.00
31)	L7 AR-1260-1	1.352	1.331	1.314	1.386	1.608	1.398	E4	8.60
32)	L7 AR-1260-2	1.615	1.566	1.585	1.652	1.930	1.670	E4	8.94
33)	L7 AR-1260-3	1.539	1.507	1.502	1.546	1.825	1.584	E4	8.61
34)	L7 AR-1260-4	1.013	0.989	1.023	1.032	1.162	1.044	E4	6.53
35)	L7 AR-1260-5	2.377	2.256	2.361	2.414	2.681	2.418	E4	6.55
36)	L8 AR-1262-1	1.831					1.831	E4	0.00
37)	L8 AR-1262-2	2.892					2.892	E4	0.00
38)	L8 AR-1262-3	1.176					1.176	E4	0.00
39)	L8 AR-1262-4	2.093					2.093	E4	0.00
40)	L8 AR-1262-5	8.488					8.488	E3	0.00
41)	L9 AR-1268-1	3.952					3.952	E4	0.00
42)	L9 AR-1268-2	3.750					3.750	E4	0.00
43)	L9 AR-1268-3	3.268					3.268	E4	0.00
44)	L9 AR-1268-4	1.107					1.107	E4	0.00
45)	L9 AR-1268-5	8.747					8.747	E4	0.00

(#) = Out of Range