

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0120718\
 Data File : P0052314.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Dec 2018 11:48
 Operator : SM/SJ
 Sample : J5996-25
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AOC1-06-A

Manual Integrations
 APPROVED

Sohil
 12/10/2018 2:19:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 21:54:00 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 04 03:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.271	3.289	14468953	3490874	16.806	15.652
2) SA Decachlor...	9.841	7.982	6773462	2374610	13.535m	11.390m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120718\
Data File : PO052314.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Dec 2018 11:48
Operator : SM/SJ
Sample : J5996-25
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AOC1-06-A


Manual Integrations
APPROVED

Sohil
12/10/2018 2:19:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 21:54:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 04 03:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.271 min
Delta R.T.: -0.012 min
Response: 14468953
Conc: 16.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AOC1-06-A


Manual Integrations
APPROVED

Sohil
12/10/2018 2:19:58 PM




#1 Tetrachloro-m-xylene

R.T.: 3.289 min
Delta R.T.: -0.007 min
Response: 3490874
Conc: 15.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.841 min
Delta R.T.: -0.031 min
Response: 6773462
Conc: 13.53 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.982 min
Delta R.T.: -0.018 min
Response: 2374610
Conc: 11.39 ng/ml m