

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121119\
 Data File : P0064544.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Dec 2019 10:48
 Operator : SM/AJ
 Sample : K6174-04
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 WATER-COMP

Manual Integrations
 APPROVED

Ankita
 12/12/2019 11:12:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 11:31:28 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 28 01:28:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.308	3.555	501199	446977	16.924	18.792m
2) SA Decachlor...	9.926	8.509	833254	893405	20.331	21.870m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121119\
Data File : P0064544.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Dec 2019 10:48
Operator : SM/AJ
Sample : K6174-04
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
WATER-COMP


Manual Integrations
APPROVED

Ankita
12/12/2019 11:12:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 11:31:28 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 28 01:28:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.308 min
Delta R.T.: -0.005 min
Response: 501199
Conc: 16.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
WATER-COMP


Manual Integrations
APPROVED

Ankita
12/12/2019 11:12:58 AM




#1 Tetrachloro-m-xylene

R.T.: 3.555 min
Delta R.T.: -0.006 min
Response: 446977
Conc: 18.79 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.926 min
Delta R.T.: -0.015 min
Response: 833254
Conc: 20.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.509 min
Delta R.T.: -0.018 min
Response: 893405
Conc: 21.87 ng/ml m