

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121218\
 Data File : P0052463.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Dec 2018 20:15
 Operator : SM/SJ
 Sample : J6322-21
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SUP-6-C-2-120718

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 00:44:58 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 04 03:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.263	3.282	11898630	3647962	13.821	16.356
2) SA Decachlor...	9.826	7.973	6055414	2947049	12.100	14.136

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121218\
Data File : PO052463.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Dec 2018 20:15
Operator : SM/SJ
Sample : J6322-21
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SUP-6-C-2-120718

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 00:44:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 04 03:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.263 min
Delta R.T.: -0.020 min
Response: 11898630
Conc: 13.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
SUP-6-C-2-120718




#1 Tetrachloro-m-xylene

R.T.: 3.282 min
Delta R.T.: -0.014 min
Response: 3647962
Conc: 16.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.826 min
Delta R.T.: -0.046 min
Response: 6055414
Conc: 12.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.973 min
Delta R.T.: -0.027 min
Response: 2947049
Conc: 14.14 ng/ml