

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121319\
 Data File : P0064587.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Dec 2019 15:40
 Operator : SM/AJ
 Sample : P0121319ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVP0121319

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 08:25:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	24.345	23.878 E3	1.9	99	0.00
2 SA Decachlorobiphenyl	30.690	30.962 E3	-0.9	105	0.00
3 L1 AR-1016-1	997.833	971.005	2.7	102	0.00
4 L1 AR-1016-2	1.452	1.404 E3	3.3	101	0.00
5 L1 AR-1016-3	884.799	864.269	2.3	101	0.00
6 L1 AR-1016-4	724.322	702.404	3.0	100	0.00
7 L1 AR-1016-5	728.773	724.568	0.6	101	0.00
31 L7 AR-1260-1	1.536	1.513 E3	1.5	102	0.00
32 L7 AR-1260-2	1.974	1.904 E3	3.5	102	0.00
33 L7 AR-1260-3	1.503	1.486 E3	1.1	102	0.00
34 L7 AR-1260-4	1.847	1.802 E3	2.4	102	0.00
35 L7 AR-1260-5	3.732	3.626 E3	2.8	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	25.349	26.836 E3	-5.9	108	0.00
2 SA Decachlorobiphenyl	34.729	35.839 E3	-3.2	108	0.00
3 L1 AR-1016-1	1.081	1.091 E3	-0.9	105	0.00
4 L1 AR-1016-2	1.504	1.526 E3	-1.5	107	0.00
5 L1 AR-1016-3	777.366	798.813	-2.8	105	0.00
6 L1 AR-1016-4	657.491	669.927	-1.9	106	0.00
7 L1 AR-1016-5	872.561	880.519	-0.9	108	0.00
31 L7 AR-1260-1	1.847	1.847 E3	0.0	107	0.00
32 L7 AR-1260-2	2.327	2.357 E3	-1.3	108	0.00
33 L7 AR-1260-3	2.117	2.146 E3	-1.4	106	0.00
34 L7 AR-1260-4	1.884	1.923 E3	-2.1	107	0.00
35 L7 AR-1260-5	4.690	4.731 E3	-0.9	106	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	259.224	0.000	100.0#	0#	-4.38#
9 L2 AR-1221-2	197.641	0.000	100.0#	0#	-4.47#
10 L2 AR-1221-3	662.448	0.000	100.0#	0#	-4.54#
11 L3 AR-1232-1	1.011	0.000 E3	100.0#	0#	-4.54#
12 L3 AR-1232-2	510.907	0.000	100.0#	0#	-5.06#
13 L3 AR-1232-3	1.118	0.000 E3	100.0#	0#	-5.35#
14 L3 AR-1232-4	555.919	0.000	100.0#	0#	-5.51#
15 L3 AR-1232-5	414.989	0.000	100.0#	0#	-5.60#
16 L4 AR-1242-1	783.664	0.000	100.0#	0#	-5.33#
17 L4 AR-1242-2	1.128	0.000 E3	100.0#	0#	-5.35#
18 L4 AR-1242-3	696.291	0.000	100.0#	0#	-5.41#
19 L4 AR-1242-4	563.778	0.000	100.0#	0#	-5.51#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121319\
 Data File : P0064587.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Dec 2019 15:40
 Operator : SM/AJ
 Sample : P0121319ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO121319

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 08:25:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
20 L4 AR-1242-5	696.929	0.000	100.0#	0#	-6.23#
21 L5 AR-1248-1	645.388	0.000	100.0#	0#	-5.33#
22 L5 AR-1248-2	924.078	0.000	100.0#	0#	-5.60#
23 L5 AR-1248-3	1.062	0.000 E3	100.0#	0#	-5.80#
24 L5 AR-1248-4	1.304	0.000 E3	100.0#	0#	-6.20#
25 L5 AR-1248-5	1.244	0.000 E3	100.0#	0#	-6.23#
26 L6 AR-1254-1	1.274	0.000 E3	100.0#	0#	-6.17#
27 L6 AR-1254-2	2.080	0.000 E3	100.0#	0#	-6.39#
28 L6 AR-1254-3	2.388	0.000 E3	100.0#	0#	-6.75#
29 L6 AR-1254-4	1.870	0.000 E3	100.0#	0#	-7.03#
30 L6 AR-1254-5	2.104	0.000 E3	100.0#	0#	-7.45#
36 L8 AR-1262-1	2.299	0.000 E3	100.0#	0#	-7.52#
37 L8 AR-1262-2	4.341	0.000 E3	100.0#	0#	-8.05#
38 L8 AR-1262-3	2.910	0.000 E3	100.0#	0#	-8.34#
39 L8 AR-1262-4	2.164	0.000 E3	100.0#	0#	-8.42#
40 L8 AR-1262-5	1.467	0.000 E3	100.0#	0#	-9.02#
41 L9 AR-1268-1	5.168	0.000 E3	100.0#	0#	-8.33#
42 L9 AR-1268-2	4.702	0.000 E3	100.0#	0#	-8.42#
43 L9 AR-1268-3	3.855	0.000 E3	100.0#	0#	-8.63#
44 L9 AR-1268-4	1.681	0.000 E3	100.0#	0#	-9.02#
45 L9 AR-1268-5	11.867	0.000 E3	100.0#	0#	-9.39#

Signal #2

8 L2 AR-1221-1	279.903	0.000	100.0#	0#	-3.67#
9 L2 AR-1221-2	212.590	0.000	100.0#	0#	-3.75#
10 L2 AR-1221-3	706.539	0.000	100.0#	0#	-3.83#
11 L3 AR-1232-1	1.118	0.000 E3	100.0#	0#	-3.83#
12 L3 AR-1232-2	1.203	0.000 E3	100.0#	0#	-4.54#
13 L3 AR-1232-3	622.398	0.000	100.0#	0#	-4.71#
14 L3 AR-1232-4	569.700	0.000	100.0#	0#	-4.80#
15 L3 AR-1232-5	634.688	0.000	100.0#	0#	-4.96#
16 L4 AR-1242-1	868.924	0.000	100.0#	0#	-4.52#
17 L4 AR-1242-2	1.193	0.000 E3	100.0#	0#	-4.54#
18 L4 AR-1242-3	634.133	0.000	100.0#	0#	-4.71#
19 L4 AR-1242-4	650.899	0.000	100.0#	0#	-4.80#
20 L4 AR-1242-5	888.950	0.000	100.0#	0#	-5.31#
21 L5 AR-1248-1	715.816	0.000	100.0#	0#	-4.52#
22 L5 AR-1248-2	981.897	0.000	100.0#	0#	-4.75#
23 L5 AR-1248-3	1.018	0.000 E3	100.0#	0#	-4.79#
24 L5 AR-1248-4	1.287	0.000 E3	100.0#	0#	-4.96#
25 L5 AR-1248-5	1.901	0.000 E3	100.0#	0#	-5.31#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121319\
 Data File : P0064587.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Dec 2019 15:40
 Operator : SM/AJ
 Sample : P0121319ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO121319

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 08:25:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
26 L6 AR-1254-1	2.031	0.000 E3	100.0#	0#	-5.31#
27 L6 AR-1254-2	1.823	0.000 E3	100.0#	0#	-5.46#
28 L6 AR-1254-3	3.106	0.000 E3	100.0#	0#	-5.86#
29 L6 AR-1254-4	2.054	0.000 E3	100.0#	0#	-6.08#
30 L6 AR-1254-5	2.985	0.000 E3	100.0#	0#	-6.50#
36 L8 AR-1262-1	1.147	0.000 E3	100.0#	0#	-6.59#
37 L8 AR-1262-2	5.464	0.000 E3	100.0#	0#	-7.03#
38 L8 AR-1262-3	2.083	0.000 E3	100.0#	0#	-7.31#
39 L8 AR-1262-4	3.981	0.000 E3	100.0#	0#	-7.38#
40 L8 AR-1262-5	1.744	0.000 E3	100.0#	0#	-7.87#
41 L9 AR-1268-1	6.395	0.000 E3	100.0#	0#	-7.31#
42 L9 AR-1268-2	5.828	0.000 E3	100.0#	0#	-7.38#
43 L9 AR-1268-3	4.777	0.000 E3	100.0#	0#	-7.58#
44 L9 AR-1268-4	2.013	0.000 E3	100.0#	0#	-7.87#
45 L9 AR-1268-5	14.466	0.000 E3	100.0#	0#	-8.14#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0