

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0121319.M
 Title : GC EXTRACTABLES
 Last Update : Fri Dec 13 08:19:24 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0064577.D 1000 =P0064575.D 750 =P0064576.D
 250 =P0064578.D 50 =P0064579.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.401	2.571	2.393	2.396	2.411	2.435 E4	3.15
2) SA	Decachlorobiphenyl	2.947	2.983	2.863	3.070	3.482	3.069 E4	7.91
3) L1	AR-1016-1	0.956	0.997	0.940	1.006	1.090	0.998 E3	5.88
4) L1	AR-1016-2	1.392	1.469	1.379	1.442	1.580	1.452 E3	5.51
5) L1	AR-1016-3	8.554	8.820	8.377	8.819	9.670	8.848 E2	5.61
6) L1	AR-1016-4	7.027	7.363	6.964	7.056	7.806	7.243 E2	4.84
7) L1	AR-1016-5	7.159	7.415	7.051	7.381	7.432	7.288 E2	2.36
8) L2	AR-1221-1	2.592					2.592 E2	0.00
9) L2	AR-1221-2	1.976					1.976 E2	0.00
10) L2	AR-1221-3	6.624					6.624 E2	0.00
11) L3	AR-1232-1	1.011					1.011 E3	0.00
12) L3	AR-1232-2	5.109					5.109 E2	0.00
13) L3	AR-1232-3	1.118					1.118 E3	0.00
14) L3	AR-1232-4	5.559					5.559 E2	0.00
15) L3	AR-1232-5	4.150					4.150 E2	0.00
16) L4	AR-1242-1	7.837					7.837 E2	0.00
17) L4	AR-1242-2	1.128					1.128 E3	0.00
18) L4	AR-1242-3	6.963					6.963 E2	0.00
19) L4	AR-1242-4	5.638					5.638 E2	0.00
20) L4	AR-1242-5	6.969					6.969 E2	0.00
21) L5	AR-1248-1	6.454					6.454 E2	0.00
22) L5	AR-1248-2	9.241					9.241 E2	0.00
23) L5	AR-1248-3	1.062					1.062 E3	0.00
24) L5	AR-1248-4	1.304					1.304 E3	0.00
25) L5	AR-1248-5	1.244					1.244 E3	0.00
26) L6	AR-1254-1	1.274					1.274 E3	0.00
27) L6	AR-1254-2	2.080					2.080 E3	0.00
28) L6	AR-1254-3	2.388					2.388 E3	0.00
29) L6	AR-1254-4	1.870					1.870 E3	0.00
30) L6	AR-1254-5	2.104					2.104 E3	0.00
31) L7	AR-1260-1	1.487	1.545	1.464	1.525	1.661	1.536 E3	4.99
32) L7	AR-1260-2	1.870	1.949	1.860	1.963	2.228	1.974 E3	7.55
33) L7	AR-1260-3	1.462	1.520	1.445	1.520	1.566	1.503 E3	3.26
34) L7	AR-1260-4	1.775	1.838	1.746	1.850	2.027	1.847 E3	5.94
35) L7	AR-1260-5	3.608	3.793	3.572	3.648	4.037	3.732 E3	5.10
36) L8	AR-1262-1	2.299					2.299 E3	0.00
37) L8	AR-1262-2	4.341					4.341 E3	0.00
38) L8	AR-1262-3	2.910					2.910 E3	0.00
39) L8	AR-1262-4	2.164					2.164 E3	0.00
40) L8	AR-1262-5	1.467					1.467 E3	0.00
41) L9	AR-1268-1	5.168					5.168 E3	0.00
42) L9	AR-1268-2	4.702					4.702 E3	0.00
43) L9	AR-1268-3	3.855					3.855 E3	0.00
44) L9	AR-1268-4	1.681					1.681 E3	0.00
45) L9	AR-1268-5	1.187					1.187 E4	0.00

Signal #2 Calibration Files

500 =P0064577.D 1000 =P0064575.D 750 =P0064576.D
 250 =P0064578.D 50 =P0064579.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.495	2.709	2.500	2.476	2.495	2.535 E4	3.85
2) SA	Decachlorobiphenyl	3.331	3.408	3.259	3.442	3.925	3.473 E4	7.56
3) L1	AR-1016-1	1.037	1.099	1.016	1.087	1.168	1.081 E3	5.48
4) L1	AR-1016-2	1.430	1.569	1.440	1.490	1.591	1.504 E3	4.89
5) L1	AR-1016-3	7.589	8.047	7.483	7.728	8.022	7.774 E2	3.26
6) L1	AR-1016-4	6.336	6.583	6.207	6.565	7.183	6.575 E2	5.70

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0121319.M
 Title : GC EXTRACTABLES
 Last Update : Fri Dec 13 08:19:24 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0064577.D 1000 =P0064575.D 750 =P0064576.D
 250 =P0064578.D 50 =P0064579.D

Compound			500	1000	750	250	50	Avg		%RSD

7)	L1	AR-1016-5	0.817	0.880	0.809	0.840	1.016	0.873	E3	9.74
8)	L2	AR-1221-1	2.799					2.799	E2	0.00
9)	L2	AR-1221-2	2.126					2.126	E2	0.00
10)	L2	AR-1221-3	7.065					7.065	E2	0.00
11)	L3	AR-1232-1	1.118					1.118	E3	0.00
12)	L3	AR-1232-2	1.203					1.203	E3	0.00
13)	L3	AR-1232-3	6.224					6.224	E2	0.00
14)	L3	AR-1232-4	5.697					5.697	E2	0.00
15)	L3	AR-1232-5	6.347					6.347	E2	0.00
16)	L4	AR-1242-1	8.689					8.689	E2	0.00
17)	L4	AR-1242-2	1.193					1.193	E3	0.00
18)	L4	AR-1242-3	6.341					6.341	E2	0.00
19)	L4	AR-1242-4	6.509					6.509	E2	0.00
20)	L4	AR-1242-5	8.890					8.890	E2	0.00
21)	L5	AR-1248-1	7.158					7.158	E2	0.00
22)	L5	AR-1248-2	9.819					9.819	E2	0.00
23)	L5	AR-1248-3	1.018					1.018	E3	0.00
24)	L5	AR-1248-4	1.287					1.287	E3	0.00
25)	L5	AR-1248-5	1.901					1.901	E3	0.00
26)	L6	AR-1254-1	2.031					2.031	E3	0.00
27)	L6	AR-1254-2	1.823					1.823	E3	0.00
28)	L6	AR-1254-3	3.106					3.106	E3	0.00
29)	L6	AR-1254-4	2.054					2.054	E3	0.00
30)	L6	AR-1254-5	2.985					2.985	E3	0.00
31)	L7	AR-1260-1	1.730	1.814	1.726	1.841	2.122	1.847	E3	8.77
32)	L7	AR-1260-2	2.187	2.304	2.189	2.320	2.633	2.327	E3	7.84
33)	L7	AR-1260-3	2.024	2.146	2.007	2.133	2.278	2.117	E3	5.16
34)	L7	AR-1260-4	1.799	1.883	1.792	1.870	2.075	1.884	E3	6.07
35)	L7	AR-1260-5	4.471	4.744	4.483	4.591	5.162	4.690	E3	6.09
36)	L8	AR-1262-1	1.147					1.147	E3	0.00
37)	L8	AR-1262-2	5.464					5.464	E3	0.00
38)	L8	AR-1262-3	2.083					2.083	E3	0.00
39)	L8	AR-1262-4	3.981					3.981	E3	0.00
40)	L8	AR-1262-5	1.744					1.744	E3	0.00
41)	L9	AR-1268-1	6.395					6.395	E3	0.00
42)	L9	AR-1268-2	5.828					5.828	E3	0.00
43)	L9	AR-1268-3	4.777					4.777	E3	0.00
44)	L9	AR-1268-4	2.013					2.013	E3	0.00
45)	L9	AR-1268-5	1.447					1.447	E4	0.00

(#) = Out of Range