

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121419\
 Data File : P0064647.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Dec 2019 2:46
 Operator : SM/AJ
 Sample : K6143-12
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 10F

Manual Integrations
 APPROVED

Ankita
 12/16/2019 10:10:42 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 03:41:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.182	3.460	392836	495341	16.136m	19.541
2)	SA Decachlor...	9.702	8.386	333597	448146	10.870	12.904
Target Compounds							
26)	L6 AR-1254-1	6.177	5.317	502737	797038	394.756	392.489
27)	L6 AR-1254-2	6.394	5.464	1270161	1542148	610.640	845.762 #
28)	L6 AR-1254-3	6.757	5.864	1618944	3740083	678.009	1203.982 #
29)	L6 AR-1254-4	7.041	6.090	1376041	1673787	735.875	814.879
30)	L6 AR-1254-5	7.456	6.505	5118151	6112854	2432.554	2047.937
31)	L7 AR-1260-1	6.919	5.992	2419074	3464666	1574.603	1876.342
32)	L7 AR-1260-2	7.175	6.179	3735196	5070835	1892.198	2179.498
33)	L7 AR-1260-3	7.531	6.331	3924940	3548065	2612.081	1675.735 #
34)	L7 AR-1260-4	7.753	6.799	4426892	4311233	2396.512	2288.464
35)	L7 AR-1260-5	8.060	7.041	7227947	10777013	1936.901	2297.859

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121419\
Data File : PO064647.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Dec 2019 2:46
Operator : SM/AJ
Sample : K6143-12
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 03:41:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.182 min
Delta R.T.: 0.005 min
Response: 392836
Conc: 16.14 ng/ml m

Instrument :
ECD_O
ClientSampleId :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM




#1 Tetrachloro-m-xylene

R.T.: 3.460 min
Delta R.T.: 0.003 min
Response: 495341
Conc: 19.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.702 min
Delta R.T.: 0.014 min
Response: 333597
Conc: 10.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.386 min
Delta R.T.: 0.010 min
Response: 448146
Conc: 12.90 ng/ml




#26 AR-1254-1

R.T.: 6.177 min
Delta R.T.: 0.007 min
Response: 502737
Conc: 394.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM




#26 AR-1254-1

R.T.: 5.317 min
Delta R.T.: 0.005 min
Response: 797038
Conc: 392.49 ng/ml




#27 AR-1254-2

R.T.: 6.394 min
Delta R.T.: 0.007 min
Response: 1270161
Conc: 610.64 ng/ml



#27 AR-1254-2

R.T.: 5.464 min
Delta R.T.: 0.005 min
Response: 1542148
Conc: 845.76 ng/ml




#28 AR-1254-3

R.T.: 6.757 min
Delta R.T.: 0.009 min
Response: 1618944
Conc: 678.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM




#28 AR-1254-3

R.T.: 5.864 min
Delta R.T.: 0.007 min
Response: 3740083
Conc: 1203.98 ng/ml




#29 AR-1254-4

R.T.: 7.041 min
Delta R.T.: 0.008 min
Response: 1376041
Conc: 735.87 ng/ml



#29 AR-1254-4

R.T.: 6.090 min
Delta R.T.: 0.005 min
Response: 1673787
Conc: 814.88 ng/ml




#30 AR-1254-5

R.T.: 7.456 min
Delta R.T.: 0.008 min
Response: 5118151
Conc: 2432.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM




#30 AR-1254-5

R.T.: 6.505 min
Delta R.T.: 0.006 min
Response: 6112854
Conc: 2047.94 ng/ml




#31 AR-1260-1

R.T.: 6.919 min
Delta R.T.: 0.008 min
Response: 2419074
Conc: 1574.60 ng/ml



#31 AR-1260-1

R.T.: 5.992 min
Delta R.T.: 0.006 min
Response: 3464666
Conc: 1876.34 ng/ml




#32 AR-1260-2

R.T.: 7.175 min
Delta R.T.: 0.008 min
Response: 3735196
Conc: 1892.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM




#32 AR-1260-2

R.T.: 6.179 min
Delta R.T.: 0.006 min
Response: 5070835
Conc: 2179.50 ng/ml




#33 AR-1260-3

R.T.: 7.531 min
Delta R.T.: 0.010 min
Response: 3924940
Conc: 2612.08 ng/ml



#33 AR-1260-3

R.T.: 6.331 min
Delta R.T.: 0.007 min
Response: 3548065
Conc: 1675.73 ng/ml




#34 AR-1260-4

R.T.: 7.753 min
Delta R.T.: 0.009 min
Response: 4426892
Conc: 2396.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
10F


Manual Integrations
APPROVED

Ankita
12/16/2019 10:10:42 AM




#34 AR-1260-4

R.T.: 6.799 min
Delta R.T.: 0.007 min
Response: 4311233
Conc: 2288.46 ng/ml



#35 AR-1260-5

R.T.: 8.060 min
Delta R.T.: 0.009 min
Response: 7227947
Conc: 1936.90 ng/ml



#35 AR-1260-5

R.T.: 7.041 min
Delta R.T.: 0.008 min
Response: 10777013
Conc: 2297.86 ng/ml