

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121619\
 Data File : P0064723.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Dec 2019 20:00
 Operator : SM/AJ
 Sample : K6309-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 30786

Manual Integrations
 APPROVED

Ankita
 12/17/2019 10:46:03 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 00:41:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.179	3.455	367870	1028229	15.111	40.562m#
2) SA Decachlor...	9.696	8.375	696092	840633	22.681	24.206

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121619\
Data File : P0064723.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Dec 2019 20:00
Operator : SM/AJ
Sample : K6309-03
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
30786


Manual Integrations
APPROVED

Ankita
12/17/2019 10:46:03 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 00:41:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.179 min
Delta R.T.: 0.002 min
Response: 367870
Conc: 15.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
30786


Manual Integrations
APPROVED

Ankita
12/17/2019 10:46:03 AM




#1 Tetrachloro-m-xylene

R.T.: 3.455 min
Delta R.T.: -0.002 min
Response: 1028229
Conc: 40.56 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.696 min
Delta R.T.: 0.009 min
Response: 696092
Conc: 22.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.375 min
Delta R.T.: -0.001 min
Response: 840633
Conc: 24.21 ng/ml