

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121719\
 Data File : P0064739.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Dec 2019 11:04
 Operator : SM/AJ
 Sample : K6294-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-1MS

Manual Integrations
 APPROVED

Ankita
 12/18/2019 9:24:45 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 17 12:29:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.180	3.454	651238	773667	26.750m	30.520m
2)	SA Decachlor...	9.705	8.379	805259	973180	26.238	28.023
Target Compounds							
3)	L1 AR-1016-1	5.334	4.520	285021	349747	285.640	323.395
4)	L1 AR-1016-2	5.356	4.538	420477	493195	289.529	327.948
5)	L1 AR-1016-3	5.417	4.711	253220	259382	286.189	333.668
6)	L1 AR-1016-4	5.514	4.753	204029	221107	281.682	336.289
7)	L1 AR-1016-5	5.805	4.963	213629	277028	293.135	317.489m
31)	L7 AR-1260-1	6.920	5.986	404717	553920	263.435m	299.984
32)	L7 AR-1260-2	7.176	6.174	532529	711063	269.772m	305.623
33)	L7 AR-1260-3	7.532	6.325	386901	624036	257.486	294.729
34)	L7 AR-1260-4	7.756	6.793	465664	542040	252.089	287.723
35)	L7 AR-1260-5	8.062	7.035	901083	1263400	241.467	269.380

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121719\
Data File : PO064739.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Dec 2019 11:04
Operator : SM/AJ
Sample : K6294-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 17 12:29:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.180 min
Delta R.T.: 0.003 min
Response: 651238
Conc: 26.75 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM




#1 Tetrachloro-m-xylene

R.T.: 3.454 min
Delta R.T.: -0.003 min
Response: 773667
Conc: 30.52 ng/ml m




#2 Decachlorobiphenyl

R.T.: 9.705 min
Delta R.T.: 0.017 min
Response: 805259
Conc: 26.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.379 min
Delta R.T.: 0.003 min
Response: 973180
Conc: 28.02 ng/ml




#3 AR-1016-1

R.T.: 5.334 min
Delta R.T.: 0.007 min
Response: 285021
Conc: 285.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM




#3 AR-1016-1

R.T.: 4.520 min
Delta R.T.: -0.001 min
Response: 349747
Conc: 323.40 ng/ml




#4 AR-1016-2

R.T.: 5.356 min
Delta R.T.: 0.006 min
Response: 420477
Conc: 289.53 ng/ml



#4 AR-1016-2

R.T.: 4.538 min
Delta R.T.: -0.001 min
Response: 493195
Conc: 327.95 ng/ml




#5 AR-1016-3

R.T.: 5.417 min
Delta R.T.: 0.007 min
Response: 253220
Conc: 286.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM




#5 AR-1016-3

R.T.: 4.711 min
Delta R.T.: -0.001 min
Response: 259382
Conc: 333.67 ng/ml




#6 AR-1016-4

R.T.: 5.514 min
Delta R.T.: 0.006 min
Response: 204029
Conc: 281.68 ng/ml



#6 AR-1016-4

R.T.: 4.753 min
Delta R.T.: -0.001 min
Response: 221107
Conc: 336.29 ng/ml




#7 AR-1016-5

R.T.: 5.805 min
Delta R.T.: 0.008 min
Response: 213629
Conc: 293.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM




#7 AR-1016-5

R.T.: 4.963 min
Delta R.T.: -0.001 min
Response: 277028
Conc: 317.49 ng/ml m




#31 AR-1260-1

R.T.: 6.920 min
Delta R.T.: 0.009 min
Response: 404717
Conc: 263.43 ng/ml m



#31 AR-1260-1

R.T.: 5.986 min
Delta R.T.: 0.000 min
Response: 553920
Conc: 299.98 ng/ml




#32 AR-1260-2

R.T.: 7.176 min
Delta R.T.: 0.009 min
Response: 532529
Conc: 269.77 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM




#32 AR-1260-2

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 711063
Conc: 305.62 ng/ml




#33 AR-1260-3

R.T.: 7.532 min
Delta R.T.: 0.011 min
Response: 386901
Conc: 257.49 ng/ml



#33 AR-1260-3

R.T.: 6.325 min
Delta R.T.: 0.000 min
Response: 624036
Conc: 294.73 ng/ml




#34 AR-1260-4

R.T.: 7.756 min
Delta R.T.: 0.011 min
Response: 465664
Conc: 252.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1MS


Manual Integrations
APPROVED

Ankita
12/18/2019 9:24:45 AM




#34 AR-1260-4

R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 542040
Conc: 287.72 ng/ml



#35 AR-1260-5

R.T.: 8.062 min
Delta R.T.: 0.012 min
Response: 901083
Conc: 241.47 ng/ml



#35 AR-1260-5

R.T.: 7.035 min
Delta R.T.: 0.002 min
Response: 1263400
Conc: 269.38 ng/ml