

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0121719\
 Data File : P0064757.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Dec 2019 17:16
 Operator : SM/AJ
 Sample : K6325-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RT-3860

Manual Integrations
 APPROVED

Ankita
 12/18/2019 9:25:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 01:00:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.179	3.454	595610	725609	24.465	28.624
2)	SA Decachlor...	9.702	8.377	609692	757269	19.866	21.805
Target Compounds							
31)	L7 AR-1260-1	6.917	5.985	32079	74241	20.881m	40.206 #
32)	L7 AR-1260-2	7.171	6.172	61986	65370	31.401m	28.097m
33)	L7 AR-1260-3	7.529	6.323	31023	50155	20.646	23.688
34)	L7 AR-1260-4	7.751	6.791	42040	34940	22.759m	18.547m
35)	L7 AR-1260-5	8.059	7.033	45360	76820	12.155	16.380m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO121719\
Data File : PO064757.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Dec 2019 17:16
Operator : SM/AJ
Sample : K6325-01
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
RT-3860


Manual Integrations
APPROVED

Ankita
12/18/2019 9:25:32 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 01:00:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.179 min
Delta R.T.: 0.002 min
Response: 595610
Conc: 24.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
RT-3860


Manual Integrations
APPROVED

Ankita
12/18/2019 9:25:32 AM




#1 Tetrachloro-m-xylene

R.T.: 3.454 min
Delta R.T.: -0.002 min
Response: 725609
Conc: 28.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.702 min
Delta R.T.: 0.015 min
Response: 609692
Conc: 19.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.377 min
Delta R.T.: 0.001 min
Response: 757269
Conc: 21.81 ng/ml




#31 AR-1260-1

R.T.: 6.917 min
Delta R.T.: 0.007 min
Response: 32079
Conc: 20.88 ng/ml m

Instrument :
ECD_O
ClientSampleId :
RT-3860


Manual Integrations
APPROVED

Ankita
12/18/2019 9:25:32 AM




#31 AR-1260-1

R.T.: 5.985 min
Delta R.T.: -0.001 min
Response: 74241
Conc: 40.21 ng/ml




#32 AR-1260-2

R.T.: 7.171 min
Delta R.T.: 0.004 min
Response: 61986
Conc: 31.40 ng/ml m



#32 AR-1260-2

R.T.: 6.172 min
Delta R.T.: -0.001 min
Response: 65370
Conc: 28.10 ng/ml m




#33 AR-1260-3

R.T.: 7.529 min
Delta R.T.: 0.007 min
Response: 31023
Conc: 20.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
RT-3860


Manual Integrations
APPROVED

Ankita
12/18/2019 9:25:32 AM




#33 AR-1260-3

R.T.: 6.323 min
Delta R.T.: -0.001 min
Response: 50155
Conc: 23.69 ng/ml




#34 AR-1260-4

R.T.: 7.751 min
Delta R.T.: 0.007 min
Response: 42040
Conc: 22.76 ng/ml m



#34 AR-1260-4

R.T.: 6.791 min
Delta R.T.: -0.001 min
Response: 34940
Conc: 18.55 ng/ml m




#35 AR-1260-5

R.T.: 8.059 min
Delta R.T.: 0.009 min
Response: 45360
Conc: 12.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
RT-3860

Manual Integrations
APPROVED

Ankita
12/18/2019 9:25:32 AM



#35 AR-1260-5

R.T.: 7.033 min
Delta R.T.: 0.000 min
Response: 76820
Conc: 16.38 ng/ml m