

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122419\
 Data File : P0064920.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Dec 2019 14:24
 Operator : SM/AJ
 Sample : K6376-05 1000X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PCB-121719-05

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 25 04:08:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0121319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 08:19:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
26)	L6	AR-1254-1	6.175	5.307	13856704 28323005	10880.492 13947.234 #
27)	L6	AR-1254-2	6.392	5.454	33280027 37995579	15999.648 20837.955 #
28)	L6	AR-1254-3	6.755	5.852	52443445 82152270	21963.160 26445.904
29)	L6	AR-1254-4	7.039	6.080	49648071 65830570	26550.648 32049.441
30)	L6	AR-1254-5	7.455	6.493	67846287 116.4E6	32245.974 39008.524



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122419\
Data File : P0064920.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Dec 2019 14:24
Operator : SM/AJ
Sample : K6376-05 1000X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PCB-121719-05

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 25 04:08:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO121319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 08:19:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.177 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PCB-121719-05




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.457 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.688 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.376 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.175 min
Delta R.T.: 0.005 min
Response: 13856704
Conc: 10880.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-121719-05




#26 AR-1254-1

R.T.: 5.307 min
Delta R.T.: -0.005 min
Response: 28323005
Conc: 13947.23 ng/ml




#27 AR-1254-2

R.T.: 6.392 min
Delta R.T.: 0.006 min
Response: 33280027
Conc: 15999.65 ng/ml



#27 AR-1254-2


R.T.: 5.454 min
Delta R.T.: -0.005 min
Response: 37995579
Conc: 20837.96 ng/ml



#28 AR-1254-3


R.T.: 6.755 min
Delta R.T.: 0.007 min
Response: 52443445
Conc: 21963.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-121719-05




#28 AR-1254-3

R.T.: 5.852 min
Delta R.T.: -0.005 min
Response: 82152270
Conc: 26445.90 ng/ml




#29 AR-1254-4

R.T.: 7.039 min
Delta R.T.: 0.007 min
Response: 49648071
Conc: 26550.65 ng/ml



#29 AR-1254-4


R.T.: 6.080 min
Delta R.T.: -0.005 min
Response: 65830570
Conc: 32049.44 ng/ml



#30 AR-1254-5

R.T.: 7.455 min
Delta R.T.: 0.007 min
Response: 67846287
Conc: 32245.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-121719-05



#30 AR-1254-5

R.T.: 6.493 min
Delta R.T.: -0.005 min
Response: 116435926
Conc: 39008.52 ng/ml