

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122618\
 Data File : P0052979.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Dec 2018 19:55
 Operator : SM/SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:23:31 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 24 02:43:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.237	3.267	30221263	6806516	21.961	22.999
2) SA Decachlor...	9.765	7.937	10835144	4013917	15.455	14.262
Target Compounds						
9) L2 AR-1221-2	4.539	0.000	344079	0	35.299	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
Data File : PO052979.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Dec 2018 19:55
Operator : SM/SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:23:31 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 24 02:43:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.237 min
Delta R.T.: -0.005 min
Response: 30221263
Conc: 21.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.267 min
Delta R.T.: 0.000 min
Response: 6806516
Conc: 23.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.765 min
Delta R.T.: -0.018 min
Response: 10835144
Conc: 15.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.937 min
Delta R.T.: -0.010 min
Response: 4013917
Conc: 14.26 ng/ml



#9 AR-1221-2

R.T.: 4.539 min
Delta R.T.: 0.001 min
Response: 344079
Conc: 35.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK



#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.544 min
Response: 0
Conc: N.D.