

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122618\
 Data File : P0052987.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Dec 2018 22:10
 Operator : SM/SJ
 Sample : J6533-10
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:27:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 24 02:43:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.238	3.268	21103915	4288488	15.335	14.491
2) SA Decachlor...	9.764	7.937	8756097	3627105	12.490	12.887
Target Compounds						
12) L3 AR-1232-2	5.166	0.000	350876	0	29.122	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
Data File : PO052987.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Dec 2018 22:10
Operator : SM/SJ
Sample : J6533-10
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:27:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 24 02:43:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.238 min
Delta R.T.: -0.004 min
Response: 21103915
Conc: 15.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-2




#1 Tetrachloro-m-xylene

R.T.: 3.268 min
Delta R.T.: 0.000 min
Response: 4288488
Conc: 14.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.764 min
Delta R.T.: -0.018 min
Response: 8756097
Conc: 12.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.937 min
Delta R.T.: -0.010 min
Response: 3627105
Conc: 12.89 ng/ml



#12 AR-1232-2

R.T.: 5.166 min
Delta R.T.: 0.034 min
Response: 350876
Conc: 29.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-2



#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.279 min
Response: 0
Conc: N.D.