

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
 Data File : PO052949.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Dec 2018 10:51
 Operator : SM/SJ
 Sample : J6191-01RE
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 NB-07-122018-ARE

Manual Integrations
 APPROVED

Sohil
 12/27/2018 9:47:37 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:09:46 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 24 02:43:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.237	3.267	18456574	4290852	13.412	14.499
2) SA Decachlor...	9.765	7.937	7233882	2653384	10.318	9.428m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
Data File : P0052949.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Dec 2018 10:51
Operator : SM/SJ
Sample : J6191-01RE
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
NB-07-122018-ARE


Manual Integrations
APPROVED

Sohil
12/27/2018 9:47:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:09:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 24 02:43:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.237 min
Delta R.T.: -0.005 min
Response: 18456574
Conc: 13.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
NB-07-122018-ARE


Manual Integrations
APPROVED

Sohil
12/27/2018 9:47:37 AM




#1 Tetrachloro-m-xylene

R.T.: 3.267 min
Delta R.T.: -0.001 min
Response: 4290852
Conc: 14.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.765 min
Delta R.T.: -0.018 min
Response: 7233882
Conc: 10.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.937 min
Delta R.T.: -0.010 min
Response: 2653384
Conc: 9.43 ng/ml m