

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
 Data File : PO052951.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Dec 2018 11:25
 Operator : SM/SJ
 Sample : J6200-01RE
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TR-05-122018RE

Manual Integrations
 APPROVED

Sohil
 12/27/2018 9:47:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:10:38 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 24 02:43:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.237	3.267	18544422	4504026	13.475	15.219
2) SA Decachlor...	9.766	7.937	7964767	2784349	11.361	9.893m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
Data File : PO052951.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Dec 2018 11:25
Operator : SM/SJ
Sample : J6200-01RE
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TR-05-122018RE


Manual Integrations
APPROVED

Sohil
12/27/2018 9:47:41 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:10:38 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 24 02:43:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.237 min
Delta R.T.: -0.005 min
Response: 18544422
Conc: 13.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
TR-05-122018RE


Manual Integrations
APPROVED

Sohil
12/27/2018 9:47:41 AM




#1 Tetrachloro-m-xylene

R.T.: 3.267 min
Delta R.T.: -0.001 min
Response: 4504026
Conc: 15.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.766 min
Delta R.T.: -0.016 min
Response: 7964767
Conc: 11.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.937 min
Delta R.T.: -0.010 min
Response: 2784349
Conc: 9.89 ng/ml m