

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122618\
 Data File : P0052972.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Dec 2018 17:57
 Operator : SM/SJ
 Sample : J6528-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 23738

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 10:33:53 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 24 02:43:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.239	3.267	8896358	5834009	6.465m	19.713m#
2) SA Decachlor...	9.766	7.938	8409923	2564771	11.996m	9.113m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122618\
Data File : PO052972.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Dec 2018 17:57
Operator : SM/SJ
Sample : J6528-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
23738

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 10:33:53 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 24 02:43:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.239 min
Delta R.T.: -0.003 min
Response: 8896358
Conc: 6.46 ng/ml m

Instrument :
ECD_O
ClientSampleId :
23738




#1 Tetrachloro-m-xylene

R.T.: 3.267 min
Delta R.T.: -0.002 min
Response: 5834009
Conc: 19.71 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.766 min
Delta R.T.: -0.016 min
Response: 8409923
Conc: 12.00 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.938 min
Delta R.T.: -0.009 min
Response: 2564771
Conc: 9.11 ng/ml m