

Method Path : P:\HPCHEM1\ECD 0\methods\
 Method File : AR1232P0122917CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Dec 29 06:39:59 2017
 Response Via : Initial Calibration

Calibration Files

400 =P0040972.D 1600 =P0040974.D 800 =P0040973.D
 200 =P0040971.D 100 =P0040970.D

Compound			400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		1.932	1.709	1.757	2.050	2.257	1.941	E5	11.49
2) SA	Decachlorobiphenyl		2.371	2.016	2.123	2.573	2.863	2.389	E5	14.33
3) L1	AR-1016-1							0.000		-1.00
4) L1	AR-1016-2							0.000		-1.00
5) L1	AR-1016-3							0.000		-1.00
6) L1	AR-1016-4							0.000		-1.00
7) L1	AR-1016-5							0.000		-1.00
8) L2	AR-1221-1							0.000		-1.00
9) L2	AR-1221-2							0.000		-1.00
10) L2	AR-1221-3							0.000		-1.00
11) L3	AR-1232-1		3.427	2.903	3.134	3.637	3.907	3.402	E3	11.68
12) L3	AR-1232-2		1.906	1.610	1.730	2.121	2.356	1.944	E3	15.42
13) L3	AR-1232-3		2.648	2.162	2.373	2.970	3.227	2.676	E3	16.15
14) L3	AR-1232-4		1.968	1.716	1.810	1.998	2.134	1.925	E3	8.52
15) L3	AR-1232-5		1.338	1.147	1.237	1.404	1.381	1.301	E3	8.26
16) L4	AR-1242-1							0.000		-1.00
17) L4	AR-1242-2							0.000		-1.00
18) L4	AR-1242-3							0.000		-1.00
19) L4	AR-1242-4							0.000		-1.00
20) L4	AR-1242-5							0.000		-1.00
21) L5	AR-1248-1							0.000		-1.00
22) L5	AR-1248-2							0.000		-1.00
23) L5	AR-1248-3							0.000		-1.00
24) L5	AR-1248-4							0.000		-1.00
25) L5	AR-1248-5							0.000		-1.00
26) L6	AR-1254-1							0.000		-1.00
27) L6	AR-1254-2							0.000		-1.00
28) L6	AR-1254-3							0.000		-1.00
29) L6	AR-1254-4							0.000		-1.00
30) L6	AR-1254-5							0.000		-1.00
31) L7	AR-1260-1							0.000		-1.00
32) L7	AR-1260-2							0.000		-1.00
33) L7	AR-1260-3							0.000		-1.00
34) L7	AR-1260-4							0.000		-1.00
35) L7	AR-1260-5							0.000		-1.00
36) L8	AR-1262-1							0.000		-1.00
37) L8	AR-1262-2							0.000		-1.00
38) L8	AR-1262-3							0.000		-1.00
39) L8	AR-1262-4							0.000		-1.00
40) L8	AR-1262-5							0.000		-1.00
41) L9	AR-1268-1							0.000		-1.00
42) L9	AR-1268-2							0.000		-1.00
43) L9	AR-1268-3							0.000		-1.00
44) L9	AR-1268-4							0.000		-1.00
45) L9	AR-1268-5							0.000		-1.00

Signal #2 Calibration Files

400 =P0040972.D 1600 =P0040974.D 800 =P0040973.D
 200 =P0040971.D 100 =P0040970.D

Compound			400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		2.704	2.416	2.448	2.908	3.137	2.723	E5	11.27
2) SA	Decachlorobiphenyl		3.826	3.450	3.542	4.085	4.433	3.867	E5	10.41

Method Path : P:\HPCHEM1\ECD 0\methods\
 Method File : AR1232P0122917CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Dec 29 06:39:59 2017
 Response Via : Initial Calibration

Calibration Files

400 =P0040972.D 1600 =P0040974.D 800 =P0040973.D
 200 =P0040971.D 100 =P0040970.D

Compound			400	1600	800	200	100	Avg	%RSD

3)	L1	AR-1016-1						0.000	-1.00
4)	L1	AR-1016-2						0.000	-1.00
5)	L1	AR-1016-3						0.000	-1.00
6)	L1	AR-1016-4						0.000	-1.00
7)	L1	AR-1016-5						0.000	-1.00
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1	4.918	4.104	4.381	5.354	5.809	4.913 E3	14.15
12)	L3	AR-1232-2	2.762	2.379	2.504	2.885	3.148	2.735 E3	11.17
13)	L3	AR-1232-3	2.984	2.572	2.743	3.224	3.275	2.959 E3	10.24
14)	L3	AR-1232-4	3.025	2.599	2.797	3.365	3.726	3.102 E3	14.50
15)	L3	AR-1232-5	3.822	3.272	3.510	3.985	4.554	3.829 E3	12.81
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1						0.000	-1.00
32)	L7	AR-1260-2						0.000	-1.00
33)	L7	AR-1260-3						0.000	-1.00
34)	L7	AR-1260-4						0.000	-1.00
35)	L7	AR-1260-5						0.000	-1.00
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range