

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010219\
 Data File : PP028613.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jan 2019 14:34
 Operator : EI\MA
 Sample : J6188-06MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BU-03-122118MSD

Manual Integrations
 APPROVED

Sohil
 1/3/2019 8:59:52 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 00:23:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120718.M
 Quant Title : 8080.M
 QLast Update : Fri Dec 07 14:01:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.096	12.209	77874919	114.2E6	298.306	255.043
Target Compounds							
1) T	Dalapon	3.341	3.009	271.7E6	275.8E6	464.868	499.108m
2) T	3,5-DICHL...	10.507	10.300	173.7E6	191.3E6	349.052m	353.416
5) T	DICAMBA	12.418	12.519	570.2E6	636.1E6	350.939	364.277
6) T	MCP	12.763	12.709	21202552	55728114	48.774	62.954 #
7) T	MCPA	13.002	13.097	39686999	75944477	55.320	56.198m
8) T	DICHLORPROP	13.615	13.634	360.6E6	230.7E6	751.891m	532.972 #
9) T	2,4-D	13.965	14.120	220.9E6	300.1E6	591.505	591.060
10) T	Pentachlo...	14.384	14.734	2582.9E6	2253.7E6	319.156m	337.461
11) T	2,4,5-TP ...	15.245	15.298	1454.0E6	2272.2E6	495.035m	888.892m#
12) T	2,4,5-T	15.635	15.945	1288.4E6	718.3E6	484.270m	330.553 #
13) T	2,4-DB	16.400	16.545	124.5E6	149.6E6	588.708m	566.529
14) T	DINOSEB	17.890	16.939	594.9E6	497.3E6	280.166	283.709
15) T	Picloram	17.657	18.321	1409.5E6	1331.9E6	350.667	378.048
16) T	DCPA	18.254	18.199	1210.3E6	1038.3E6	281.966	317.235

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010219\
Data File : PP028613.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jan 2019 14:34
Operator : EI\MA
Sample : J6188-06MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
BU-03-122118MSD

Manual Integrations
APPROVED

Sohil
1/3/2019 8:59:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 00:23:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120718.M
Quant Title : 8080.M
QLast Update : Fri Dec 07 14:01:32 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

