

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP010319.M
 Title : 8080.M
 Last Update : Fri Jan 04 01:40:06 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP028620.D 1000 =PP028621.D 1500 =PP028622.D
 500 =PP028619.D 200 =PP028618.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	6.032	5.901	5.690	6.188	6.765	6.115 E5	6.65
2) T	3,5-DICHLOROBENZOI...	4.993	4.897	4.705	5.034	5.358	4.997 E5	4.77
3) T	4-Nitrophenol	4.069	4.183	4.209	3.818	3.496	3.955 E5	7.58
4) S	2,4-DCAA	2.573	2.511	2.444	2.601	2.726	2.571 E5	4.11
5) T	DICAMBA	1.574	1.545	1.500	1.576	1.633	1.566 E6	3.10
6) T	MCPP	5.072	5.183	5.209	4.963	4.041	4.894 E5	9.94
7) T	MCPA	8.147	8.048	7.918	7.996	7.145	7.851 E5	5.13
8) T	DICHLORPROP	3.212	3.106	3.001	3.406	3.450	3.235 E5	5.94
9) T	2,4-D	3.092	3.090	3.166	2.894	3.106	3.069 E5	3.36
10) T	Pentachlorophenol	7.912	7.571	7.025	8.157	8.591	7.851 E6	7.55
11) T	2,4,5-TP (SILVEX)	2.881	2.798	2.677	2.914	2.940	2.842 E6	3.74
12) T	2,4,5-T	2.614	2.554	2.461	2.609	2.373	2.522 E6	4.11
13) T	2,4-DB	1.836	1.882	1.918	1.639	1.123	1.680 E5	19.61
14) T	DINOSEB	2.159	2.079	2.015	2.226	2.463	2.188 E6	7.90
15) T	Picloram	4.364	4.342	4.273	4.287	3.565	4.166 E6	8.12
16) T	DCPA	4.195	4.070	3.859	4.220	4.299	4.129 E6	4.16

Signal #2 Calibration Files

750 =PP028620.D 1000 =PP028621.D 1500 =PP028622.D
 500 =PP028619.D 200 =PP028618.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.790	5.635	5.506	5.896	6.402	5.846 E5	5.89
2) T	3,5-DICHLOROBENZOI...	5.146	5.041	4.989	5.099	5.489	5.153 E5	3.83
3) T	4-Nitrophenol	2.072	2.072	2.094	1.971	1.625	1.967 E5	10.01
4) S	2,4-DCAA	4.379	4.277	4.192	4.390	4.591	4.366 E5	3.43
5) T	DICAMBA	1.741	1.691	1.637	1.748	1.818	1.727 E6	3.91
6) T	MCPP	9.671	9.518	9.043	9.444	8.570	9.249 E5	4.81
7) T	MCPA	1.499	1.482	1.479	1.464	1.392	1.463 E6	2.85
8) T	DICHLORPROP	4.352	4.190	4.091	4.390	4.689	4.343 E5	5.26
9) T	2,4-D	4.952	4.845	4.790	4.878	4.923	4.878 E5	1.31
10) T	Pentachlorophenol	6.538	6.165	5.693	6.818	7.296	6.502 E6	9.41
11) T	2,4,5-TP (SILVEX)	2.528	2.440	2.343	2.555	2.485	2.470 E6	3.37
12) T	2,4,5-T	2.237	2.172	2.106	2.224	2.107	2.169 E6	2.86
13) T	2,4-DB	2.436	2.494	2.555	2.269	2.134	2.378 E5	7.28
14) T	DINOSEB	1.763	1.702	1.678	1.775	1.855	1.755 E6	3.94
15) T	Picloram	3.834	3.714	3.614	3.860	3.594	3.723 E6	3.29
16) T	DCPA	3.288	3.125	2.943	3.400	3.573	3.266 E6	7.45

(#) = Out of Range