Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010421\
Data File : PO074365.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Jan 2021 9:19
Operator : DD\AJ
Sample : PO010421ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 00:11:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 05:03:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.942	3.990	4453639	2472196	46.350	46.410	
2)	SA Decachlor...	10.967	9.251	4485559	2615179	48.250	50.236	
	Target Compounds							
3)	L1 AR-1016-1	6.266	5.237	1749437	1015775	488.278	470.414	
4)	L1 AR-1016-2	6.291	5.257	2429083	1420775	488.979	473.325	
5)	L1 AR-1016-3	6.357	5.447	1457441	769936	492.646	480.930	
6)	L1 AR-1016-4	6.464	5.500	1228922	633601	494.109	500.927	
7)	L1 AR-1016-5	6.780	5.727	1193006	786630	510.046	497.427	
8)	L2 AR-1221-1	5.189	4.243	146401	95764	129.725	141.356	
9)	L2 AR-1221-2	5.293	4.342	213175	137141	258.639	267.760	
10)	L2 AR-1221-3	5.380	4.430	818703	496104	310.382	319.537	
11)	L3 AR-1232-1	5.380	4.430	818703	496104	391.759	399.941	
12)	L3 AR-1232-2	5.970	5.257	1146265	1420775	1087.968	1098.744	
13)	L3 AR-1232-3	6.291	5.447	2429083	769936	1106.543	1108.238	
14)	L3 AR-1232-4	6.464	5.544	1228922	732395	1126.387	1225.860	
15)	L3 AR-1232-5	6.562	5.727	982903	786630	1303.429	1214.762	
16)	L4 AR-1242-1	6.266	5.237	1749437	1015775	598.603	572.735	
17)	L4 AR-1242-2	6.291	5.257	2429083	1420775	591.771	583.785	
18)	L4 AR-1242-3	6.357	5.447	1457441	769936	589.049	581.441	
19)	L4 AR-1242-4	6.464	5.544	1228922	732395	596.442	582.312	
20)	L4 AR-1242-5	7.250	6.104	193220	606946	84.588	357.874	#
21)	L5 AR-1248-1	6.266	5.237	1749437	1015775	791.914	739.654	
22)	L5 AR-1248-2	6.562	5.500	982903	633601	342.010	343.112	
23)	L5 AR-1248-3	6.780	5.544	1193006	732395	353.999	394.638	
24)	L5 AR-1248-4	7.209	5.727	246263	786630	58.390	354.964	#
25)	L5 AR-1248-5	7.250	6.146	193220	109183	48.397	46.223	
26)	L6 AR-1254-1	7.180	6.104	838234	606946	203.396	167.613	
27)	L6 AR-1254-2	7.409	6.262	941451	563201	144.708	178.086	
28)	L6 AR-1254-3	7.793	6.698f	517720	1152071	78.322	229.599	#
29)	L6 AR-1254-4	8.088	6.918	418464	193763	73.547	53.049	#
30)	L6 AR-1254-5	8.517	7.343	3234156	2064643	597.691	438.116	#
31)	L7 AR-1260-1	7.960	6.816	2154892	1483514	520.833	498.107	
32)	L7 AR-1260-2	8.225	7.012	2989646	2025684	515.256	497.137	
33)	L7 AR-1260-3	8.592	7.166	2417158	1708176	465.621	490.717	
34)	L7 AR-1260-4	8.823	7.645	2385421	1493135	477.745	522.221	
35)	L7 AR-1260-5	9.150	7.892	5456846	3907400	481.308	508.812	
36)	L8 AR-1262-1	8.592	7.343	2417158	2064643	290.202	797.573	#
37)	L8 AR-1262-2	9.150	7.892	5456846	3907400	378.138	400.213	
38)	L8 AR-1262-3	0.000	8.177	0	921290	N.D.	232.343	#
39)	L8 AR-1262-4	9.536	8.239	2173427	2647773	300.548	396.862	#
40)	L8 AR-1262-5	10.221	8.735	1608606	999344	319.235	329.085	
41)	L9 AR-1268-1	0.000	8.177	0	921290	N.D.	77.117	#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010421\
Data File : PO074365.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Jan 2021 9:19
Operator : DD\AJ
Sample : PO010421ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 00:11:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 05:03:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.536	8.239	2173427	2647773	140.942	266.732 #
43)	L9 AR-1268-3	9.781	8.447	109811	45814	7.992	5.330 #
44)	L9 AR-1268-4	10.221	8.735	1608606	999344	281.634	286.512
45)	L9 AR-1268-5	10.632	9.011	414053	248379	10.164	10.143

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010421\
Data File : PO074365.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Jan 2021 9:19
Operator : DD\AJ
Sample : PO010421ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 00:11:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 05:03:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.942 min
Delta R.T.: 0.009 min
Response: 4453639
Conc: 46.35 ng/ml

Instrument :
ECD_P
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 3.990 min
Delta R.T.: 0.000 min
Response: 2472196
Conc: 46.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.967 min
Delta R.T.: 0.012 min
Response: 4485559
Conc: 48.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.251 min
Delta R.T.: 0.001 min
Response: 2615179
Conc: 50.24 ng/ml



#3 AR-1016-1

R.T.: 6.266 min
Delta R.T.: 0.010 min
Response: 1749437
Conc: 488.28 ng/ml

Instrument :
ECD_P
ClientSampleId :



#3 AR-1016-1

R.T.: 5.237 min
Delta R.T.: 0.000 min
Response: 1015775
Conc: 470.41 ng/ml



#4 AR-1016-2

R.T.: 6.291 min
Delta R.T.: 0.010 min
Response: 2429083
Conc: 488.98 ng/ml



#4 AR-1016-2

R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 1420775
Conc: 473.33 ng/ml



#5 AR-1016-3

R.T.: 6.357 min
Delta R.T.: 0.010 min
Response: 1457441
Conc: 492.65 ng/ml

Instrument :
ECD_P
ClientSampleId :



#5 AR-1016-3

R.T.: 5.447 min
Delta R.T.: 0.000 min
Response: 769936
Conc: 480.93 ng/ml



#6 AR-1016-4

R.T.: 6.464 min
Delta R.T.: 0.010 min
Response: 1228922
Conc: 494.11 ng/ml



#6 AR-1016-4

R.T.: 5.500 min
Delta R.T.: 0.000 min
Response: 633601
Conc: 500.93 ng/ml



#7 AR-1016-5

R.T.: 6.780 min
Delta R.T.: 0.010 min
Response: 1193006
Conc: 510.05 ng/ml

Instrument :
ECD_P
ClientSampleId :



#7 AR-1016-5

R.T.: 5.727 min
Delta R.T.: 0.001 min
Response: 786630
Conc: 497.43 ng/ml



#8 AR-1221-1

R.T.: 5.189 min
Delta R.T.: 0.004 min
Response: 146401
Conc: 129.72 ng/ml



#8 AR-1221-1

R.T.: 4.243 min
Delta R.T.: -0.002 min
Response: 95764
Conc: 141.36 ng/ml



#9 AR-1221-2

R.T.: 5.293 min
Delta R.T.: 0.008 min
Response: 213175
Conc: 258.64 ng/ml

Instrument :
ECD_P
ClientSampleId :



#9 AR-1221-2

R.T.: 4.342 min
Delta R.T.: -0.001 min
Response: 137141
Conc: 267.76 ng/ml



#10 AR-1221-3

R.T.: 5.380 min
Delta R.T.: 0.009 min
Response: 818703
Conc: 310.38 ng/ml



#10 AR-1221-3

R.T.: 4.430 min
Delta R.T.: -0.001 min
Response: 496104
Conc: 319.54 ng/ml



#11 AR-1232-1

R.T.: 5.380 min
Delta R.T.: 0.010 min
Response: 818703
Conc: 391.76 ng/ml

Instrument :
ECD_P
ClientSampleId :



#11 AR-1232-1

R.T.: 4.430 min
Delta R.T.: 0.000 min
Response: 496104
Conc: 399.94 ng/ml



#12 AR-1232-2

R.T.: 5.970 min
Delta R.T.: 0.010 min
Response: 1146265
Conc: 1087.97 ng/ml



#12 AR-1232-2

R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 1420775
Conc: 1098.74 ng/ml



#13 AR-1232-3

R.T.: 6.291 min
Delta R.T.: 0.011 min
Response: 2429083
Conc: 1106.54 ng/ml

Instrument :
ECD_P
ClientSampleId :



#13 AR-1232-3

R.T.: 5.447 min
Delta R.T.: 0.001 min
Response: 769936
Conc: 1108.24 ng/ml



#14 AR-1232-4

R.T.: 6.464 min
Delta R.T.: 0.009 min
Response: 1228922
Conc: 1126.39 ng/ml



#14 AR-1232-4

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 732395
Conc: 1225.86 ng/ml



#15 AR-1232-5

R.T.: 6.562 min
Delta R.T.: 0.010 min
Response: 982903
Conc: 1303.43 ng/ml

Instrument :
ECD_P
ClientSampleId :



#15 AR-1232-5

R.T.: 5.727 min
Delta R.T.: 0.000 min
Response: 786630
Conc: 1214.76 ng/ml



#16 AR-1242-1

R.T.: 6.266 min
Delta R.T.: 0.009 min
Response: 1749437
Conc: 598.60 ng/ml



#16 AR-1242-1

R.T.: 5.237 min
Delta R.T.: 0.000 min
Response: 1015775
Conc: 572.73 ng/ml



#17 AR-1242-2

R.T.: 6.291 min
Delta R.T.: 0.011 min
Response: 2429083
Conc: 591.77 ng/ml

Instrument :
ECD_P
ClientSampleId :



#17 AR-1242-2

R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 1420775
Conc: 583.78 ng/ml



#18 AR-1242-3

R.T.: 6.357 min
Delta R.T.: 0.010 min
Response: 1457441
Conc: 589.05 ng/ml



#18 AR-1242-3

R.T.: 5.447 min
Delta R.T.: 0.001 min
Response: 769936
Conc: 581.44 ng/ml



#19 AR-1242-4

R.T.: 6.464 min
Delta R.T.: 0.010 min
Response: 1228922
Conc: 596.44 ng/ml

Instrument :
ECD_P
ClientSampleId :



#19 AR-1242-4

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 732395
Conc: 582.31 ng/ml



#20 AR-1242-5

R.T.: 7.250 min
Delta R.T.: 0.009 min
Response: 193220
Conc: 84.59 ng/ml



#20 AR-1242-5

R.T.: 6.104 min
Delta R.T.: 0.000 min
Response: 606946
Conc: 357.87 ng/ml



#21 AR-1248-1

R.T.: 6.266 min
Delta R.T.: 0.010 min
Response: 1749437
Conc: 791.91 ng/ml

Instrument :
ECD_P
ClientSampleId :



#21 AR-1248-1

R.T.: 5.237 min
Delta R.T.: 0.000 min
Response: 1015775
Conc: 739.65 ng/ml



#22 AR-1248-2

R.T.: 6.562 min
Delta R.T.: 0.011 min
Response: 982903
Conc: 342.01 ng/ml



#22 AR-1248-2

R.T.: 5.500 min
Delta R.T.: 0.000 min
Response: 633601
Conc: 343.11 ng/ml



#23 AR-1248-3

R.T.: 6.780 min
Delta R.T.: 0.010 min
Response: 1193006
Conc: 354.00 ng/ml

Instrument :
ECD_P
ClientSampleId :



#23 AR-1248-3

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 732395
Conc: 394.64 ng/ml



#24 AR-1248-4

R.T.: 7.209 min
Delta R.T.: 0.010 min
Response: 246263
Conc: 58.39 ng/ml



#24 AR-1248-4

R.T.: 5.727 min
Delta R.T.: 0.001 min
Response: 786630
Conc: 354.96 ng/ml



#25 AR-1248-5

R.T.: 7.250 min
Delta R.T.: 0.010 min
Response: 193220
Conc: 48.40 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 6.146 min
Delta R.T.: 0.000 min
Response: 109183
Conc: 46.22 ng/ml



#26 AR-1254-1

R.T.: 7.180 min
Delta R.T.: 0.010 min
Response: 838234
Conc: 203.40 ng/ml



#26 AR-1254-1

R.T.: 6.104 min
Delta R.T.: 0.000 min
Response: 606946
Conc: 167.61 ng/ml



#27 AR-1254-2

R.T.: 7.409 min
Delta R.T.: 0.009 min
Response: 941451
Conc: 144.71 ng/ml

Instrument :
ECD_P
ClientSampleId :



#27 AR-1254-2

R.T.: 6.262 min
Delta R.T.: 0.000 min
Response: 563201
Conc: 178.09 ng/ml



#28 AR-1254-3

R.T.: 7.793 min
Delta R.T.: 0.010 min
Response: 517720
Conc: 78.32 ng/ml



#28 AR-1254-3

R.T.: 6.698 min
Delta R.T.: 0.020 min
Response: 1152071
Conc: 229.60 ng/ml



#29 AR-1254-4

R.T.: 8.088 min
Delta R.T.: 0.008 min
Response: 418464
Conc: 73.55 ng/ml

Instrument :
ECD_P
ClientSampleId :



#29 AR-1254-4

R.T.: 6.918 min
Delta R.T.: 0.001 min
Response: 193763
Conc: 53.05 ng/ml



#30 AR-1254-5

R.T.: 8.517 min
Delta R.T.: 0.007 min
Response: 3234156
Conc: 597.69 ng/ml



#30 AR-1254-5

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 2064643
Conc: 438.12 ng/ml



#31 AR-1260-1

R.T.: 7.960 min
Delta R.T.: 0.009 min
Response: 2154892
Conc: 520.83 ng/ml

Instrument :
ECD_P
ClientSampleId :



#31 AR-1260-1

R.T.: 6.816 min
Delta R.T.: 0.001 min
Response: 1483514
Conc: 498.11 ng/ml



#32 AR-1260-2

R.T.: 8.225 min
Delta R.T.: 0.009 min
Response: 2989646
Conc: 515.26 ng/ml



#32 AR-1260-2

R.T.: 7.012 min
Delta R.T.: 0.001 min
Response: 2025684
Conc: 497.14 ng/ml



#33 AR-1260-3

R.T.: 8.592 min
Delta R.T.: 0.010 min
Response: 2417158
Conc: 465.62 ng/ml

Instrument :
ECD_P
ClientSampleId :



#33 AR-1260-3

R.T.: 7.166 min
Delta R.T.: 0.000 min
Response: 1708176
Conc: 490.72 ng/ml



#34 AR-1260-4

R.T.: 8.823 min
Delta R.T.: 0.009 min
Response: 2385421
Conc: 477.75 ng/ml



#34 AR-1260-4

R.T.: 7.645 min
Delta R.T.: 0.001 min
Response: 1493135
Conc: 522.22 ng/ml



#35 AR-1260-5

R.T.: 9.150 min
Delta R.T.: 0.009 min
Response: 5456846
Conc: 481.31 ng/ml

Instrument :
ECD_P
ClientSampleId :



#35 AR-1260-5

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 3907400
Conc: 508.81 ng/ml



#36 AR-1262-1

R.T.: 8.592 min
Delta R.T.: 0.008 min
Response: 2417158
Conc: 290.20 ng/ml



#36 AR-1262-1

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 2064643
Conc: 797.57 ng/ml



#37 AR-1262-2

R.T.: 9.150 min
Delta R.T.: 0.009 min
Response: 5456846
Conc: 378.14 ng/ml

Instrument :
ECD_P
ClientSampleId :



#37 AR-1262-2

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 3907400
Conc: 400.21 ng/ml



#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 9.451 min
Response: 0
Conc: N.D.



#38 AR-1262-3

R.T.: 8.177 min
Delta R.T.: 0.000 min
Response: 921290
Conc: 232.34 ng/ml



#39 AR-1262-4

R.T.: 9.536 min
Delta R.T.: -0.008 min
Response: 2173427
Conc: 300.55 ng/ml

Instrument :
ECD_P
ClientSampleId :



#39 AR-1262-4

R.T.: 8.239 min
Delta R.T.: 0.000 min
Response: 2647773
Conc: 396.86 ng/ml



#40 AR-1262-5

R.T.: 10.221 min
Delta R.T.: 0.009 min
Response: 1608606
Conc: 319.24 ng/ml



#40 AR-1262-5

R.T.: 8.735 min
Delta R.T.: 0.002 min
Response: 999344
Conc: 329.09 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 9.451 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#41 AR-1268-1

R.T.: 8.177 min
Delta R.T.: 0.000 min
Response: 921290
Conc: 77.12 ng/ml



#42 AR-1268-2

R.T.: 9.536 min
Delta R.T.: -0.011 min
Response: 2173427
Conc: 140.94 ng/ml



#42 AR-1268-2

R.T.: 8.239 min
Delta R.T.: -0.002 min
Response: 2647773
Conc: 266.73 ng/ml



#43 AR-1268-3

R.T.: 9.781 min
Delta R.T.: 0.009 min
Response: 109811
Conc: 7.99 ng/ml

Instrument :
ECD_P
ClientSampleId :



#43 AR-1268-3

R.T.: 8.447 min
Delta R.T.: 0.000 min
Response: 45814
Conc: 5.33 ng/ml



#44 AR-1268-4

R.T.: 10.221 min
Delta R.T.: 0.011 min
Response: 1608606
Conc: 281.63 ng/ml



#44 AR-1268-4

R.T.: 8.735 min
Delta R.T.: 0.002 min
Response: 999344
Conc: 286.51 ng/ml



#45 AR-1268-5

R.T.: 10.632 min
Delta R.T.: 0.010 min
Response: 414053
Conc: 10.16 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.011 min
Delta R.T.: 0.000 min
Response: 248379
Conc: 10.14 ng/ml