

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010421\
 Data File : PP032541.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jan 2021 19:40
 Operator : DD\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 05:45:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 05 05:44:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.747	4.034	2922177	2573807	50.000	50.000
2)	SA Decachlor...	10.587	9.318	2297326	2262340	50.000	50.000
Target Compounds							
21)	L5 AR-1248-1	6.050	5.283	534062	474914	500.000	500.000
22)	L5 AR-1248-2	6.344	5.545	693656	637174	500.000	500.000
23)	L5 AR-1248-3	6.558	5.589	862416	684968	500.000	500.000
24)	L5 AR-1248-4	6.983	5.773	951087	829398	500.000	500.000
25)	L5 AR-1248-5	7.021	6.196	955275	814615	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010421\
Data File : PP032541.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Jan 2021 19:40
Operator : DD\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 05:45:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 05:44:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 2922177
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 2573807
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.587 min
Delta R.T.: 0.000 min
Response: 2297326
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.318 min
Delta R.T.: 0.000 min
Response: 2262340
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 6.050 min
Delta R.T.: 0.000 min
Response: 534062
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 5.283 min
Delta R.T.: 0.000 min
Response: 474914
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 693656
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 637174
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.558 min
Delta R.T.: 0.000 min
Response: 862416
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 684968
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 951087
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 829398
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 7.021 min
Delta R.T.: 0.000 min
Response: 955275
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 6.196 min
Delta R.T.: 0.000 min
Response: 814615
Conc: 500.00 ng/ml