

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010421\
 Data File : PP032562.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Jan 2021 1:31
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP010421AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 07:54:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 05 07:49:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.747	4.033	2940734	2520971	53.773	52.084
2) SA Decachlor...	10.586	9.319	2219749	2171137	49.583	48.898
Target Compounds						
26) L6 AR-1254-1	6.952	6.153	937262	1170177	470.505	466.071
27) L6 AR-1254-2	7.181	6.311	1425612	1038121	470.571	484.421
28) L6 AR-1254-3	7.559	6.731	1537466	1687375	471.907	471.822
29) L6 AR-1254-4	7.853	6.971	1109230	969596	481.402	470.712
30) L6 AR-1254-5	8.281	7.398	1246416	1531993	507.547	495.416



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010421\
Data File : PP032562.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Jan 2021 1:31
Operator : DD\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP010421AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 07:54:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 07:49:02 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.747 min
Delta R.T.: -0.002 min
Response: 2940734
Conc: 53.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP010421AR1254




#1 Tetrachloro-m-xylene

R.T.: 4.033 min
Delta R.T.: -0.002 min
Response: 2520971
Conc: 52.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.586 min
Delta R.T.: -0.002 min
Response: 2219749
Conc: 49.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.319 min
Delta R.T.: 0.000 min
Response: 2171137
Conc: 48.90 ng/ml



#26 AR-1254-1


R.T.: 6.952 min
Delta R.T.: 0.000 min
Response: 937262
Conc: 470.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP010421AR1254




#26 AR-1254-1

R.T.: 6.153 min
Delta R.T.: 0.000 min
Response: 1170177
Conc: 466.07 ng/ml




#27 AR-1254-2

R.T.: 7.181 min
Delta R.T.: 0.000 min
Response: 1425612
Conc: 470.57 ng/ml



#27 AR-1254-2


R.T.: 6.311 min
Delta R.T.: 0.000 min
Response: 1038121
Conc: 484.42 ng/ml



#28 AR-1254-3


R.T.: 7.559 min
Delta R.T.: 0.000 min
Response: 1537466
Conc: 471.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP010421AR1254




#28 AR-1254-3

R.T.: 6.731 min
Delta R.T.: 0.000 min
Response: 1687375
Conc: 471.82 ng/ml




#29 AR-1254-4

R.T.: 7.853 min
Delta R.T.: 0.000 min
Response: 1109230
Conc: 481.40 ng/ml



#29 AR-1254-4


R.T.: 6.971 min
Delta R.T.: 0.000 min
Response: 969596
Conc: 470.71 ng/ml



#30 AR-1254-5

R.T.: 8.281 min
Delta R.T.: 0.000 min
Response: 1246416
Conc: 507.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP010421AR1254



#30 AR-1254-5

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 1531993
Conc: 495.42 ng/ml