

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP010422.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jan 05 05:31:42 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042502.D 1000 =PP042500.D 750 =PP042501.D
 250 =PP042503.D 50 =PP042504.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.001	1.964	1.954	1.985	1.750	1.931 E4	5.31
2) SA	Decachlorobiphenyl	2.045	2.018	2.017	2.015	1.881	1.995 E4	3.26
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	2.064					2.064 E2	0.00
9) L2	AR-1221-2	1.582					1.582 E2	0.00
10) L2	AR-1221-3	5.185					5.185 E2	0.00
11) L3	AR-1232-1	4.222					4.222 E2	0.00
12) L3	AR-1232-2	1.942					1.942 E2	0.00
13) L3	AR-1232-3	3.564					3.564 E2	0.00
14) L3	AR-1232-4	1.772					1.772 E2	0.00
15) L3	AR-1232-5	1.242					1.242 E2	0.00
16) L4	AR-1242-1	4.721					4.721 E2	0.00
17) L4	AR-1242-2	6.811					6.811 E2	0.00
18) L4	AR-1242-3	4.133					4.133 E2	0.00
19) L4	AR-1242-4	3.499					3.499 E2	0.00
20) L4	AR-1242-5	3.173					3.173 E2	0.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	7.829					7.829 E2	0.00
37) L8	AR-1262-2	1.337					1.337 E3	0.00
38) L8	AR-1262-3	6.591					6.591 E2	0.00
39) L8	AR-1262-4	4.833					4.833 E2	0.00
40) L8	AR-1262-5	4.879					4.879 E2	0.00
41) L9	AR-1268-1	1.654	1.615	1.631	1.614	1.667	1.636 E3	1.46
42) L9	AR-1268-2	1.534	1.504	1.508	1.526	1.491	1.513 E3	1.15
43) L9	AR-1268-3	1.372	1.366	1.352	1.342	1.296	1.346 E3	2.23
44) L9	AR-1268-4	5.138	5.299	5.273	5.451	6.210	5.474 E2	7.78
45) L9	AR-1268-5	4.485	4.580	4.558	4.467	4.224	4.463 E3	3.17

Signal #2 Calibration Files

500 =PP042502.D 1000 =PP042500.D 750 =PP042501.D
 250 =PP042503.D 50 =PP042504.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.706	2.615	2.641	2.747	2.621	2.666 E4	2.17
2) SA	Decachlorobiphenyl	3.357	3.087	3.200	3.594	3.799	3.407 E4	8.52
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP010422.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jan 05 05:31:42 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042502.D 1000 =PP042500.D 750 =PP042501.D
 250 =PP042503.D 50 =PP042504.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	3.459					3.459 E2	0.00
9) L2	AR-1221-2	2.685					2.685 E2	0.00
10) L2	AR-1221-3	8.236					8.236 E2	0.00
11) L3	AR-1232-1	6.635					6.635 E2	0.00
12) L3	AR-1232-2	5.950					5.950 E2	0.00
13) L3	AR-1232-3	3.269					3.269 E2	0.00
14) L3	AR-1232-4	2.945					2.945 E2	0.00
15) L3	AR-1232-5	3.201					3.201 E2	0.00
16) L4	AR-1242-1	8.213					8.213 E2	0.00
17) L4	AR-1242-2	1.111					1.111 E3	0.00
18) L4	AR-1242-3	6.393					6.393 E2	0.00
19) L4	AR-1242-4	6.214					6.214 E2	0.00
20) L4	AR-1242-5	7.403					7.403 E2	0.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	8.681					8.681 E2	0.00
37) L8	AR-1262-2	1.386					1.386 E3	0.00
38) L8	AR-1262-3	1.069					1.069 E3	0.00
39) L8	AR-1262-4	1.927					1.927 E3	0.00
40) L8	AR-1262-5	9.018					9.018 E2	0.00
41) L9	AR-1268-1	2.993	2.783	2.861	3.156	3.358	3.030 E3	7.64
42) L9	AR-1268-2	2.820	2.630	2.704	2.951	3.053	2.831 E3	6.13
43) L9	AR-1268-3	2.383	2.200	2.272	2.531	2.692	2.416 E3	8.23
44) L9	AR-1268-4	1.021	0.948	0.999	1.041	1.017	1.005 E3	3.50
45) L9	AR-1268-5	7.025	6.688	6.813	7.219	7.849	7.119 E3	6.40

(#) = Out of Range